


Instructions For Use

Rotors and Tubes

For Beckman Coulter
J2, J6, and Avanti J Series Centrifuges



PN JR-IM-10AG
October 2014



Beckman Coulter, Inc.
250 S. Kraemer Blvd.
Brea, CA 92821 U.S.A.



Rotors and Tubes
For J2, J6, and Avanti J Series Centrifuges
PN JR-IM-10AG (October 2014)

© 2014 Beckman Coulter, Inc.
All rights reserved

No part of this document may be reproduced or transmitted in any form or by any means, electronic, mechanical, photocopying, recording, or otherwise, without prior written permission from Beckman Coulter, Inc.

Beckman Coulter, the stylized logo, Avanti, Aerosolve, HarvestLine, Microfuge, and Quick-Seal are trademarks of Beckman Coulter, Inc. and are registered in the USPTO.

All other trademarks, service marks, products, or services are trademarks or registered trademarks of their respective holders.

Find us on the World Wide Web at:
www.beckmancoulter.com

Printed in U.S.A.

Revision History

This document applies to the latest software listed and higher versions. When a subsequent software version changes the information in this document, a new issue will be released.

Revision AE, December 2013

Added the JA-14.50 rotor: See [Table 1.1, Rotors Used in Beckman Coulter J Series Centrifuges](#) and [Table 4.1, General Specifications for Beckman Coulter J Series Fixed-Angle Rotors](#).

Revision AF, February 2014

Changed Polyallomer to polypropylene:

- [CHAPTER 1, Pelleting \(Differential Separation\)](#)
- [CHAPTER 2, Characteristics and Chemical Resistances of Tube and Bottle Materials](#)
- [CHAPTER 2, Labware Types](#)
- [CHAPTER 3, Gradient Preparation](#)
- [CHAPTER 2, Gradient Formation and Fractionation](#)
- [CHAPTER 2, Labware Types](#)
- [CHAPTER 3, Gradient Preparation](#)
- [CHAPTER 3, General Filling and Sealing Requirements for Tubes and Bottles](#)
- [CHAPTER 3, Open-Top Polypropylene Tubes](#)
- [CHAPTER 3, Capping Tubes](#)
- [CHAPTER 3, Filling and Sealing Quick-Seal Tubes](#)
- [CHAPTER 7, Cleaning](#)
- [CHAPTER 7, Sterilization and Disinfection](#)
- [CHAPTER 7, Tube and Bottle Sterilization and Disinfection](#)

Revision AG, October 2014

Added Avanti JXN-30 Centrifuge, Avanti JXN-26 Centrifuge, and JA-18 Rotor: See [CHAPTER 1, Rotors Used in Beckman Coulter J Series Centrifuges](#).

Safety Notice

Read all product manuals and consult with Beckman Coulter-trained personnel before attempting to operate instrument. Do not attempt to perform any procedure before carefully reading all instructions. Always follow product labeling and manufacturer's recommendations. If in doubt as to how to proceed in any situation, contact your Beckman Coulter Representative.

Alerts for Warning, Caution, Important, and Note

WARNING

WARNING indicates a potentially hazardous situation which, if not avoided, could result in death or serious injury.

CAUTION

CAUTION indicates a potentially hazardous situation, which, if not avoided, may result in minor or moderate injury.

IMPORTANT IMPORTANT is used for comments that add value to the step or procedure being performed. Following the advice in the Important adds benefit to the performance of a piece of equipment or to a process.

NOTE NOTE is used to call attention to notable information that should be followed during installation, use, or servicing of this equipment.

This safety notice summarizes information basic to the safe operation of the rotors and accessories described in this manual. The international symbol displayed above is a reminder that all safety instructions should be read and understood before use or maintenance of rotors or accessories. When you see the symbol on other pages, pay special attention to the safety information presented. Also observe any safety information contained in applicable rotor and centrifuge manuals. Observance of safety precautions will help to avoid actions that could cause personal injury, as well as damage or adversely affect the performance of the centrifuge/rotor/tube system.

Chemical and Biological Safety

Normal operation may involve the use of solutions and test samples that are pathogenic, toxic, or radioactive. Such materials should not be used in these rotors, however, unless all necessary safety precautions are taken.

- Observe all cautionary information printed on the original solution containers prior to their use.
- Handle body fluids with care because they can transmit disease. No known test offers complete assurance that they are free of micro-organisms. Some of the most virulent—Hepatitis (B and C) and HIV (I–V) viruses, atypical mycobacteria, and certain systemic fungi—further emphasize

the need for aerosol protection. Handle other infectious samples according to good laboratory procedures and methods to prevent spread of disease. Because spills may generate aerosols, observe proper safety precautions for aerosol containment. Do not run toxic, pathogenic, or radioactive materials in the rotor without taking appropriate safety precautions. Biosafe containment should be used when Risk Group II materials (as identified in the *World Health Organization Laboratory Biosafety Manual*) are handled; materials of a higher group require more than one level of protection.

- Dispose of all waste solutions according to appropriate environmental health and safety guidelines.
- If disassembly reveals evidence of leakage, you should assume that some fluid escaped the container or rotor. Apply appropriate decontamination procedures to the centrifuge, rotor, and accessories.

Mechanical Safety

- Use only the rotors, components, and accessories designed for use in the rotor and centrifuge being used (refer to the applicable rotor manual). *The safety of rotor components and accessories made by other manufacturers cannot be ascertained by Beckman Coulter. Use of other manufacturers' components or accessories in Beckman Coulter rotors may void the rotor warranty and should be prohibited by your laboratory safety officer.*
- Rotors are designed for use at the speeds indicated; however, speed reductions may be required because of weight considerations of tubes, adapters, and/or the density of the solution being centrifuged. Be sure to observe the instructions in the applicable rotor manual.
- NEVER attempt to slow or stop a rotor by hand.
- The strength of containers can vary between lots, and will depend on handling and usage. We highly recommend that you pretest them in the rotor (using buffer or gradient of equivalent density to the intended sample solution) to determine optimal operating conditions. Scratches (even microscopic ones) significantly weaken glass and polycarbonate containers.

To help prevent premature failures or hazards by detecting stress corrosion, metal fatigue, wear or damage to anodized coatings, and to instruct laboratory personnel in the proper care of rotors, Beckman Coulter offers the Field Rotor Inspection Program (FRIP). This program involves a visit to your laboratory by a specially trained Beckman Coulter representative, who will inspect all of your rotors for corrosion or damage. The representative will recommend repair or replacement of at-risk rotors to prevent potential rotor failures. Contact your local Beckman Coulter office to request this service.

It is your responsibility to decontaminate the rotors and accessories before requesting service by Beckman Coulter Field Service.

RoHS Notice

These labels and materials declaration table (the Table of Hazardous Substance's Name and Concentration) are to meet People's Republic of China Electronic Industry Standard SJ/T11364-2006 "Marking for Control of Pollution Caused by Electronic Information Products" requirements.



China RoHS Caution Label — This label indicates that the electronic information product contains certain toxic or hazardous substances. The center number is the Environmentally Friendly Use Period (EFUP) date, and indicates the number of calendar years the product can be in operation. Upon the expiration of the EFUP, the product must be immediately recycled. The circling arrows indicate the product is recyclable. The date code on the label or product indicates the date of manufacture.



China RoHS Environmental Label — This label indicates that the electronic information product does not contain any toxic or hazardous substances. The center "e" indicates the product is environmentally safe and does not have an Environmentally Friendly Use Period (EFUP) date. Therefore, it can safely be used indefinitely. The circling arrows indicate the product is recyclable. The date code on the label or product indicates the date of manufacture.

Contents

Revision History, iii

Safety Notice, v

Alerts for Warning, Caution, Important, and Note, v

Chemical and Biological Safety, v

Mechanical Safety, vi

RoHS Notice, vii

Scope, xvii

Scope of this Manual, xvii

Rotors, 1-1

Introduction, 1-1

General Description, 1-1

Rotor Designations, 1-1

Materials, 1-3

Drive Pins, 1-4

Rotor Selection, 1-4

Pelleting (Differential Separation), 1-11

Isopycnic Separations, 1-15

Rate Zonal Separations, 1-15

Blood Component Separations, 1-16

General Operating Information, 1-16

Rotor Balance, 1-16

Rotor Tie-Down, 1-17

Overspeed Protection, 1-18

Allowable Run Speeds, 1-18

Temperature Compensation, 1-19

Avanti J Centrifuges, 1-19

Avanti J2 Series Centrifuges (No longer manufactured), 1-20

Analog J6 Series Centrifuges (No longer manufactured), 1-20

Microprocessor-Controlled Centrifuges, 1-21

Tubes, Bottles, and Accessories, 2-1

Introduction, 2-1

Labware Selection Criteria , 2-1

Labware Material Compatibility with Solvents and Sample, 2-2

Gradient Formation and Fractionation, 2-3

Labware Types, 2-3

Polypropylene Tubes, 2-3

Open-Top Polypropylene Tubes, 2-3

Quick-Seal Polypropylene Tubes, 2-4

Polycarbonate Tubes, 2-4

Polyethylene Tubes, 2-4

Ultra-Clear Tubes, 2-4

Stainless Steel Tubes, 2-5

Microfuge Tubes, 2-5

Bottles, 2-5

Multiwell Titer Plates, 2-5

96-Well Microtiter Plates, 2-5

Deep-Well Titer Plates (and Caps), 2-6

Square-Well Titer Plates, 2-6

Temperature Limits, 2-6

Spacers and Floating Spacers, 2-7

Adapters, 2-7

Bottle Adapters, 2-7

Multitube Adapters, 2-7

Solid Multitube Adapters, 2-8

Modular Disk Adapters, 2-8

Bottle and Tube Caps, 2-9

Aerosolve Canisters, 2-9

Blood Bag Cups, 2-10

Rotor Labware Assemblies, 2-10

Tubes, Bottles, and Accessories, 3-1

Introduction, 3-1

Gradient Preparation, 3-1

General Filling and Sealing or Capping Requirements, 3-2

Working with Physiological Fluids, 3-4

Filling Open-Top Tubes, 3-5

Open-Top Polypropylene Tubes, 3-5

Swinging-Bucket Rotors, 3-5

Fixed-Angle Rotors, 3-5

Other Open-Top Tubes, 3-5

Polycarbonate, 3-5

Ultra-Clear, 3-5

Polypropylene, 3-6
Polyethylene, 3-6
Stainless Steel, 3-6
Capping Tubes, 3-6
Filling and Capping Tubes, 3-7
Three-Piece Assemblies, 3-7
JLA-8.1000 and JLA-9.1000 Bottle Cap/Closure, 3-7
Filling and Loading Cups in the JS-5.0 Rotor, 3-8
Filling and Sealing Quick-Seal Tubes, 3-9
Capping Multiwell Titer Plates, 3-14
Cap Strips, 3-14
Aluminum Foil Lids, 3-14
Using Adapters, 3-14
Using Solid Multitube Adapters, 3-14
Using Modular Disk Multitube Adapters, 3-15
Using Aerosolve Canisters, 3-16
Using Canisters as Wide-Mouth Bottles, 3-17
Using Canisters with Tube Racks, 3-18
Using Blood Bag Cups, 3-19
Sample Recovery, 3-20
Capped Tubes, 3-21
JS-5.0 Cups, 3-21
Quick-Seal Tubes, 3-22
Making Ultra-Clear Tubes Wettable, 3-23
Using Fixed-Angle Rotors, 4-1
Introduction, 4-1
Description, 4-1
Tubes and Bottles, 4-6
Rotor Preparation and Loading, 4-7
Prerun Safety Check, 4-7
Rotor Preparation, 4-7
Operation, 4-8
Installing the Rotor, 4-9
Removal and Sample Recovery, 4-10
Using Swinging-Bucket Rotors, 5-1
Introduction, 5-1
Description, 5-1
Labware, 5-4
Rotor Preparation and Loading, 5-5
Prerun Safety Check, 5-5

- Rotor Preparation, 5-5
 - Special Preparation Instructions for JS-24 Series Rotors, 5-6
- Loading the Rotor Yoke, 5-7
 - Loading JS-24 Series Rotors, 5-8
 - Symmetric and Balanced Loading, 5-9
- Loading Buckets, 5-11
 - Using Modular Disk Adapters, 5-11
 - Using Blood Bag Cups, 5-12
- Loading Buckets Into the Rotor, 5-13
- Using Microtiter Plate Carriers, 5-14
 - Micro Plus Carriers, 5-14
 - JS-5.9 and JS-5.3 Plate Carriers, 5-15
 - J6 Carriers, 5-16

Operation, 5-17

Sample Recovery, 5-17

- Removing JS-24 Series Rotors, 5-17

Using Vertical-Tube and Rack-Type Rotors, 6-1

Introduction, 6-1

Description, 6-1

- Vertical-Tube Rotors, 6-2
- Rack-Type Rotors, 6-2

Using a Vertical-Tube Rotor, 6-3

- Tubes and Bottles, 6-3
- Rotor Preparation and Loading, 6-3
 - Prerun Safety Checks, 6-3
 - Rotor Preparation, 6-3
- Operation, 6-4
- Installing the Rotor, 6-5
- Removal and Sample Recovery, 6-6

Using a Rack-Type Rotor, 6-7

- Trays and Tubes, 6-7
- Rotor Preparation and Loading, 6-7
 - Prerun Safety Checks, 6-7
 - Rotor Preparation, 6-7
- Operation, 6-8
- Installing the Rotor, 6-9
- Removal and Sample Recovery, 6-10

Care and Maintenance, 7-1

Introduction, 7-1

Rotor Care, 7-1

- Cleaning, 7-2
- Decontamination, 7-3

Sterilization and Disinfection, 7-4
Inspection, 7-4
Lubrication, 7-5
O-Rings, 7-5
Pivot Pins and Buckets, 7-5
Field Rotor Inspection Program, 7-6
Tube, Bottle, and Accessory Care, 7-6
Cleaning, 7-6
Decontamination, 7-7
Sterilization and Disinfection, 7-7
Inspection, 7-9
Tube and Bottle Storage, 7-9
Returning a Rotor or Accessory to the Factory, 7-10
Diagnostic Hints, 7-10
 Chemical Resistances for Beckman Coulter Centrifugation Products, A-1
List of Chemical Resistances, A-1
 Temperature Compensation Tables, B-1
Introduction, B-1
Temperature Compensation, B-1
 Gradient Materials, C-1
Introduction, C-1
 Blood Component Separation, D-1
Introduction, D-1
Blood Bank Collection Overview, D-1
Components and Typical Usage, D-2
Single-Donor Fresh Plasma, D-2
Single-Donor Plasma, D-2
Packed Red Blood Cells (PBC), D-2
Leukocyte-Depleted Red Blood Cells, D-2
Shelf Life, D-3
Freezing, D-3
Separation of Blood Components by Centrifugation, D-3
Tips for Optimum Centrifugation Runs, D-6
 References, E-1
List of References, E-1

Glossary

Beckman Coulter, Inc.
J Series Rotor Warranty

Illustrations

- 1.1 Fixed-Angle, Swinging-Bucket, Vertical-Tube, and Rack-Type Rotors, 1-2
- 1.2 Particle Separation in Fixed-Angle, Swinging-Bucket, and Vertical-Tube Rotors., 1-10
- 1.3 Sedimentation Coefficients (in Svedberg Units) for Some Common Biological Materials, 1-12
- 1.4 Nomogram for J2 Series Centrifuges, 1-13
- 1.5 Nomogram for J6Series Centrifuges, 1-14
- 1.6 Arranging Tubes Symmetrically in a Fixed-Angle, Vertical-Tube, or JS-24 Series Swinging-Bucket Rotor, 1-17
- 3.1 The Cordless Quick-Seal Tube Topper, 3-10
- 4.1 Examples of Fixed-Angle Rotors, 4-2
- 5.1 Examples of Swinging-Bucket Rotors, 5-2
- 5.2 Examples of Correctly and Incorrectly Loaded Buckets and Carriers, 5-10
- 5.3 Typical Blood Bag Loading Procedures (JS-24.3 Rotor Shown), 5-13
- 5.4 The Micro Plus Microtiter Plate Carrier, Base, Pad, and Deep-Well Microtiter Plate, 5-15
- 6.1 Vertical-Tube Rotor, 6-2
- 6.2 Rack-Type Rotor, 6-2
- D.1 Blood Component Separation, D-4

Tables

1.1	Rotors Used in Beckman Coulter J Series Centrifuges, 1-5
2.1	Characteristics and Chemical Resistances of Tube and Bottle Materials, 2-2
3.1	General Filling and Sealing Requirements for Tubes and Bottles, 3-3
3.2	Aerosolve Tube Rack, 3-18
4.1	General Specifications for Beckman Coulter J Series Fixed-Angle Rotors, 4-3
5.1	General Specifications for Beckman Coulter J Series Swinging-Bucket Rotors, 5-3
5.2	Microplate Carriers Used with J6-Series Rotors, 5-16
6.1	General Specifications for Beckman Coulter J Series Vertical-Tube and Rack-Type Rotors, 6-1
7.1	Tube and Bottle Sterilization and Disinfection, 7-8
7.2	Troubleshooting Chart, 7-11
B.1	Temperature Compensation Settings for the J2-HC Centrifuge, B-1
B.2	Temperature Compensation Settings for the J2-21, J2-21B, J2-21C, and J2-HS Centrifuges, B-3
B.3	Temperature Compensation Settings for the J2-MI, J2-21M, J2-MC, and J2-21M/E Centrifuges, B-5
B.4	Temperature Compensation Settings for the J6 Centrifuges, B-6
C.1	Commonly Used Gradient Materials with Their Solvents, C-2
C.2	Density, Refractive Index, and Concentration Data—Cesium Chloride at 25°C, Molecular Weight = 168.37, C-3
C.3	Density, Refractive Index, and Concentration Data—Sucrose at 20°C, Molecular Weight = 342.3, C-4
C.4	Density Conversion for Cesium and Rubidium Salts at 20°C, C-5
D.1	Blood Component Storage, D-3
D.2	Blood Bank Methods, D-5

Scope of this Manual

This manual contains general information for properly preparing a rotor for centrifugation in a Beckman Coulter J series centrifuge. This manual should be used with the individual rotor instruction manual shipped with each rotor. The rotor manuals provide specific information for each rotor, including special operating procedures and precautions, tube, bottle, and adapter part numbers, and equations to calculate maximum allowable rotor speeds. Each manual has a code number in the bottom left-hand corner of the cover page that can be used for reordering. To reorder, call Beckman Coulter Customer Service at 1-800-742-2345 (U.S.A. or Canada), outside the U.S., contact your local Beckman Coulter representative.

A lot of information is compiled in this manual, and we urge you to read it carefully — especially if this is your first experience with Beckman Coulter products.

- [CHAPTER 1](#) describes, by usage, Beckman Coulter's currently produced J series rotors; this should help you determine the appropriate rotor to use for a particular application. Also included in this section is a discussion of rotor materials, components, and centrifugation techniques.
- [CHAPTER 2](#) describes various tubes, adapters, spacers, and cannisters to help you choose a particular container for your application.
- [CHAPTER 3](#) provides instructions for using tubes, bottles, cannisters, and related accessories.
- [CHAPTER 4](#) contains step-by-step procedures for preparing a fixed angle rotor for a centrifuge run. Similar information is available for swinging bucket rotors in [CHAPTER 5](#), and [CHAPTER 6](#) contains the same type of information for vertical tube and rack-type rotors. (Elutriation, zonal, and continuous flow rotors are not covered in this manual.)
- [CHAPTER 7](#) provides rotor, tube, and accessory care and maintenance information, as well as some diagnostic hints. Please read it. Proper rotor care results in longer rotor life.

- Several appendixes contain information that may be of special interest:
 - [APPENDIX A](#) lists chemical resistances for rotor and accessory materials to help determine compatibility with a variety of solutions.
 - [APPENDIX B](#) contains Temperature Compensation Tables for various rotors.
 - [APPENDIX C](#) contains reference information on some commonly used gradient materials.
 - [APPENDIX D](#) provides information about separation of blood components using J series centrifuges.
 - [APPENDIX E](#) lists references for further reading.
 - [Glossary](#) provides a glossary of terms.

Introduction

This chapter is an introduction to the Beckman Coulter family of J series rotors, providing general information on rotor design, selection, and operation. Rotor designs described are fixed angle, swinging bucket, vertical tube, and rack type. Specific instructions for using each type of rotor are contained in [CHAPTER 4](#), [CHAPTER 5](#), and [CHAPTER 6](#). Care and maintenance information for all of these rotors is contained in [CHAPTER 7](#). Elutriator, continuous flow, and zonal rotors are not covered in this manual. The elutriator rotors are described in detail in their respective rotor instruction manuals, publications JE6B-IM-9 and JE5-IM-13; the continuous flow/zonal rotor, JCF-Z, is described in publication JCFZ-IM-12.

General Description


Rotor Designations

Beckman Coulter J series rotors are usually named according to the type of rotor and the rotor's maximum allowable revolutions per minute (in thousands), referred to as rated speed. For example, the JA-12 rotor is a fixed-angle rotor with a maximum speed of 12,000 rpm. However, the naming system for J series rotors was changed slightly in early 1994.

- Rotors released before 1994 (for example, the JA-18.1): JA designates that it is a fixed-angle rotor used in a J series centrifuge; the 18 indicates that the rated speed of the rotor is 18,000 rpm; the decimal unit (.1) distinguishes between different rotors with the same rated speed.
- Rotors released after January 1994 (for example, the JA-25.50): JA still designates that it is a fixed-angle rotor used in a J series centrifuge; the 25 still identifies the rated speed of the rotor (25,000 rpm); but the decimal unit (.50) describes the nominal volume of the largest tube or bottle (in mL) used in the rotor.

An example of each rotor type is shown in [Figure 1.1](#).

Figure 1.1 Fixed-Angle, Swinging-Bucket, Vertical-Tube, and Rack-Type Rotors



Containers in *fixed-angle rotors* (designated **JA**) are held at an angle to the axis of rotation in tube cavities.

Containers in *J-Lite fixed-angle rotors* (designated **JLA**) are also held at an angle to the axis of rotation; the rotor construction results in reduced overall weight.


Containers in *swinging-bucket rotors* (designated **JS**) are held in rotor buckets or multitube carriers attached to the rotor body by hinge pins. The buckets or carriers swing out to a horizontal position as the rotor accelerates.

Tubes in *vertical-tube rotors* (designated **JV**) are held parallel to the axis of rotation. These rotors have plugs, screwed into the rotor cavities over sealed tubes, that keep the tubes in the cavities and provide support for the hydrostatic forces generated by centrifugation.


Tubes in the *rack-type rotor* (designated **JR**) are held in gamma-counter racks. Racks are loaded into special plastic trays, which are then loaded into carriers at a resting angle. During centrifugation, the carriers swing out to a completely horizontal position.

Materials


Most Beckman Coulter J series rotors are made of aluminum and are anodized to protect the metal from corrosion. (The JS-13.1 and JS-7.5 rotors are painted with polyurethane paint and are not anodized.) The anodized coating is a thin, hard layer of aluminum oxide formed electrochemically in the final stages of rotor fabrication. A black or colored dye may be applied over the oxide for rotor family identification. The coating can be damaged if careful cleaning procedures are not followed. Therefore, it is especially important to clean aluminum rotors with brushes that will not scratch the anodized coating and to use a noncorrosive, neutral-pH detergent. Refer to [CHAPTER 7](#) for cleaning and maintenance procedures.



Some J series rotors have attached windshields to reduce air friction. The windshields are made of anodized aluminum



Canisters used in some J-Lite rotors are made of lightweight carbon fiber epoxy composite. The lightweight canisters make the overall rotor weight significantly lighter than a comparably sized all-aluminum rotor. Each canister has a sleeve washer, made of polytetrafluoroethylene and polyetherimide (PEI), which acts as a sleeve between the canister and the aluminum rotor body. A lubricated ethylene propylene rubber O-ring inside the canister closure helps create a secondary seal during centrifugation.




Transparent plastic covers are available for some swinging-bucket rotor buckets, to help contain spills and glass particles in the event of tube breakage. The covers are made of high-impact polyetherimide (PEI). Each cover requires an O-ring.


The O-rings or gaskets in rotor assemblies with lids are made of Buna N elastomer and maintain atmospheric pressure in the rotor if they are kept clean and lightly coated with silicone vacuum grease. Plug gaskets in vertical-tube rotors are made of thermoplastic polyester elastomer and do not require coating.

Drive Pins

Currently produced J series rotors have drive pins in the drive hole. These pins mesh with teeth on the centrifuge drive spindle hub when the rotor is installed to ensure that the rotor does not slip on the hub during initial acceleration. Most drive pins are oriented horizontally (or angled) in the drive hole; however, some are oriented vertically.



All rotors used in Avanti J series centrifuges must have drive pins in the rotor drive hole. Some Beckman Coulter rotors, including the JA-10 and the JS-7.5, were previously manufactured without drive pins because pins were not needed when these rotors were used in J2 series centrifuges. Check all J series rotors for drive pins before using them in an Avanti J series centrifuge. To check for drive pins, hold the rotor up or turn it on its side and look into the drive hole. If you do not see two metal pins near the top of the hole, do not use the rotor in the Avanti J. Call your local Beckman Coulter office for information on returning the rotor to the factory for upgrading.



In fixed-angle and vertical-tube rotors manufactured since early 1997, the rotor pins are positioned parallel to the **BECKMAN** name engraved at the center of the rotor body. Knowing the pin orientation before you install the rotor will help to ensure that you position the rotor properly on the hub, minimizing the chance of hub damage.

Rotor Selection

Rotors used in Beckman Coulter J series centrifuges are listed in [Table 1.1](#). General rotor specifications for each fixed-angle rotor are in [Table 4.1](#), swinging buckets in [Table 5.1](#), and vertical-tube and rack-type in [Table 6.1](#). Detailed descriptions of each rotor are included in the applicable rotor manual.

Table 1.1 Rotors Used in Beckman Coulter J Series Centrifuges^a

Rotor	Nominal Rotor Capacity	Max Speed (rpm)	Max RCF (x g)	k Factor	Avanti J-30 I	Avanti JXN 30	(Avanti J-20 XP Series)	(Avanti J-25 Series)	(Avanti J-26 XP Series)	Avanti J-26S XP Series	Avanti JXN-26	Avanti J-E	Avanti J-HC	(J2-MC)	(J2-HS)	(J2-HC)	J6-MI	(J6-MC)	(J6-HC)
Fixed Angle																			
JA-30.50 Ti	400 mL	30,000	108,860	280	X	X	X	X	X	X	X			X	X	X			
JA-25.50	400 mL	25,000	76,600 ^b	418	X	X	X	X	X	X	X	X		X	X	X			
JA-25.15	360 mL	25,000	74,200 (outer row) 60,200 (inner row)	265 380	X	X	X	X	X	X	X			X	X	X			
JA-21	180 mL	21,000	50,400	470	X	X	X	X	X	X	X	X		X	X	X	X	X	X
JA-20.1	180 mL	20,000	51,500 (outer row) 43,900 (inner row)	325 371	X	X	X	X	X	X	X	X		X	X	X	X	X	X
JA-20	400 mL	20,000	48,400	770	X	X	X	X	X	X	X	X		X	X	X	X	X	X
JA-18.1 (45° angle adapter)	43.2 mL	18,000 ^c	42,100	156	X	X	X	X	X	X	X			X	X	X	X	X	X
JA-18.1 (25° angle adapter)	43.2 mL	17,000 ^d	36,300	91	X	X	X	X	X	X	X	X			X	X	X	X	X
JA-18	1 liter	18,000	47,900	56x	X	X	X	X	X	X	X	X		X	X	X	X	X	X
JA-17	700 mL	17,000 ^e	39,800	690	X	X	X	X	X	X	X	X		X	X	X	X	X	X

1-6 **Table 1.1** Rotors Used in Beckman Coulter J Series Centrifuges^a (Continued)

Rotor	Nominal Rotor Capacity	Max Speed (rpm)	Max RCF (x g)	k Factor	Avanti J-30 I	Avanti JXN 30	(Avanti J-20 XP Series)	(Avanti J-25 Series)	(Avanti J-26 XP Series)	Avanti J-26S XP Series	Avanti JXN-26	Avanti J-E	Avanti J-HC	(J2-MC)	(J2-HS)	(J2-HC)	J6-MI	(J6-MC)	(J6-HC)
JLA-16.250	1.5 liter	16,000 ^f	38,400	1090	X	X	X	X	X	X	X	X		X	X				
JA-14	1.5 liter	14,000	30,100	1764	X	X	X	X	X	X	X	X		X	X	X	X	X	X
F14BCI-14x50cy	700 mL	14,000	33,500	789	X		X	X	X	X		X							
F14BCI-6x250y	1500 mL	14,000	30,000	1690	X		X	X	X	X		X							
JA-12	600 mL	12,000	23,200	1244	X	X	X	X	X	X	X	X		X	X	X			
JA-10 ^g	3 liters	10,000 ^h	17,700	3610	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X
JLA-10.500	3 liters	10,000 ^h	18,500	2850	X		X	X	X	X	X	X		X	X	X			
F10BCI-6x500y	3 liters	10,000	17,696	3417	X	X	X	X	X	X	X	X							
JLA-9.1000	4 liters	9,000 ⁱ	16,800	2540	X	X	X	X	X	X	X	X							
JLA-8.1000	6 liters	8,000	15,900	2500			X		X	X	X		X						
JA-14.50	800 mL	14,000	35,000	787	X	X	X	X	X	X	X	X							
Swinging Bucket																			
JS-24.38	231 mL	24,000 ^j	103,900	334	X	X	X	X	X	X	X								
JS-24.15	90 mL	24,000 ^j	110,500	376	X	X	X	X	X	X	X								
JS-13.1	300 mL	13,000	26,500	1841	X	X	X	X	X	X	X	X		X	X	X	X	X	X
JS-7.5	200 mL	7,500	10,400	1090	X	X	X	X	X	X	X			X	X	X	X	X	X

Table 1.1 Rotors Used in Beckman Coulter J Series Centrifuges^a (Continued)

Rotor	Nominal Rotor Capacity	Max Speed (rpm)	Max RCF (x g)	k Factor	Avanti J-30 I	Avanti JXN 30	(Avanti J-20 XP Series)	(Avanti J-25 Series)	(Avanti J-26 XP Series)	Avanti J-26S XP Series	Avanti JXN-26	Avanti J-E	Avanti J-HC	(J2-MC)	(J2-HS)	(J2-HC)	J6-MI	(J6-MC)	(J6-HC)
JS-5.9	384 mL	5,900	6,570		X			X											
JS-5.3	691 mL	5,300	6,130				X		X	X	X	X							
JS-5.2	4 liters	5,200	6,840	9051														X	X
JS-5.0	9 liters	5,000	7,480	9171									X						
JS-4.3	3 liters	4,300	4,220	16,635			X		X	X	X					X			
JS-4.2	6 liters	4,200	5,020	11,502									X			X	X	X	X
JS-4.2A	6 liters	4,200	5,020	11,502									X				X	X	X
JS-4.2SM	6 quad blood bags	4,200	4,900														X	X	X
JS-4.2SMA	6 quad blood bags	4,200	4,900														X	X	X
JS-4.0	4 liters	4,000	4,050	15,298			X		X	X	X						X	X	X
(JS-3.4A-1250)	7.5 liters	3,400	3,370	18,066									X						
JS-3.0	6 liters	3,000	2,560	22,598													X	X	X
(JS-2.9)	6 liters	2,900	2,500	24,400													X	X	X
(JV-20)	312 mL	20,000	41,619	206										X	X	X			
(JR-3.2)	320 mL	2,280	25,606														X	X	X


Table 1.1 Rotors Used in Beckman Coulter J Series Centrifuges^a (Continued)

Rotor	Nominal Rotor Capacity	Max Speed (rpm)	Max RCF (x g)	k Factor	Avanti J-30 I	Avanti JXN 30	(Avanti J-20 XP Series)	(Avanti J-25 Series)	(Avanti J-26 XP Series)	Avanti J-26S XP Series	Avanti JXN-26	Avanti J-E	Avanti J-HC	(J2-MC)	(J2-HS)	(J2-HC)	J6-MI	(J6-MC)	(J6-HC)
-------	------------------------	-----------------	---------------	----------	---------------	---------------	-------------------------	----------------------	-------------------------	------------------------	---------------	------------	-------------	---------	---------	---------	-------	---------	---------


Zonal and Continuous Flow (see applicable rotor manual for rotor description and use)																			
JCF-Z	100 L/hr (HF seal assembly) 45 L/hr (SF seal assembly)	20,000	39,900		X	X	X	X	X	X	X			X	X	X			
JE-5.0	1000 mL	5,000	4,700				X		X	X							X	X	X
JE-6B	100 mL	6,000	5,080		X			X						X	X	X			

- a. Rotors and centrifuges in parentheses are no longer manufactured.
- b. Maximum speed in an Avanti J-E centrifuge is 21,000 rpm (18,000 rpm at 2°C at 35°C ambient and 95 percent humidity).
- c. When a JA-18.1 rotor is used in a J2-HC centrifuge, derate the rotor as follows: when the 45° adapters are used, do not run the rotor above 15,000 rpm; when 25° adapters are used, do not run the rotor above 16,000 rpm.
- d. When a JA-18.1 rotor is used in a J2-HC centrifuge, derate the rotor as follows: when the 45° adapters are used, do not run the rotor above 15,000 rpm; when 25° adapters are used, do not run the rotor above 16,000 rpm.
- e. Maximum speed in an Avanti J-E for the rotor with magnets; without magnets maximum is 13,000 rpm. (Maximum speed at 2° C in a 50-Hz centrifuge is 15,000 rpm.)
- f. Maximum speed in an Avanti J series centrifuge. Maximum speed in a J2 series centrifuge is 14,000 rpm.
- g. Newer JA-10 rotors with magnets, p/n 369867, cannot be used in J2 or J6 series centrifuges. Older JA-10 rotors without magnets, p/n 334833, can be used in J2 and J6 series centrifuges.
- h. Maximum speed in an Avanti J-E for rotor without magnets is 6300 rpm.
- i. Maximum speed for rotor in an Avanti J-E centrifuge is 6300 rpm.
- j. The JS-24.38 and JS-24.15 rotors can achieve 24,000 rpm in an Avanti J-30I centrifuge only. In Avanti J-26S XP series, J-26 XP series, J-25 series, J-20 series, and J-20 XP series centrifuges, the maximum speed for these rotors is 10,000 rpm.


Selection of a rotor depends on a variety of factors, such as sample volume, number of sample components to be separated, particle size, run time, required quality of separation, type of separation, and the centrifuge in use. Fixed-angle, swinging-bucket, vertical-tube, and rack-type rotors are designed to provide optimal separations for a variety of sample types. (For especially large sample volumes, continuous flow and zonal rotors are available.)




- *Fixed-angle rotors* are general-purpose rotors that are especially useful for pelleting subcellular particles and in short-column banding of viruses and subcellular organelles. Tubes are held at an angle (usually 20 to 45 degrees) to the axis of rotation. The tube angle shortens the particle pathlength (see [Figure 1.2](#)), compared to swinging-bucket rotors, resulting in reduced run times. Tubes can be placed directly in a rotor cavity if the diameters of the tube and the cavity are the same. Using adapters, more than one type and size of tube can be centrifuged together, provided that the loads are properly balanced. Refer to [CHAPTER 4](#) for specific information about the use of fixed-angle rotors.



- *Swinging-bucket rotors* are used for pelleting, isopycnic studies (separation as a function of density), and rate zonal studies (separation as a function of sedimentation coefficient). Large swinging-bucket rotors are used to obtain cell-free plasma or for cell packing. These rotors can be equipped with racks or microplate carriers to hold a variety of tubes, bottles, blood bags, or multiwell plates. Refer to [CHAPTER 5](#) for specific information about the use of swinging-bucket rotors.




- *Vertical-tube rotors* hold tubes parallel to the axis of rotation; therefore, bands separate across the diameter of the tube rather than down the length of the tube (see [Figure 1.2](#)). Only Quick-Seal tubes are used in vertical-tube rotors, making tube caps unnecessary. Refer to [CHAPTER 6](#) for specific information about the use of vertical-tube rotors.



- *Rack-type rotors* hold tubes in gamma-counter racks. Racks are loaded into special plastic trays, which are then loaded into carriers at a resting angle. During centrifugation, the carriers swing out to a completely horizontal position. Refer to [CHAPTER 6](#) for specific information about the use of rack-type rotors.

Figure 1.2 Particle Separation in Fixed-Angle, Swinging-Bucket, and Vertical-Tube Rotors.*



* Dark gray represents pelleted material, light gray is floating components, and bands are indicated by black lines.

Pelleting (Differential Separation)

Pelleting separates particles of different sedimentation coefficients, the largest particles in the sample traveling to the bottom of the tube (or bottle) first. Differential centrifugation is the successive pelleting of particles of decreasing sedimentation velocities, using increasingly higher forces and/or long run times. The relative pelleting efficiency of each rotor is measured by its k factor (clearing factor):

$$k = \frac{\ln(r_{\max}/r_{\min})}{\omega^2} \times \frac{10^{13}}{3600} \quad \text{EQ 1}$$

where ω is the angular velocity of the rotor in radians per second ($2\pi\text{RPM}/60$, or $\omega = 0.10472 \times \text{rpm}$), $r_{\max}$ is the maximum radius, and $r_{\min}$ is the minimum radius.

After substitution,

$$k = \frac{(2.533 \times 10^{11}) \ln(r_{\max}/r_{\min})}{\text{rpm}^2} \quad \text{EQ 2}$$

This factor can be used in the following equation to estimate the time t (in hours) required for pelleting:

$$t = \frac{k}{s} \quad \text{EQ 3}$$

where s is the sedimentation coefficient* of the particle of interest in Svedberg units. (Because s values in seconds are such small numbers, they are generally expressed in Svedberg units (S), where 1 S is equal to 10^{-13} seconds). It is usual practice to use the standard sedimentation coefficient $s_{20,w}$ based on sedimentation in water at 20°C. Clearing factors can be calculated at speeds other than maximum rated speed by use of the following formula:

$$k_{\text{adj}} = k \left(\frac{\text{rated speed of rotor}}{\text{actual run speed}} \right)^2 \quad \text{EQ 4}$$

Run times can also be calculated from data established in prior experiments when the k factor of the previous rotor is known. For any two rotors, a and b :


$$\frac{t_a}{t_b} = \frac{k_a}{k_b} \quad \text{EQ 5}$$

where the k factors have been adjusted for the actual run speed used.

Figure 1.3 lists sedimentation coefficients for some common biological materials. The k factors at rated speeds for Beckman Coulter J series rotors are provided in the table of general specifications in each rotor use section.

* $s = dr/dt \times 1/\omega^2 r$, where dr/dt is the sedimentation velocity.


Figure 1.3 Sedimentation Coefficients (in Svedberg Units) for Some Common Biological Materials



The centrifugal force exerted at a given radius in a rotor is a function of the rotor speed. The nomograms for J2 series and J6 series centrifuges in [Figure 1.4](#) and [Figure 1.5](#) allow you to determine relative centrifugal field (RCF) for a given radius and rotor speed. In Avanti J series centrifuges, the RCF is calculated automatically by the centrifuge software.


Run times can be shortened by using partially filled thickwall polypropylene and polycarbonate tubes. The short pathlength means less distance for particles to travel in the portion of the tube experiencing greatest centrifugal force, and hence shortened run times. The k factors for half-filled tubes can be calculated by using an approximate $r_{\max}$ and r_{av} in k factor [EQ 1](#).

Figure 1.4 Nomogram for J2 Series Centrifuges*



* Align a straightedge through known values in two columns; read the figure where the straightedge intersects the third column.

Figure 1.5 Nomogram for J6Series Centrifuges*



* Align a straightedge through known values in two columns; read the figure where the straightedge intersects the third column.

Isopycnic Separations

A sedimentation-equilibrium, or isopycnic, method separates particles on the basis of particle buoyant density. Each component in the sample travels through the gradient until it reaches an equilibrium position. Particle velocity due to differences in density is given in the following expression:

$$v = \left[\frac{d^2(\rho_p - \rho_c)}{18\mu} \right] \times g \quad \text{EQ 6}$$

where

v = sedimentation velocity (dr/dt)
 d = particle diameter
 ρ_p = particle density
 ρ_c = solution density
 μ = viscosity of liquid media
 g = standard acceleration of gravity

At equilibrium, $\rho_p - \rho_c$ is zero, and particle velocity is therefore zero.

The gradient may be preformed before the run or generated during centrifugation. For gradients formed by centrifugation, the time it takes to form a gradient depends on the sedimentation and diffusion coefficients of the gradient material, the pathlength, and the rotor speed. For a given gradient material, the shorter the pathlength and the higher the rotor speed, the faster the gradient will form. In general, the time required for gradients to reach equilibrium in swinging-bucket rotors will be longer than in fixed-angle rotors. One way to reduce run times is to use partially filled tubes. Refer to the applicable rotor manual to determine the maximum allowable speed and solution density when using partially filled tubes.

Rate Zonal Separations

Particle separation achieved with rate zonal separation is a function of the particles' sedimentation coefficient (density, size, and shape) and the viscosity of the gradient material. Sucrose is especially useful as a gradient material for rate zonal separation because its physical characteristics are well known and it is readily available. Samples are layered on top of the gradient. Under centrifugal force, particles migrate as zones. Rate zonal separation is time dependent; if the particles are more dense than the most dense portion of the gradient, some or all of the particles will pellet unless the run is stopped at the appropriate time.

A separation is sometimes a combination of rate zonal and isopycnic. Depending on particle buoyant densities and sedimentation coefficients, some particles may be separated by their differential rates of sedimentation, while others may reach their isopycnic point in the gradient.

In most cases, when banding two or three components by rate zonal separation, run times can be shortened considerably if reduced fill levels are used. Tubes are partially filled with gradient, but the sample volume is not changed (however, gradient capacity will be reduced). Thickwall tubes should be used for this technique, since thinwall tubes will collapse if not full.

Blood Component Separations

Centrifugation is the primary method for processing blood because it provides the required high throughput, reproducibility, and versatility. Most blood components can be separated in one or two runs. Generally, two types of runs are performed.

- Soft spin runs, short centrifugation runs (3 to 5 minutes) at low g -forces (2000 to 3000 $\times g$) at ambient temperature, are used to keep small cells or platelets in suspension while the larger cells sediment. This type of run is used to obtain platelet-rich plasma and red blood cell concentrate from whole blood.
- Hard spin runs are longer (5 to 7 minutes), at higher g -forces (4000 to 5000 $\times g$), at ambient temperatures or at 4°C, and are used to separate fresh plasma from cellular components.


Soft spin and hard spin techniques are often combined. Refer to [APPENDIX D](#) for further information about separation of blood components by centrifugation.

General Operating Information

Careful centrifugation technique is essential, because forces generated in high-speed centrifugation can be enormous. For example, 10 grams at the bottom of a JA-25.50 fixed-angle rotor rotating at 25,000 rpm exerts the gravitational equivalent of 0.8 ton of centrifugal mass at the bottom of the tube cavity.

NOTE Specific information about filling, sealing, and capping containers, loading rotors, etc., is contained in later sections.

Rotor Balance




The mass of a properly loaded rotor is evenly distributed on the centrifuge drive hub, causing the rotor to turn smoothly with the drive. An improperly loaded rotor will be unbalanced; consistent running of unbalanced rotors will reduce centrifuge drive life. To balance the rotor load, fill all opposing containers to the same level with liquid of the same density. Weight of opposing containers must be distributed equally. Place tubes in a fixed-angle, vertical- tube, or JS-24 series swinging-bucket rotor symmetrically, as illustrated in [Figure 1.6](#). Detailed information about balancing other swinging-bucket rotors is contained in [CHAPTER 5](#).


If sample quantity is limited and the rotor is not balanced, do one of the following to balance the rotor, depending on the rotor in use:

- Load the opposite rotor cavities or buckets with tubes containing a liquid of the same density as opposing tubes.
- Layer a low-density, immiscible liquid, such as mineral oil, on top of the sample to fill opposing tubes to the same level.

Figure 1.6 Arranging Tubes Symmetrically in a Fixed-Angle, Vertical-Tube, or JS-24 Series Swinging-Bucket Rotor*




Rotor Tie-Down




To secure the rotor to the drive spindle hub during centrifugation, J series rotors are equipped with devices that screw into the hub. If the rotor is left in the centrifuge between runs, tighten the tie-down device before each run.

- Some rotors are equipped with tie-down assemblies. These may be knobs that can be hand-tightened when the rotor is installed, and between runs if the rotor is left in the centrifuge. Other tie-down assemblies are tightened by turning the rotor lid knob.



- Some new and modified rotors have dual-locking lid mechanisms. The dual-locking lid mechanism consists of a daisy knob that secures the lid to the rotor, and a tie-down knob that attaches the rotor to the centrifuge drive hub. (Daisy refers to the knob shape. The grooves between each “petal” let your fingers grip the knob firmly and provide leverage for turning.) The daisy knob allows you to attach the lid to the rotor before placing the rotor into the centrifuge, and to remove the rotor from the centrifuge with the lid attached.



CAUTION

Always loosen the rotor knob before loosening the daisy knob to avoid jamming the knobs.

- Other rotors are secured to the centrifuge drive spindle hub by a tie-down screw. A torque bar is supplied with the rotor to provide leverage to securely fasten the rotor.

* For example, two, three, four, or six tubes can be arranged symmetrically in a six-place rotor.

Overspeed Protection


Rotors are specifically designed to withstand a maximum load (that is, volume and density of the rotor contents) at rated speed. At greater speeds, or at rated speeds with heavier loads, rotors are subject to failure.

- In *Avanti J series centrifuges*, an electronic recognition system identifies the rotor, thereby limiting speed to the rated speed of the rotor.
- In J2 and J6 series centrifuges with *analog controls*, the rotor speed is limited by the physical properties of the rotor. Friction created by the air in the centrifuge chamber interacting with the rotor surfaces during centrifugation in most cases prevents rotors from exceeding their rated speeds.
- In *microprocessor-controlled J2 and J6 series centrifuges*, internal circuitry monitors the rotor speed and prevents a rotor from exceeding its rated speed. The rotor entry code listed in the applicable rotor manual sets the allowable speed.

At rated speeds with heavier loads, rotors are subject to failure. It is the operator's responsibility to limit rotor speed when centrifuging dense solutions or when using heavy containers; refer to [Allowable Run Speeds](#) below.

Allowable Run Speeds

Under some conditions, the maximum allowable speed of the rotor (indicated by the rotor name) must be reduced to ensure that neither the rotor nor the labware are overstressed during centrifugation.



- **Dense Solutions.** When using dense solutions (> 1.2 g/mL) in J2 series rotors, determine maximum run speed using the following square-root reduction formula:

$$\text{reduced run speed} = \text{maximum rated speed} \sqrt{\frac{\rho_A}{\rho_B}} \quad \text{EQ 7}$$

where ρ_A is the maximum permissible density of the tube contents for a particular rotor (from the rotor manual), and ρ_B is the actual density of the tube contents to be centrifuged.

When using dense solutions in J6 series rotors, determine maximum run speed using the following square-root reduction formula:

$$\text{reduced run speed} = \text{maximum rated speed} \sqrt{\frac{A}{B}} \quad \text{EQ 8}$$


where A is 2500 grams for JS rotors or 1500 grams for the JR-3.2 rotor, and B is the weight in grams of a total load (bucket with adapter and sample; bucket with blood bag cup and filled blood bag; tray with racks, tubes, and sample).

NOTE The maximum speed for Avanti J or J2 series rotors in J6 series instruments is 6000 rpm with solutions of density no greater than 2.0 g/mL. *Solutions of density greater than 2.0 g/mL should not be centrifuged.*

- **Critical Speed Range.** The critical speed range of a rotor is the range of speeds in which, during acceleration, the rotor shifts so as to rotate about its center of mass. While passing through this speed range, the rotor will usually vibrate. Do not set operating speeds that are within a rotor's critical speed range (as listed in the rotor manual).
- **Minimum Speeds.** Some buckets or carriers will not achieve their full horizontal position if the rotor is run below minimum rotating speed. Refer to the individual rotor manual for speed requirements.

Temperature Compensation

To ensure that the rotor reaches the required temperature during centrifugation, some temperature compensation may be required because of the mass of these rotors. Tables listing temperature compensation units for various rotors are contained in [APPENDIX B](#) and individual rotor manuals. Follow the instructions below for the model of centrifuge being used.



Avanti J Centrifuges

Avanti J series centrifuges provide automatic temperature compensation. Enter the run temperature according to the instructions in your centrifuge instruction manual. No additional input is required.

Avanti J2 Series Centrifuges (No longer manufactured)

Set temperature compensation in analog J2 model centrifuges (models J2-HS, J2-21, and J2-HC) as follows.

- 1 Turn the **SET** knob on the centrifuge panel to the required sample temperature.



- 2 Find the compensation value in [APPENDIX B](#) (or in the applicable rotor manual) that corresponds with the required temperature and run speed.

Set the **COMP** dial to that setting.

(Interpolate if intermediate values are required.)



NOTE Temperature settings for J-21 series centrifuges must be empirically determined.

Analog J6 Series Centrifuges (No longer manufactured)

Set temperature compensation in analog J6 model centrifuges (models J6-HC and J6-B) as follows.


- 1 Find the compensation value in [APPENDIX B](#) (or in the applicable rotor manual) that corresponds with the required temperature and run speed.

(Interpolate if intermediate values are required.)

- 2 Turn the **SET** knob on the centrifuge control panel to the required sample temperature.

Microprocessor-Controlled Centrifuges

Operating temperatures for most rotors are contained in memory in microprocessor-controlled centrifuges (model J6-MI, and discontinued models J6-MC, J2-MI, J2-21M, J2-MC, and J2-21M/E). Set temperature compensation as follows for rotors not in centrifuge memory.



- 1 Press the **TEMP** key on the centrifuge control panel and then use the keypad to enter the required sample temperature.
- 2 Find the compensation value in [APPENDIX B](#) (or in the applicable rotor manual) that corresponds with the required temperature and run speed.
- 3 Press **COMP ADJ.**
The word “**COMP**” flashes in the **TEMPERATURE** display and the display flashes.
- 4 Use the keypad to enter the compensation value.
Press the $\pm$. key to enter a minus sign; pressing it again will remove the minus sign.
- 5 Check the temperature display.
If the entry is incorrect, press **ce** and reenter the digits.
- 6 When the entry is correct, press **ENTER/RECALL**.

Tubes, Bottles, and Accessories

Introduction

*This section describes various labware used in Beckman Coulter J series rotors. General instructions for using containers follow in [CHAPTER 3](#). Care and maintenance instructions are in [CHAPTER 7](#). General rotor use instructions are in [CHAPTER 4](#), [CHAPTER 5](#), and [CHAPTER 6](#). The individual rotor manual that comes with each rotor provides specific instructions on the tubes, bottles, and accessories that can be used in a particular rotor. *A table of chemical resistances can be found in [APPENDIX A](#) of this manual.*

Labware Selection Criteria

No single tube or bottle design or material meets all application requirements. Labware choice is usually based on a number of factors.

- The centrifugation technique to be used, including the rotor in use, quantity of sample to be centrifuged, need for sterilization, importance of band visibility, and so forth
- Chemical resistance—the nature of the sample and any solvent or gradient media
- Temperature and speed considerations
- Whether tubes or bottles are to be reused

[Table 2.1](#) contains an overview of some of the characteristics of tube and bottle materials.

* A complete list of tubes, bottles, and accessories is provided in the latest edition of the Beckman Coulter *High Performance, High Speed, High Capacity Rotors, Tubes, and Accessories* catalog (BR-8102), available at www.beckmancoulter.com.

Table 2.1 Characteristics and Chemical Resistances of Tube and Bottle Materials^a

Tube or Bottle Type	Optical Property	Puncturable	Sliceable	Reusable	Acids (dilute or weak)	Acids (strong)	Alcohols (aliphatic)	Aldehydes	Bases	Esters	Hydrocarbons (aliphatic)	Hydrocarbons (aromatic and halogenated)	Ketones	Oxidizing Agents (strong)	Salts
thinwall polypropylene	transparent	yes	yes	no	S	U	U	M	S	U	U	U	U	U	S
thickwall polypropylene	translucent	no	no ^b	yes	S	S	S	M	S	M	M	U	M	U	S
Ultra-Clear	transparent	yes	yes	no	S	U	U	S	U	U	U	U	U	U	M
polycarbonate	transparent	no	no	yes	M	U	U	M	U	U	U	U	U	M	M
polypropylene	translucent/ transparent	no	no ^b	yes	S	S	S	M	S	M	S	M	M	M	S
polyethylene	transparent/ translucent	yes	no	yes	S	S	S	S	S	S	U	M	M	M	S
cellulose propionate	transparent	no	no ^b	no	S	U	U	U	U	M	S	S	U	M	S
stainless steel	opaque	no	no	yes	S	U	S	S	M	S	S	S	M	S	M

S = satisfactory resistance

M = marginal resistance

U = unsatisfactory resistance

a. Refer to Appendix A for information about specific solutions.

b. Polypropylene and cellulose propionate tubes with diameters of 5 to 13 mm may be sliced using the Centritube Slicer (part number 347960) and appropriate adapter plate.

NOTE This information has been consolidated from a number of sources and is provided *only* as a guide to the selection of tube or bottle materials. Soak tests at 1 *g* (at 20°C) established the data for most of the materials; reactions may vary under the stress of centrifugation, or with extended contact or temperature variations. To prevent failure and loss of valuable sample, **ALWAYS TEST SOLUTIONS UNDER OPERATING CONDITIONS BEFORE USE.**



WARNING

Do not use flammable substances in or near operating centrifuges.

Labware Material Compatibility with Solvents and Sample

The chemical compatibility of tube or bottle materials with the gradient-forming medium or other chemicals in the solution is an important consideration. Although neutral sucrose and salt solutions cause no problems, alkaline solutions cannot be used in Ultra-Clear tubes or in polycarbonate tubes and bottles. Polycarbonate and Ultra-Clear tubes are incompatible with DMSO, sometimes used in the preparation of sucrose gradients for sedimentation of denatured DNA.

Gradient Formation and Fractionation

Consideration should be given to gradient formation and fractionation when choosing a tube for a density gradient run. If the bands or zones formed during centrifugation are indistinct, they may not be visible through a translucent material such as polypropylene. If optimum band visualization is important, Ultra-Clear, polycarbonate, or cellulose propionate tubes should be used. Whenever collection of bands or zones must be done by slicing or puncturing the tube, a thin, flexible tube wall is required. Ultra-Clear or polypropylene tubes should be used in these cases, depending on the need for transparency.

Labware Types

NOTE Tubes made of cellulose nitrate were formerly popular for various separations, particularly rate-zonal separations. Beckman Coulter discontinued the use of cellulose nitrate for tube manufacture in 1980, due to inconsistent physical properties inherent in the material. If you currently have cellulose nitrate tubes, dispose of them. Consult your laboratory safety officer for proper disposal procedures.

Polypropylene Tubes

Polypropylene tubes are translucent or transparent in appearance, depending on wall thickness, and are non-wettable (although some polypropylene tubes can be chemically treated to make them wettable). Polypropylene tubes are reusable unless deformed during centrifugation or autoclaving. Polypropylene tubes have good tolerance to gradient media, including alkalines. They are satisfactory for many acids, bases, alcohols, DMSO, and some organic solvents. They can be used with or without caps in fixed-angle rotors. Speed reduction is sometimes required with these tubes if run with less than full volume (refer to your rotor manual). Several types of polypropylene tubes are available.




Open-Top Polypropylene Tubes

Thinwall open-top tubes are used in swinging bucket and fixed-angle rotors. In swinging-bucket rotors, thinwall tubes should be filled to within 2 or 3 mm of the tube top for proper tube support. Caps are usually required in fixed-angle rotors. Thinwall tubes are designed for one-time use and should be discarded after use.

Thickwall open-top tubes offer the convenience of centrifuging partially filled tubes without tube caps in fixed-angle and swinging-bucket rotors. Because the solution reorients during centrifugation, the maximum partial fill volume depends on the tube angle. For greater fill volumes, use tubes with caps. Refer to the applicable rotor manual for fill volumes and speed reduction requirements. Thickwall polypropylene tubes are typically reusable unless deformed during centrifugation or autoclaving.

Quick-Seal Polypropylene Tubes



Heat-sealed Quick-Seal tubes can be used in some fixed-angle rotors and in the JS-24 series rotors; they *must be used* in the JV-20 vertical-tube rotor. Single-use Quick-Seal tubes are a convenient form of sealable tube; they are especially useful for the containment of radioactive or pathogenic samples. There are two Quick-Seal tube designs, dome-top and bell-top.

- The bell-top simplifies removal of materials that float during centrifugation.
- Dome-top tubes hold more volume than their bell-top equivalents.

Detailed information about Quick-Seal tubes is contained in publication IN-181.

Polycarbonate Tubes



Polycarbonate is tough, rigid, nonwetable, and glass-like in appearance. Polycarbonate tubes are reusable and can be used with or without caps in fixed-angle rotors, and at least half full in swinging-bucket rotors. Speed reduction may be required in some rotors if the tubes are not completely filled.

Although polycarbonate tubes may be autoclaved, doing so greatly reduces the usable life of these tubes. Cold sterilization methods are recommended. Washing with alkaline detergents can cause failure. *Crazing*—the appearance of fine cracks in the tube—is the result of stress “relaxation” and can affect tube performance. These cracks will gradually increase in size and depth, becoming more visible. Tubes should be discarded before cracks become large enough for fluid to escape. These tubes have good tolerance to all gradient media except alkalines (pH greater than 8). They are satisfactory for some weak acids, but are unsatisfactory for all bases, alcohol, and other organic solvents.

Polyethylene Tubes

Polyethylene tubes are translucent or transparent and have a good tolerance for use with strong acids and bases. They are reusable but cannot be autoclaved. In swinging-bucket rotors, they are used without caps, and with or without caps in fixed-angle rotors.

Ultra-Clear Tubes



Ultra-Clear tubes, made of a tough thermoplastic, are thinwall and not wettable (but can be made wettable; see [CHAPTER 3](#)). Ultra-Clear tubes are available in two types—open-top and Quick-Seal. They are transparent centrifuge tubes, offering easy location of visible banded samples. Standard straight-wall Ultra-Clear tubes must be filled completely and capped for use in fixed-angle rotors.

Ultra-Clear tubes, which can be used one time only, have good resistance to most weak acids and some weak bases, but are unsatisfactory for DMSO and most organic solvents, including all alcohols. Ultra-Clear tubes should not be autoclaved.

Stainless Steel Tubes



Stainless steel tubes offer excellent resistance to organic solvents and heat, but should not be used with most acids or bases. They offer only marginal resistance to most gradient-forming materials other than sucrose and glycerol. Stainless steel tubes are very strong and can be centrifuged when filled to any level. Because of their weight, however, run speeds must often be reduced (see publication L5-TB-072). Stainless steel tubes can be used indefinitely if they are undamaged and not allowed to corrode. They may be autoclaved after use as long as they are thoroughly dried before storage.

Microfuge Tubes



Microfuge tubes, 1.5-mL tubes with attached caps, are made of clear or colored polypropylene. The tubes are placed in adapters for use in some fixed-angle rotors. They are also used in multitube adapters in the buckets or carriers of swinging-bucket rotors. The number and arrangement of tubes in opposing adapters should be balanced.

Bottles




Bottles are available in polycarbonate (hard and clear), polypropylene (translucent), and Sealed polypropylene or polycarbonate bottles, available for most fixed-angle rotors, have a three-piece liquid-tight cap assembly. Other bottles have screw-on caps. Cap assemblies should always be removed before autoclaving bottles. Bottle selection depends on the rotor in use and the specific application; refer to the applicable rotor manual.

Multiwell Titer Plates


Titer plates can be run in specially designed carriers in some swinging-bucket rotors. Carriers are used by installing them on the pivot pins in place of the buckets normally used with the rotor, or in buckets designed to run plates. Because the plates can break under the stresses of high-speed centrifugation, speed reduction is usually required when running multiwell plates. Multiwell plates are also used in adapters in the rack-type rotor.

96-Well Microtiter Plates




The 96-well plates are manufactured of specially formulated polystyrene. These flat-bottom, nonsterile plates normally hold 300 µL per well of sample and solvent.

Deep-Well Titer Plates (and Caps)




These plates are manufactured of sterile or nonsterile polystyrene or polypropylene. The plates can contain up to 1.2 mL per well of sample and solvent in a single 96-well plate when run uncapped. When used with caps, which come in 96-cap strips, each well accommodates 1.0 mL.

Square-Well Titer Plates



Square-well plates are manufactured of nonsterile polypropylene. The square-well format provides 2 mL per well capacity in each 96-well plate.


Temperature Limits



Each labware material has a specified temperature range. Although some high-speed centrifuges can achieve temperatures as high as 45°C, only certain tube or bottle materials can be run under these conditions. Most containers are made of thermoplastic materials that soften at elevated temperatures. This temperature-induced softening, together with such factors as the centrifugal force, the run duration, the type of rotor, previous run history, and the tube angle, can cause labware to collapse. Therefore, if high-temperature runs—above 25°C—are required, it is best to pretest labware under the actual experimental conditions, using buffer or gradient of similar density rather than a valuable sample. (Stainless steel tubes can be centrifuged at any temperature.)

- Plastic labware has been centrifuge tested for use at temperatures between 2 and 25°C. For centrifugation at other temperatures, pretest tubes under anticipated run conditions.
- If plastic containers are frozen before use, make sure that they are thawed to at least 2°C prior to centrifugation.

Spacers and Floating Spacers



Quick-Seal tubes require spacers made of anodized aluminum, with or without floating spacers. The particular combination depends on the type of rotor being used and the tube size.

- In swinging-bucket and fixed-angle rotors, the top of the tube must be supported.
- In vertical-tube rotors, the entire cavity must be filled.


Plastic spacers have been tested for centrifugation between 2 and 25°C. If spacers are centrifuged at temperatures significantly greater than 25°C, deformation of the spacer and tube may occur.

Adapters

Many rotors can accommodate a variety of tube sizes by using adapters that line the tube cavity or bucket. Adapters are fabricated of several different kinds of materials, depending on the rotor and the tube to be used in them. Some of the common materials are acetal, modified polyphenylene oxide, polyetherimide (PEI), polyethylene, rubber, polypropylene, and glass-filled or foamed polypropylene.

Tubes or bottles used with adapters can be filled (and capped, if applicable), according to the type of tube and the design of the rotor being used.

Bottle Adapters




Bottles are often supported during centrifugation in bottle adapters that fit inside the rotor buckets or cavities. The adapters are usually ribbed for strength and support a variety of bottle sizes.

To prevent the bottles from stretching or breaking, a plastic sleeve, or adapter, must be used around each Beckman Coulter 1-liter bottle during centrifugation in J6 series rotors. In other rotors, if the bottles fit snugly in the buckets, the adapters are not required. (Refer to the applicable rotor manual.)

Multitube Adapters



Adapters are used to enable centrifugation of multiple tubes in the bucket of a swinging-bucket rotor or in a fixed-angle tube cavity.

Solid Multitube Adapters



These solid adapters, available in several tube configurations, are made of polyetherimide (PEI), modified polyphenylene oxide, polypropylene, or aluminum that is anodized for corrosion protection. They can be filled and loaded into rotor buckets or cavities without any preparation. These adapters can also be used as tube racks in the laboratory.

Modular Disk Adapters



These adapters can also be used as tube racks in the laboratory. The adapter disks are color-coded by the tube size they accommodate; the number of disks used in an adapter assembly depends upon the length of tubes used. Refer to the applicable rotor manual to determine the kind of adapter required for the tubes you are using. A tube decanter is available to hold tubes securely in some adapters, allowing all tubes to be decanted at once.

Do not intermix modular adapters (or their individual parts) from Beckman Coulter's J6 series rotors with those for the JS-4.3 rotor. While the adapters are similar in appearance, they have very different weights. J6 adapters have bails (vertical supports) that are curved at the top; bails for the JS-4.3 adapters are straight. Keep J6 and JS-4.3 disks and bases separate from each other—mixing them can cause imbalance. In addition, the J6 adapter bails will interfere with the JS-4.3 rotor yoke when the buckets swing up to horizontal position.

Bottle and Tube Caps


The need for caps depends on such factors as the kind of rotor being used, the type of container, and the amount of sample being centrifuged.

Some tubes must be capped before centrifugation, as in the case of thinwall tubes. The thickness and strength of some containers, such as thickwall plastic and stainless steel tubes used in fixed-angle and swinging-bucket rotors, allows them to be run without caps, but they must be only partially filled. (Refer to the applicable rotor manual for allowable capless fill levels.) When greater fill volumes are required in these tubes, caps must be used for sample retention.

When closed containers are required, several choices are available:


- Cap assemblies—threaded caps with inserts and O-rings, or one-piece caps with O-rings, that provide a leakproof closure to accommodate a capacity container load (that is, to the bottom of the insert).
- Threaded caps without inserts or O-rings—these are not as liquid-tight as the cap assemblies; therefore, the meniscus must be kept lower to prevent leakage.
- Snap-on caps—these caps are simple to use but are not as liquid-tight as the cap assemblies or threaded caps. They require an even lower meniscus to prevent leakage.

Aerosolve Canisters



Aerosolve canisters, used in the JS-4.3 swinging-bucket rotor, are designed to minimize aerosol leakage and liquid spills. The canister is transparent, enabling you to see broken labware and take proper precautions before opening the canister.

The canister and lid are made of polyphenylsulfone, tube racks are made of polypropylene, and the O-ring is ethylene-propylene rubber. Refer to [APPENDIX A](#), Chemical Resistances, to determine compatibilities with specific chemicals.




Each canister can hold a variety of tube sizes in tube racks that are specifically designed to fit in the canisters. The canister can also be used as a 500-mL wide-mouth bottle.

WARNING

When centrifuging hazardous materials, always open canisters in an appropriate blood or biological safety cabinet.


Blood Bag Cups




Polypropylene blood bag cups are available for use in swinging-bucket rotors to obtain cell-free plasma for cell packing or for leukolyte depletion. Different sizes of cups are available to accommodate single, double, triple, or quad pack blood bags. Refer to the applicable rotor manual to determine the correct blood bag cup to use. Blood bag cups are autoclavable.

Blood bags should be loaded into the cups outside of the centrifuge to avoid tripping the centrifuge imbalance monitor during loading.

Rotor Labware Assemblies



The JS-5.0 labware assembly has an available HarvestLine System liner. If liners are not used, the sample can be loaded directly into the cup and a partition can be inserted to minimize sample disturbance at low g forces. The gasket and the cup and cover surfaces that contact the gasket must be dry to ensure sealing. Gasket 36926 (green) is used when liners are used; gasket 369257 (red) is used when the cup is used alone, with or without a partition. The cup cover top surface can be written on to identify the assembly or sample.



The HarvestLine System for the JLA-8.1000 and JLA-9.1000 rotors provides a convenient method of loading, recovering, and storing samples run in these rotors. Up to six rotor bottles are placed in the filling rack, and a liner is placed into each bottle. The liners are loaded with sample through a funnel or fermentor hose. The valve in the neck of each liner is then sealed and the liner necks folded to fit inside the bottles. The bottles are sealed with rotor plugs and cap/closures, and the sealed bottles are placed into the rotor canisters for centrifugation. After centrifugation, the liner valves are cut off and the supernatant decanted, either for storage or disposal. The liners can then be heat-sealed for pellet storage or disposal.

Tubes, Bottles, and Accessories


Introduction

*This section contains general instructions for filling and capping the labware used in Beckman Coulter J series rotors, for selecting and using the appropriate accessories, and for recovering samples after a run. Individual rotor manuals provide specific instructions on tubes, bottles, and accessories that can be used in a particular rotor.**

Rotor use instructions are in [CHAPTER 4](#) for fixed-angle rotors, in [CHAPTER 5](#) for swinging-bucket rotors, and in [CHAPTER 6](#) for vertical-tube and rack-type rotors. A table of chemical resistances is in [APPENDIX A](#) of this manual. Reference information on some commonly used gradient materials is in [APPENDIX C](#).

Gradient Preparation

Many commercial gradient formers are available. These devices usually load a tube by allowing the gradient solutions to run down the side of the tube. The heaviest concentration is loaded first, followed by successively lighter concentrations. This method is acceptable for wettable tubes; however, loading a nonwetable tube (such as Ultra-Clear, polypropylene, and polycarbonate) by allowing solutions to run down the side of the tube can cause mixing.




Gradients in nonwetable tubes can be prepared using a gradient former by placing a long syringe needle or tubing to the tube bottom and reversing the gradient chambers. In that way the lightest gradient concentration is loaded first, underlayed by increasingly heavier concentrations.

* A complete list of tubes, bottles, and adapters is provided in the latest edition of the Beckman Coulter *High Performance, High Speed, High Capacity Rotors, Tubes & Accessories* catalog (BR-8102), available at www.beckmancoulter.com.




You can also prepare preformed step gradients by hand, using a pipette. Carefully layer solutions of decreasing concentration by placing the tip of the pipette at the angle formed by the tube wall and the meniscus, or float the lighter gradient concentrations up by adding increased density solutions to the tube bottom using a hypodermic syringe with a long needle such as a pipetting needle.



Another way to form a linear gradient is to allow a step gradient to diffuse to linearity. Depending on the concentration differential between steps and the cross-sectional area, allow 3 to 6 hours for diffusion at room temperature, and about 16 hours at 0 to 4°C. For diffusion of step gradient in Quick-Seal and capped straightwall tubes, slowly lay the tube on its side (tube contents will not spill, but make sure the tube does not roll). After 2 hours at room temperature, slowly set the tube upright.

Once the gradient is prepared, layer the sample on top of the gradient.



For *thinwall* tubes only partially filled with gradient, add a buffer solution to fill the tube to provide tube wall support. Although the gradient volume is reduced, sample volume is not changed.

NOTE If a partially filled *thickwall* tube is centrifuged, the tube does not require liquid support, and therefore, the buffer solution is not required.

General Filling and Sealing or Capping Requirements

See [Table 3.1](#) for general filling and sealing or capping requirements for tubes and bottles used in J series rotors. Maximum fill volume includes sample and gradient. Refer to individual rotor manuals for specific filling and capping requirements.

Table 3.1 General Filling and Sealing Requirements for Tubes and Bottles

Tube or Bottle	Filling Level Requirements		
	Swinging-Bucket Rotors	Fixed-Angle Rotors	Vertical-Tube Rotors
Polypropylene thinwall tubes thickwall tubes Quick-Seal tubes bottles	within 2 to 3 mm of top at least 1/2 full full and heat sealed min to max (see rotor manual) with screw-on cap or cap assembly	full with cap 1/2 full to max capless level or full with cap full and heat sealed 1/2 full to max (see rotor manual) with screw-on cap or cap assembly	not used not used full and heat sealed not used
Ultra-Clear open-top tubes Quick-Seal tubes	within 2 to 3 mm of top not used	full with cap full and heat sealed	not used full and heat sealed
Polycarbonate thickwall tubes bottles	at least 1/2 full at least 1/2 full	1/2 full to max capless level or full with cap min to max (see rotor manual) with screw-on cap or cap assembly	not used not used
Stainless Steel tubes	any level	any level with cap or cap assembly	not used
Polypropylene tubes and bottles	at least 1/2 full	1/2 full to max capless level or full with cap or cap assembly	not used
Polyethylene tubes	at least 1/2 full	1/2 full to max capless level or full with cap	not used
Cellulose Propionate tubes and bottles	at least 1/2 full	1/2 full to max capless level	not used

Working with Physiological Fluids

WARNING

Handle body fluids with care because they can transmit disease. No known test offers complete assurance that they are free of micro-organisms. Some of the most virulent—Hepatitis (B and C) and HIV (I–V) viruses, atypical mycobacteria, and certain systemic fungi—further emphasize the need for aerosol protection. Handle other infectious samples according to good laboratory procedures and methods to prevent spread of disease. Because spills may generate aerosols, observe proper safety precautions for aerosol containment. Do not run toxic, pathogenic, or radioactive materials in this rotor without taking appropriate safety precautions. Biosafe containment should be used when Risk Group II materials (as identified in the World Health Organization *Laboratory Biosafety Manual*) are handled; materials of a higher group require more than one level of protection.



When working with potentially hazardous materials, always fill or open containers in an appropriate hood or biological safety cabinet. Three levels of containment are offered by Beckman Coulter, and may be used singly or combined, depending upon your application.


1. Capped tubes or bottles are designed to provide fluid containment. We strongly recommend that all containers carrying physiological fluids be capped to prevent leakage.
2. Rotor or bucket covers are designed to minimize the possibility of fluid leakage during centrifugation.
 - Bucket covers for swinging bucket rotors help to contain fluids within the bucket in the event of tube breakage or blood-bag failure.
 - Some fixed-angle rotors have available dual-locking lid mechanisms that provide added biosafety by allowing the rotor to be loaded into and removed from the centrifuge with the lid in place. The rotor may be placed under a safety hood before the lid is attached or removed.
3. Aerosolve canisters are designed to minimize the possibility of aerosol (and fluid) leakage during centrifugation.

Filling Open-Top Tubes

Open-Top Polypropylene Tubes

Open-top polypropylene tubes are used in swinging bucket and fixed-angle rotors.


Swinging-Bucket Rotors



Fill all opposing tubes to the same level.

- *Thinwall Tubes*—Fill to within 2 or 3 mm of the top for proper tube wall support.
- *Thickwall Tubes*—Fill at least half full.

Fixed-Angle Rotors



Fill all opposing tubes to the same level.

- *Thinwall Tubes*—Must be completely filled; liquid and cap for support of the tube wall is critical.
- *Thickwall Tubes*—Can be partially filled and centrifuged as indicated in the applicable rotor manual. Speed reductions may be required for these partially filled tubes. For greater fill volumes and faster speeds, tube caps should be used. Refer to the rotor manual for fill volumes and speed limitations.

Other Open-Top Tubes

Open-top tubes of other materials can also be used in fixed-angle and swinging-bucket rotors. (Vertical-tube rotors use only Quick-Seal tubes.) Fill these tubes as indicated below.

Polycarbonate

Thickwall polycarbonate tubes can be centrifuged partially filled. Observe maximum rotor speeds and fill volumes listed in the applicable rotor manual.

Ultra-Clear

For *swinging-bucket* rotors, fill to within 2 or 3 mm of the top of the tube. Refer to the applicable rotor manual.

Polypropylene

Fill all opposing tubes to the same level.

- For *swinging-bucket* rotors, fill to within 2 or 3 mm of the top of the tube.
- Fill thickwall polypropylene tubes at least half full to maximum level in *fixed-angle* rotors. Speed reduction is required. Refer to the applicable rotor manual.

Polyethylene

For *swinging bucket* and *fixed-angle* rotors, fill these tubes from half full to maximum level. Refer to the applicable rotor manual.

Stainless Steel


Because of their strength, stainless steel tubes can be centrifuged while filled to any level (with all opposing tubes filled to the same level). However, run speeds must be reduced due to their weight. The criteria for speed reduction depends on the tube-cap material and the strength of the rotor being used. Refer to the applicable rotor manual or *Run Speeds for Stainless Steel Tubes* (publication L5-TB-072) for correct run speeds.

Capping Tubes


Caps must be used with thinwall polypropylene and Ultra-Clear tubes in fixed-angle rotors. To prevent spillage, thickwall polypropylene, polycarbonate, and stainless steel tubes must be capped when fill levels exceed the maximum level for uncapped tubes as listed in the applicable rotor manual.

Cap requirements depend on the tube or bottle material, diameter, and wall thickness, as well as on the rotor. The applicable rotor manual specifies which cap should be used with a particular tube or bottle; use of the wrong cap could cause a rotor mishap.


When closed containers are required, several choices are available:



- Cap assemblies—threaded caps with inserts and O-rings that provide a leakproof closure to accommodate a capacity container load (that is, to the bottom of the insert). Single-piece cap assemblies have the insert permanently attached.



- Threaded caps without inserts or O-rings—these are not as liquid-tight as the cap assemblies; therefore, the meniscus must be kept lower to prevent leakage. Speed reductions may also be required with lower fill volumes.



- Snap-on caps—these caps are simple to use but are not as liquid-tight as the cap assemblies or threaded caps. They require an even lower meniscus to prevent leakage.


Filling and Capping Tubes

To prevent spillage and provide support, polycarbonate and polypropylene bottles used in fixed-angle rotors must be capped when fill levels exceed the maximum level allowed for uncapped bottles. Bottles should be filled to maximum fill levels when spun at full rated speeds. Unless specified otherwise, the minimum recommended volume for bottles is half full; this will require reduced rotor speed for optimum labware performance. Refer to the applicable rotor manual for bottle fill levels and cap requirements.

Three-Piece Assemblies


Cap bottles with three-piece cap assemblies as follows:

- 1 Be sure the O-ring, plug, and bottle lip are dry and free of lubrication.
- 2 Place the O-ring on the underside of the plug.
- 3 Insert the plug into the neck of the bottle, ensuring that no fluid contacts the O-ring.



- 4 Tighten the cap by hand.

JLA-8.1000 and JLA-9.1000 Bottle Cap/Closure



Place the plug on the bottle, then screw on the cap/closure by hand as tightly as possible. Tighten until the scribe mark on the cap/closure is aligned with or goes past the vent line on the bottle.


Filling and Loading Cups in the JS-5.0 Rotor

NOTE Four labware cups must be used for every run and must be balanced to within 25 grams of each other. *Do not load the rotor with two filled cups and two empty cups.*

-
- 1** Insert four labware cups into two cup racks with the cup latch hinges toward the center of the racks.
-
- 2** Make sure that the gaskets and sealing surfaces on each cup and cover are clean and dry.
 - a.** Place a gasket around the top edge of each cup, carefully pushing the gasket down until it is fully seated on the cup.
 - b.** Use green gaskets (369261) if you are using liners.
 - c.** Use red gaskets (369257) if you are using cups alone, with or without partitions.
-
- 3** Place a liner in each cup (if applicable).
-
- 4** You may fill the cups now, or close the lid and fill through the spout.

NOTE If liners are not used, partitions (369259) may be inserted into the slots inside the cups. Remove the red cup gaskets (369257) before inserting partitions, and be sure to reinstall the gaskets.
-
- 5** Place a cover on each cup and fasten the latch securely.
 - a.** If the latch will not fasten, check to make sure that the gasket is properly installed.
The latch cannot be fastened if the gasket is not fully seated.
 - b.** *Be sure that the latch is fastened before lifting the cup by the handle.*

- 6** If the cups were not filled previously, load sample into each cup through the cover spout using a funnel, tubing (1.27-cm [$1/2$ -in.] O.D.), or a pipette.
- Use the fill line indicators to assist in filling all four cups to the same level.
All four cups must balance to within 25 grams of each other.
 - When loading is complete, snap a plug into place in each cover spout.



Filling and Sealing Quick-Seal Tubes


Fill each tube to the base of the neck, using a syringe with a 13-gauge or smaller needle.* A small air space (no larger than 3 mm) may be left, but an air bubble that is too large can cause the tube to deform, disrupting gradients or sample. Spacer and/or floating spacer requirements for Quick-Seal tubes are described in the individual rotor manuals. The neck of the tube should be clean and dry before sealing.

There are two tube sealers for use with Quick-Seal tubes—the hand-held Cordless Tube Topper, and the older tabletop model (no longer available). Refer to *How to Use Quick-Seal Tubes with the Beckman Coulter Cordless Tube Topper* (publication IN-181) for detailed information about the Tube Topper. Instructions for using the older tabletop tube sealer are in *How to Use Quick-Seal Tubes with the Beckman Tube Sealer* (publication IN-163).

Quick-Seal tubes are heat-sealed quickly and easily using the Beckman Coulter Cordless Tube Topper (see [Figure 3.1](#)). The following procedures provide the two methods for heat-sealing Quick-Seal tubes using the hand-held Tube Topper. Use the applicable tube rack listed in the rotor manual.

* A sample application block (part number 342694) is available for holding and compressing tubes, and can be used to layer samples on preformed gradients in polypropylene Quick-Seal tubes.

Figure 3.1 The Cordless Quick-Seal Tube Topper



⚠ CAUTION


Before plugging in the Tube Topper, be sure that you have a proper power source (120 V, 50 or 60 Hz). Charge your Cordless Tube Topper only in the charging stand supplied with it.

- 1 Remove the Tube Topper from the charging stand.
 - a. Leave the pushbutton turned to LOCK position.
 - b. Insert the ends of the Tube Topper tip into the two openings of the copper strips at the end of the Tube Topper device.

⚠ WARNING

Touching the heated tip of the Tube Topper will cause burns. When the pushbutton is pressed, the tip heats almost immediately. Make sure the pushbutton is turned to LOCK position *unless you are actually sealing a tube.*

- 2 Place a seal former on each tube stem. (The polytetrafluorethylene coating on the seal formers is permanent.
Do not scratch the interior of the formers, as you may damage this coating.)




- 3** Seal each tube using Method A (With the Seal Guide) or Method B (Without the Seal Guide).
Method A is preferable when sealing smaller tubes or when resealing a tube that leaks.

CAUTION


Always keep the Tube Topper in its charging stand when not in use. Do not lay the unit against any surface after use until the tip has cooled (3 to 5 minutes after shut off).

Method A — With the Seal Guide




- a. Place a seal guide (with the flat side down) over the seal former.
- b. Turn the Tube Topper pushbutton to USE position. Press the pushbutton and wait 3 to 5 seconds for the tip to heat.
- c. Apply the tip of the Tube Topper vertically to the seal former. Press down gently for about 10 seconds. The seal guide should move down the tube stem until it rests on the tube shoulder. Using the seal guide prevents the seal former from being pressed into the tube shoulder.

NOTE. Always apply the tip of the Tube Topper vertically to the seal former. Apply gentle pressure when sealing the tube.




- d. When the seal guide has moved to the correct position, remove the Tube Topper and pinch the circular seal guide to hold the seal former in place.
- e. Place the heat sink (small end) over the cap for 2 to 3 seconds while the plastic cools — do NOT let the seal former pop up. (If the seal former does pop up, the tube may not have an adequate seal and may need to be resealed.)
- f. Remove the heat sink and seal guide. When the seal former cools, remove it by hand or with the removal tool (361668). Save the seal guide and former for future use.

Method B — Without the Seal Guide



NOTE Always apply the tip of the Tube Topper vertically to the seal former. Apply gentle pressure when sealing the tube.

- a. Turn the Tube Topper pushbutton to USE position. Press the pushbutton and wait 3 to 5 seconds for the tip to heat.
- b. Apply the tip of the Tube Topper vertically to the seal former. The seal former should move down the tube stem until it just rests on the tube shoulder. Be careful NOT to press the seal former into the tube shoulder; it may cause the tube to leak.



NOTE It is very important to apply the heat sink immediately. To do so, we recommend that you have it in one hand, ready to apply as soon as needed.

- c. Remove the Tube Topper. IMMEDIATELY place the large end of the heat sink over the seal former. Hold it there for a few seconds while the plastic cools—do NOT let the seal former pop up. (If the seal former does pop up, the tube may not have an adequate seal and may need to be resealed.)
- d. Remove the heat sink. When the seal former cools, remove it by hand or with the removal tool (361668).

- 4 After completing either heat-sealing method, squeeze the tube gently (if the tube contents may be disturbed) to test the seal for leaks. If the tube does leak, try resealing it using Method A.



- 5 The tube is now ready for centrifugation. Seal the remaining tubes.


- 6 Return the Tube Topper to its charging stand when finished.

Capping Multiwell Titer Plates

Multiwell titer plates—regular, deep-well, and square-well—can be run uncovered or using one of the available cover types.


Cap Strips

Available sterile or nonsterile 96-cap strips can be used with deep-well plates. (When these caps are used, the capacity of each well is reduced to 1.0 mL.)



Aluminum Foil Lids

Seal & Sample aluminum foil lids (538619) have a bioinert adhesive backing, enabling complete plate sealing. The lids are presized for multiwell, deep-well, and square-well plates, and cause no reduction in well capacity. A 4-inch soft-rubber roller (538618) is required to ensure secure sealing of the foil lids



Using Adapters


Tubes and bottles used with adapters can be filled (and capped, if applicable) according to the type of container and the design of the rotor being used.

Using Solid Multitube Adapters

The solid adapters, available in several tube configurations, can be filled and loaded into rotor buckets or cavities without any preparation. They can also be used as tube racks in the laboratory.


Using Modular Disk Multitube Adapters

These adapters can also be used as tube racks in the laboratory. The adapter disks are color-coded by the tube size they accommodate; the number of disks used in an adapter assembly depends upon the length of tubes used. Refer to the applicable rotor manual to determine the kind of adapter required for the tubes you are using.



Assemble modular disk adapters as follows.

- 1** Select the appropriate adapter base and attach a bail to it.
- 2** Place the base and bail in an empty bucket or on the lab bench (not in the rotor).
- 3** Position one of the disks so that its grooves are aligned with the bail.
Push the disk down until the bail snaps into the grooves.
- 4** Add more disks until the height of the assembly is nearly as tall as the tubes you will be using.
(If the height of the disks is very tall, you may have to push the bail into the grooves of the top disks by hand.)
Remove or add disks to the bail to accommodate shorter or longer tubes.
If the tubes fit too snugly in the adapter's rubber base, apply a light film of dusting power, such as talcum powder, to prevent the tubes from sticking.




Using Aerosolve Canisters


Aerosolve canisters can be used in the JS-4.3 rotor to minimize aerosol leakage and liquid spills from rotor buckets during centrifugation. Each canister can hold a variety of tube sizes in tube racks that are specifically designed to fit in the canisters. The canister can also be used as a 500-mL wide-mouth bottle.

 WARNING

When centrifuging hazardous materials, always open canisters in an appropriate hood or biological safety cabinet.




- 1 Inspect canister assemblies before use.
Do not use damaged components.
- 2 Before placing the canister in a bucket, remove the bucket-cover O-ring seated on the ledge inside the bucket.
If this O-ring is not removed, a vacuum will be created between the bucket and canister that will make removing the canister from the bucket difficult.



NOTE Do not run chloroformed samples in Aerosolve canisters. Chloroform vapors can damage the canister material.


- 3 Fill the canister as described under *Using Canisters as Wide-Mouth Bottles*, or *Using Canisters with Tube Racks*, below.
- 4 Screw the lid on until closing resistance is first felt, then tighten it an additional 60 degrees.
The scribe marks around the rim of the canister and the corrugated finger grips on the lid are all placed 60 degrees apart.



To tighten, turn the lid from (A) to (B).
Tightening down the lid more than this will make it difficult to remove.

Using Canisters as Wide-Mouth Bottles

- 1 Fill each canister only to the fill-level line (maximum is 500 mL of 1.2 g/mL liquid).




- 2 Run another canister, filled to the same level with liquid of the same density, in the opposite bucket.

Using Canisters with Tube Racks

The racks designed to hold tubes in the Aerosolve canister are listed in [Table 3.2](#). Tube racks are easily disassembled by unscrewing the handle and lifting off the top plate.

- 1 Press a rubber cushion (if applicable—see [Table 3.2](#)) into each tube hole in the rack base.



Aerosolve
Tube Rack

- 2 Load filled tubes symmetrically into tube racks.
 - Opposing loads should weigh about the same, within 10 grams.
 - Do not exceed the rated maximum load for each bucket (1000 grams).
 - Maximum bucket load includes the bucket, cushion (if applicable), canister, tube rack, tubes, and sample.

NOTE Partially filled tube racks should contain the same number of balanced tubes. Each tube in a rack must be balanced by a tube in a diametrically opposed position in the opposite rack.

Table 3.2 Aerosolve Tube Rack

Rack Color	Nominal Tube Volume (mL)	Nominal Tube Diameter (mm)	Maximum Number Tubes per Adapter	Part Number		Tube Cushion ^a Part Number
				(set of four racks)	(set of two racks)	
white	1.5	11	24	—	354495	none
blue	3 & 5	12	24	359160	359482	344117
tan	5	13	24	358993	359489	none
orange	10	14	18	359161	359483	344118
purple	12 3 & 5	16 12	12 6	359162	359484	344119
white (vials)	15	14	10	—	344517	none
green	15 & 20 3 & 5	18 12	12 6 ^b	359163	359485	344120

Table 3.2 Aerosolve Tube Rack (*Continued*)

Rack Color	Nominal Tube Volume (mL)	Nominal Tube Diameter (mm)	Maximum Number Tubes per Adapter	Part Number		Tube Cushion ^a Part Number
				(set of four racks)	(set of two racks)	
light green (conical)	15 3 & 5	17 12	6 6	358991	359487	none
lime green (conical)	50 3 & 5	30 12	4 4	358992	359488	none
yellow	50 3 & 5	29 12	4 4	359164	359486	344121

- a. These cushions are inserted into the tube holes in the base of the rack. An additional pad (part number 349948), inserted in the rotor bucket beneath the entire canister, is also needed.
- b. If using 15-mL Vacutainers, only four may be loaded into this tube rack (the two outer positions are restricted by the cover height). Vacutainer is a registered trademark of Beckton, Dickinson and Company.

Using Blood Bag Cups

WARNING

Ask your laboratory safety officer to advise you about the level of containment required for your application and about the proper decontamination or sterilization procedures to follow if fluids escape from containers.


Different cups, color-coded for capacity identification, can accommodate single, double, triple, or quad pack blood bags. Refer to the applicable rotor manual to determine the correct blood bag cup to use. *Do not pour liquid directly into blood bag cups. Fit blood bags into cups before loading the cups into the rotor buckets.* Stuffing blood bags directly into the rotor while it is installed in the centrifuge can trip the imbalance detector.

- 1 Load the bags as far down into the cups as possible.

Make sure the bags stay as vertical as possible, with no folds at the top or corners.

If folds are present, blood cells could remain in the folds and then mix with the plasma when the bag is removed.

- 2 Sandwich the tubing between the blood bag and any transfer packs.



- 3 Make sure the loaded blood bag cups opposite each other on the rotor yoke are approximately the same weight (within 1 gram).
(Balancing pads can be used with some rotors, if necessary, to maintain weight balance.)

Load blood bag cups into the rotor buckets. To reduce the possibility of bag breakage, align the blood bag seam with the rotor pivot pins with the label facing out (away from the axis of rotation).

Sample Recovery

CAUTION

If disassembly reveals evidence of leakage, you should assume that some fluid escaped the container or rotor. Apply appropriate decontamination procedures to the centrifuge, rotor, and accessories.

You can recover labware from most J series rotors while the rotor or yoke remains in the centrifuge.

- Rotor buckets or carriers can be removed from the rotor yoke, then unloaded on a lab bench or table or under a protective hood. Blood bags must *always* be removed from blood bag cups outside of the centrifuge.
- You can remove the lid from most fixed-angle rotors and extract the tubes or bottles using a removal tool (specified in the applicable rotor manual).

NOTE Vertical-tube rotors cannot be unloaded inside the centrifuge. The rotor must be removed from the centrifuge and placed in a rotor vise to loosen the tube cavity plugs.

Sample recovery depends on the type of labware used, the component(s) isolated, and the analysis desired. The Beckman Universal Fraction Recovery System (343890) can be useful when recovering sample from tubes (see publication L5-TB-081).

Capped Tubes

The usual methods of recovering supernatants or pellets include decanting or withdrawing the gradient and scraping pellets from the tube bottom.



- 1 Remove tube caps carefully to avoid sample mixing.
- 2 If tubes will be reused, scrape pellets out with a plastic or wooden tool; scratches on tube interiors caused by abrasive or sharply pointed tools can result in tube failure during subsequent runs.

JS-5.0 Cups


- 1 Remove the rotor lid and hang it on the black rubber block on the inside of the centrifuge door.
- 2 Remove the plug from the labware cup cover and pour the supernatant out of the cup through the spout.
Or, remove the cup cover and pour the supernatant over the cup edge.
- 3 If a liner was used, remove the liner from the cup.
Fold or heat seal the liner* and store the pellet as required.
- 4 If a liner was not used, first remove the red cup gasket (369257), remove the partition (if used), and then use the spatula (367891) to remove pellet from the cup.
Do not use a metal tool to remove pellet, as metal could damage the cup and shorten its useful life.




* Beckman Coulter recommends Cole-Parmer heat sealer Model U-03018-10, adjusted to setting 3 or 4. Contact Cole-Parmer at (800) 323-4340, by Fax at (847) 247-2929, or at www.coleparmer.com.

Quick-Seal Tubes


There are several methods of recovering fractions from Quick-Seal tubes. One of the following procedures may be used.




NOTE If you plan to collect particles from the tube side or bottom, first create an air passage by snipping the stem or inserting a hollow hypodermic needle in the top of the tube.




- Puncture the side of the tube just below the band with a needle and syringe and draw the sample off. Take care when piercing the tube to avoid pushing the needle out the opposite side.



- Puncture the bottom of the tube and collect the drops.



- Aspirate the sample from the tube top by snipping off the tube stem, then aspirating the sample with a Pasteur pipette or needle and syringe.



Slice the tube, using the Beckman CentriTube Slicer (347960). Refer to publication L-TB-010 for instructions for using the CentriTube Slicer.

For additional information on fraction recovery systems available from Beckman Coulter, refer to the latest edition of *High Performance, High Speed, High Capacity Rotors, Tubes & Accessories* (publication BR-8102) available at www.beckmancoulter.com.

Making Ultra-Clear Tubes Wettable

The following method of making Ultra-Clear tubes wettable has proven successful for some users:

1. Polyvinyl alcohol (2 g) was dissolved in distilled water (50 mL) by stirring and heating to gentle reflux.
2. Isopropanol (50 mL) was slowly added to the hot solution and stirring and heating continued until a clear solution was obtained.
3. The solution was then allowed to cool to room temperature.
4. Ultra-Clear tubes were filled with the coating solution, then aspirated out with a water pump after 15 minutes, leaving a thin film on the tube walls. A small amount of solution that collected in the tube bottoms after standing was removed with a pipette.
5. The tubes were left open to dry at room temperature overnight, then filled with distilled water. After standing overnight at room temperature, the distilled water was poured out.
6. Finally, the tubes were briefly flushed with water, tapped to remove excess liquid, and left to dry.


Using Fixed-Angle Rotors

Introduction

This section contains instructions for using fixed-angle rotors in J series centrifuges. In addition to these instructions, observe procedures and precautions provided in the applicable rotor and centrifuge manuals.

Refer to [CHAPTER 2](#) for labware selection information, and [CHAPTER 3](#) for recommended filling and sealing or capping requirements and for sample recovery procedures. Refer to [CHAPTER 7](#) for information on the care of rotors and accessories.

Description




Fixed-angle rotors (see [Figure 4.1](#)) are general-purpose rotors that are especially useful for pelleting subcellular particles and in short-column banding of viruses and subcellular organelles. Refer to [Table 4.1](#) for general rotor specifications.

Tubes in fixed-angle rotors are held at an angle (usually 20 to 45 degrees) to the axis of rotation. The tube angle shortens the particle pathlength compared to swinging-bucket rotors, resulting in reduced run times. Tubes can be placed directly in a rotor cavity if the diameters of the tube and the cavity are the same. Using adapters, more than one type and size of tube can be centrifuged together, provided that the load is properly balanced.


O-rings, made of Buna N rubber, are located in the rotor lid. The O-rings help to maintain atmospheric pressure inside a fixed-angle rotor during centrifugation, when they are properly lubricated.

A tie-down device or lid-locking knob is used to secure the rotor to the centrifuge drive spindle hub before the run begins.


Figure 4.1 Examples of Fixed-Angle Rotors



JA-20



JA-25.15



JLA-10.500

Table 4.1 General Specifications for Beckman Coulter J Series Fixed-Angle Rotors








Rotor Profile and Name	Max Speed ^a / RCF/ <i>k</i> Factor	Critical Speed Range ^b (rpm)	Radial Distances (mm)			Number of Tubes × Nominal Capacity of Largest Tube	Nominal Rotor Capacity
			<i>r</i> _{max}	<i>r</i> _{av}	<i>r</i> _{min}		
 JA-30.50 TI (34° Angle)	30,000 rpm 108,860 × <i>g</i> 280	600 to 800	108	74	40	8 × 50 mL	400 mL
 JA-25.50 (34° Angle)	25,000 rpm ^c 75,600 × <i>g</i> 418	600 to 800	108	73.2	38.5	8 × 50 mL	400 mL
 JA-25.15 (25° Angle)	25,000 rpm 74,200 × <i>g</i> (outer row) 265	600 to 800	106	79	52	24 × 15 mL	360 mL
	60,200 × <i>g</i> (inner row) 380		86	59	32		
 JA-21 (40° Angle)	21,000 rpm 50,400 × <i>g</i> 470	600 to 800	102	73	45	18 × 10 mL	180 mL
 JA-20.1 (23° Angle)	20,000 rpm 51,500 × <i>g</i> (outer row)	600 to 800	115	107	64	32 × 15 mL	480 mL
	43,900 × <i>g</i> (inner row) 371		98	73	47		
 JA-20 (34° Angle)	20,000 rpm 48,400 × <i>g</i> 770	600 to 800	108	70	32	8 × 50 mL	400 mL
 JA-18.1 (45° Angle Adapter) or (25° Angle Adapter)	18,000 rpm ^d 42,100 × <i>g</i> 156	600 to 800	116	105	95	24 × 1.8 mL	43.2 mL
	17,000 rpm ^d 36,300 × <i>g</i> 91		112	106	101	24 × 1.8 mL	43.2 mL

Table 4.1 General Specifications for Beckman Coulter J Series Fixed-Angle Rotors (*Continued*)















Rotor Profile and Name	Max Speed ^a / RCF/ k Factor	Critical Speed Range ^b (rpm)	Radial Distances (mm)			Number of Tubes × Nominal Capacity of Largest Tube	Nominal Rotor Capacity
			$r_{\max}$	r_{av}	$r_{\min}$		
 JA-14.50 (35° Angle)	14,000 rpm $35,000 \times g$ 787	400 to 1000	160	124	87	16 × 50 mL	800 mL
 JA-18 (23° Angle)	18,000 rpm ^e $47,900 \times g$ 566	600 to 800	132	98	64	10 × 100 mL	1 liter
 JA-17 (25° Angle)	17,000 rpm ^f $39,800 \times g$ 690	600 to 800	123	90	58	14 × 50 mL	700 mL
 JLA-16.250 (25° Angle)	16,000 rpm ^g $38,400 \times g$ 1090	600 to 800	134	90	46	6 × 250 mL	1.5 liter
 JA-14 (25° Angle)	14,000 rpm $30,100 \times g$ 1764	600 to 800	137	86	35	6 × 250 mL	1.5 liter
 F14BCI-14x50cy (34° Angle)	14,000 rpm $33,500 \times g$ 789	600 to 1200	153	118	83	14 × 50 mL	700 mL
 F14BCI-6x250y (23° Angle)	14,000 rpm $30,000 \times g$ 1690	600 to 1200	137	87	37	6 × 250 mL	1 500 mL
 JA-12 (35° Angle)	12,000 rpm $23,200 \times g$ 1244	400 to 1000	144	108	71	12 × 50 mL	600 mL
 JA-10 (25° Angle)	10,000 rpm $17,700 \times g$ 3610	600 to 800	158	98	38	6 × 500 mL	3 liters

Table 4.1 General Specifications for Beckman Coulter J Series Fixed-Angle Rotors (*Continued*)

Rotor Profile and Name	Max Speed ^a / RCF/ k Factor	Critical Speed Range ^b (rpm)	Radial Distances (mm)			Number of Tubes × Nominal Capacity of Largest Tube	Nominal Rotor Capacity
			$r_{\max}$	r_{av}	$r_{\min}$		
 F10BCI-6x500y (23° Angle)	10,000 rpm $17,696 \times g$ 3417	600 to 1200	158	100	41	6 × 500 mL	3 liters
 JLA-10.500 ^h (20° Angle)	10,000 rpm ⁱ $18,600 \times g$ 2850	600 to 800	166	110	64	6 × 500 mL	3 liters
 JLA-9.1000 ⁱ (20° Angle) (use only in Avanti J series centrifuges)	9000 rpm ^j $16,800 \times g$ 2540	200 to 400	185	134	82	4 × 1000 mL	4 liters
 JLA-8.1000 ⁱ (20° Angle) (use only in Avanti J-26S XP series, Avanti J-26 XP series, Avanti J-HC, and discontinued Avanti J-20 XP series and J-20 series centrifuges)	8000 rpm $15,900 \times g$ 2500	200 to 400	222.8	171	119	6 × 500 mL	6 liters


- Maximum speeds are based on a solution density of 1.2 g/mL in all rotors except for the JA-18.1, which is rated for a density of 1.4 g/mL.
- Critical speed range is the range of speeds over which the rotor shifts so as to rotate about its center of mass. Passing through or running at the critical speed range is characterized by some vibration.
- Maximum speed in an Avanti J-E centrifuge is 21,000 rpm.
- When a JA-18.1 rotor is used in the J2-HC centrifuge, derate the rotor as follows: when the 45° adapters are used, do not run the rotor above 15,000 rpm; when 25° adapters are used, do not run the rotor above 16,000 rpm.
- Maximum speed in an Avanti J series centrifuge, except Avanti J-E; maximum speed in an Avanti J-E for a rotor with magnets, maximum speed for rotor without magnets is 13,000 rpm. Maximum speed in a J2 series centrifuge is 14,000 rpm.
- Maximum speed in an Avanti J-E for a rotor with magnets, maximum speed for a rotor without magnets is 13,000 rpm.
- Maximum speed in an Avanti J-E for the rotor with magnets; without magnets maximum is 14,000 rpm. (Maximum speed at 2°C in a 50-Hz centrifuge is 14,000 rpm.)
- Do not put bottles directly into the rotor without canisters.
- Maximum speed for rotor without magnets in an Avanti J-E centrifuge is 6300 rpm.
- Maximum speed for rotor in an Avanti J-E centrifuge is 6300 rpm.



Some rotors have dual-locking lid mechanisms consisting of a daisy knob that secures the lid to the rotor, and a round rotor knob that attaches the rotor to the centrifuge drive spindle hub. (Daisy refers to the knob shape. The grooves between each “petal” let your fingers grip the knob firmly and provide leverage for turning.) The dual-locking capability provides added biosafety by allowing the rotor to be loaded into and removed from the centrifuge with the lid in place. The rotor may be placed under a safety hood before the lid is attached or removed.



Always loosen the rotor knob before loosening the daisy knob to avoid jamming the knobs.



A feature of many Beckman Coulter fixed-angle rotors is a patented fluid-containment annulus, located below the O-ring sealing surface. If tubes are overfilled or if leakage occurs during centrifugation, the annulus holds enough volume that all of the liquid is kept inside the rotor—even if all tubes leak at the same time. This feature virtually eliminates the escape of liquid into the centrifuge chamber.

NOTE Although rotor components and accessories made by other manufacturers may fit in the Beckman Coulter rotor you are using, their safety in the rotor cannot be ascertained by Beckman Coulter. Use of other manufacturers’ components or accessories in the Beckman Coulter rotor may void the rotor warranty, and should be prohibited by your laboratory safety officer. Only the components and accessories listed in the applicable rotor manual should be used.

Tubes and Bottles

Fixed-angle rotors can accommodate a variety of tube types, listed in the rotor manual. Refer to [CHAPTER 3](#) for tube filling and sealing requirements. Observe the maximum rotor speeds and fill volumes listed in the rotor manual.

NOTE JLA-8.1000 and JLA-9.1000 rotors run only the specially designed bottles with polypropylene AutoVent plug and polyphenylsulfone (PPSU) or polyphenylene sulfide (PPS) cap/closures. Refer to the applicable rotor manual for instructions on use of these bottles and accessories.

Rotor Preparation and Loading

For runs at other than room temperature, refrigerate or warm the rotor beforehand for fast equilibration.

Prerun Safety Check



Read all safety information in the rotor manual before using the rotor.

1 Make sure that the rotor and lid are clean and show no signs of corrosion or cracking.

2 Check the chemical compatibilities of all materials used.
(Refer to *Chemical Resistances* (publication IN-175.)


3 Verify that the tubes and bottles being used are listed in the applicable rotor manual.

4 If fluid containment is required, use capped tubes or bottles.
We strongly recommend capping all containers carrying physiological fluids to prevent leakage.

Rotor Preparation


1 Be sure that metal threads in the rotor are clean and lightly but evenly lubricated with Spinkote lubricant (306812).
Also ensure that O-rings are lightly but evenly coated with silicone vacuum grease (335148).

2 Load the filled containers symmetrically into the rotor.
Opposing tubes must be filled to the same level with liquid of the same density.
Refer to [Rotor Balance](#) in [CHAPTER 1](#).



Operation

Refer to the applicable centrifuge instruction manual for detailed operating information. For low-temperature runs, precool the rotor in the centrifuge or in a refrigerator before use—especially before short runs—to ensure that the rotor reaches the set temperature. (To ensure that the rotor reaches the required temperature during centrifugation, some temperature compensation may be required because of the mass of these rotors. Refer to [APPENDIX B](#) or to the applicable rotor manual for tables listing temperature compensation units for various rotors.)




- If you are using an Avanti J series centrifuge (except J-E), select the rotor number.
- If you are using an Avanti J-E or a microprocessor-controlled J2 or J6 series centrifuge, enter the rotor code (if the JA-10 rotor is used for example, enter code **10**).

Installing the Rotor

CAUTION


The centrifuge drive spindle can be bent or broken if the rotor is forced sideways or dropped onto it. Install the rotor by centering it over the spindle and carefully lowering it straight down.

- 1 Carefully lower the rotor straight down onto the drive spindle.
 - a. Rotate it by hand until the drive pins seat on the drive spindle hub.
 - In *older model centrifuges*—be sure the pins in the rotor drive hole are located in the grooves of the drive spindle hub.



Older Model Centrifuges

- In *newer model centrifuges*—be sure the pins in the rotor drive hole are not sitting on top of the teeth on the drive spindle hub.




Newer Model Centrifuges

CAUTION

The pins located in the rotor drive hole must be seated correctly on the centrifuge drive spindle. Running a rotor that is not seated properly may result in severe rotor damage.

- 2 After the rotor is seated on the drive spindle hub, place the lid on the rotor.

- 3 Press down on the knob, then screw it down *tight*.
(Turning the knob to the right [clockwise] attaches the rotor to the hub; the lid on some fixed-angle rotors remains free and may be slipped on or off while the rotor remains secured in the centrifuge.)




NOTE The JA-18 rotor *must be run with the lid on* in Avanti J series centrifuges.

Removal and Sample Recovery

CAUTION


If disassembly reveals evidence of leakage, you should assume that some fluid escaped the container or rotor. Apply appropriate decontamination procedures to the centrifuge, rotor, and accessories.

- 1 Unscrew the rotor lid knob to release the rotor from the spindle hub.




NOTE Labware can be recovered from most fixed-angle rotors while the rotor remains in the centrifuge. You can remove the lid and extract the tubes or bottles using the removal tool specified in the applicable rotor manual. If the rotor is left in the centrifuge between runs, be sure that it is securely tied down before each run. Remove the rotor regularly and clean the drive spindle assembly.


-
- 2 To remove the rotor, lift it straight up and off the drive spindle.



NOTE If the rotor sticks to the drive spindle, screw the short end of the rotor lifting handle assembly into the threaded opening to force the rotor off of the drive spindle hub. Lubrication of the centrifuge drive spindle hub with Spinkote should prevent the rotor from sticking on all centrifuges except Avanti J series. Avanti J series centrifuges have acetal rings on the spindle hubs to prevent sticking and do not require lubrication.



-
- 3 Remove spacers, tubes, and bottles with the appropriate removal tool.



Quick-Seal Tube
Removal Tool
(361668)

-
- 4 Refer to [CHAPTER 3](#) for sample recovery methods.
-


Using Swinging-Bucket Rotors

Introduction

This section contains instructions for using swinging-bucket rotors in J series centrifuges. In addition to these instructions, observe procedures and precautions provided in the applicable rotor and centrifuge manuals.

Refer to [CHAPTER 2](#) for tube selection information, and [CHAPTER 3](#) for recommended labware filling and sealing requirements and for sample recovery procedures. Refer to [CHAPTER 7](#) for information on the care of rotors and accessories.

Description



Swinging-bucket rotors (see [Figure 5.1](#)) are normally used for density gradient separations, either isopycnic or rate zonal. Refer to [Table 5.1](#) for general rotor specifications. A tie-down device or lid-locking knob is used to lock the rotor to the centrifuge drive hub before the run begins.




Tubes or bottles in swinging-bucket rotors are held in the rotor buckets that are attached to the rotor body by hinge pins. The buckets swing out to horizontal position as the rotor is accelerated, and stay horizontal until rotor deceleration begins. During deceleration, the buckets gradually return to vertical position.




NOTE Although rotor components and accessories made by other manufacturers may fit in the Beckman Coulter rotor you are using, their safety in the rotor cannot be ascertained by Beckman Coulter. Use of other manufacturers' components or accessories in the rotor may void the rotor warranty, and should be prohibited by your laboratory safety officer. Only the components and accessories listed in the applicable rotor manual should be used.


Figure 5.1 Examples of Swinging-Bucket Rotors



JS-7.5



JS-4.3



JS-13.1

Table 5.1 General Specifications for Beckman Coulter J Series Swinging-Bucket Rotors














Rotor Profile and Name		Max Speed ^a / RCF/ <i>k</i> Factor	Critical Speed Range ^b (rpm)	Radial Distances (mm)			Number of Tubes x Nominal Capacity of Largest Tube	Nominal Rotor Capacity
				<i>r</i> _{max}	<i>r</i> _{av}	<i>r</i> _{min}		
	JS-24.38	24,000 rpm 110,500 × <i>g</i> 334	N/A	161.0	118.2	75.3	6 × 38.5 mL	231 mL
	JS-24.15	24,000 rpm 103,900 × <i>g</i> 376	N/A	171.3	122.1	72.9	6 × 15 mL	90 mL
	JS-13.1	13,000 rpm 26,500 × <i>g</i> 1841	400 to 1450	140	91	41	6 × 50 mL	300 mL
	JS-7.5	7500 rpm 10,400 × <i>g</i> 1090	600 to 800	165	108	51	4 × 250 mL	1 liter
	JS-5.9	5900 rpm 6570 × <i>g</i>	500 to 1200	194.8	179.6	164.3	10 microplates 4 deep-well plates 2 squarewell plates	384 mL
	JS-5.3	5300 rpm 6130 × <i>g</i> (deep-well plates)	500 to 1200	168.5	153.4	138.6	24 microplates 8 deep-well plates 4 square well-plates	768 mL
		6870 × <i>g</i> (500-mL bottles)		218.4	155.6	92.7	4 × 500 mL	2 liters
	JS-5.2	5200 rpm 6840 × <i>g</i> 9051	600 to 800	226	156	86	4 × 1 liter 4 blood bags 12 microplates 148 RIA tubes	4 liters
	JS-5.0	5000 rpm 7480 × <i>g</i> 9171	300 to 600	267	188	108	4 × 2.25 liters	9 liters
	JS-4.3	4300 rpm 4220 × <i>g</i> 16,635	400 to 1450	204 (buckets) 163 (carriers)			4 × 750 mL 4 blood bags 12 microplates 148 RIA tubes	3 liters

Table 5.1 General Specifications for Beckman Coulter J Series Swinging-Bucket Rotors (*Continued*)

Rotor Profile and Name		Max Speed ^a / RCF/ k Factor	Critical Speed Range ^b (rpm)	Radial Distances (mm)			Number of Tubes x Nominal Capacity of Largest Tube	Nominal Rotor Capacity
				$r_{\max}$	r_{av}	$r_{\min}$		
	JS-4.2	4200 rpm 5020 × <i>g</i> 11,502	600 to 800	254	184	114	6 × 1 liter 6 blood bags 18 microplates 336 RIA tubes	6 liters
	JS-4.2SM (use only in J6 series centrifuges)	4200 rpm 4900 × <i>g</i>	600 to 800	248	182	116	6 triple or quad pack blood bags	6 liters
	JS-4.2A	4200 rpm 5020 × <i>g</i> 11,502	600 to 800	254	184	114	6 × 1 liter 6 blood bags 18 microplates 336 RIA tubes	6 liters
	JS-4.2SMA (use only in J6 series centrifuges)	4200 rpm 4900 × <i>g</i>	600 to 800	248	182	116	6 triple or quad pack blood bags	6 liters
	JS-4.0	4000 rpm 4044 × <i>g</i> 15,296	600 to 800	226	156	86	4 × 1 liter 4 blood bags 12 microplates 148 RIA tubes	4 liters
	JS-3.0	3000 rpm 2560 × <i>g</i> 22,548	600 to 800	254	184	114	6 × 1 liter 6 blood bags 18 microplates 336 RIA tubes	6 liters
	JS-2.9	2900 rpm 2500 × <i>g</i> 24,400	600 to 800	265	192	118	12 × 500 mL blood bags	6 liters

a. Maximum speeds are based on a solution density of 1.2 g/mL.

b. Critical m speeds are based on a solution density of 1.2 g/mL.

† Critical speed range is the range of speeds over which the rotor shifts so as to rotate about its center of mass. Passing through or running at the critical speed range is characterized by some vibration.

Labware

Swinging-bucket rotors can accommodate a variety of tubes, bottles, multiwell titer plates, and blood bags, listed in individual rotor manuals. Refer to [CHAPTER 3](#) for tube filling and sealing requirements. Observe the maximum rotor speeds and fill volumes listed in the applicable rotor manual.

Rotor Preparation and Loading

For runs at other than room temperature, refrigerate or warm the rotor beforehand for fast equilibration.

Prerun Safety Check



Read all safety information in the rotor manual before using the rotor.

-
- 1** Make sure that the rotor and buckets or carriers are clean and show no signs of corrosion or cracking.

 - 2** Check the chemical compatibilities of all materials used.
(Refer to [APPENDIX A.](#))

 - 3** Verify that the tubes, bottles, or carriers being used are listed in the applicable rotor manual.

 - 4** If fluid containment is required, use capped tubes or bottles and/or Aerosolve canisters.
Beckman Coulter strongly recommends capping all containers carrying physiological fluids to prevent leakage.
-


Rotor Preparation

-
- 1** Be sure that metal threads in the rotor yoke are clean and lightly but evenly lubricated with Spinkote lubricant (306812).

 - 2** Ensure that O-rings are in good condition and are lightly but evenly coated with silicone vacuum grease (335148).

 - 3** Ensure that all sealing surfaces are smooth and undamaged for proper sealing.


- 4 Before each use of the rotor, make sure that bucket pin sockets are lubricated with Paint On Graphite Lubricant (977212) as described in [CHAPTER 7](#).



Special Preparation Instructions for JS-24 Series Rotors

Place the rotor on the rotor stand (362785) when it is not in the centrifuge.

- 1 Load the filled containers into the buckets.
Complete loading by placing the correct floating spacers (if required) over the tubes.
- 2 Ensure that bucket O-rings are lightly but evenly coated with silicone vacuum grease.
Do not run a bucket without an O-ring, as the bucket will leak.
- 3 Be sure that metal threads in the bucket caps are clean and lightly but evenly lubricated with Spinkote lubricant.
Put bucket caps on the buckets and screw them down manually.



- 4 Hook all buckets, loaded or empty, on to the rotor, and be sure that both hooks are on the crossbar.
All six buckets must be in the same size.
Do not intermix the smaller and larger buckets in a single run.
 - If fewer than six tubes are being run they must be arranged symmetrically in the rotor.
 - Opposing tubes must be filled to the same level with liquid of the same density.


Loading the Rotor Yoke




CAUTION

The centrifuge drive spindle can be bent or broken if the rotor is forced sideways or dropped onto it. Install the rotor by centering it over the spindle and carefully lowering it straight down.

- 1 Carefully lower the rotor yoke straight down onto the drive spindle.
 - a. Rotate it by hand until the drive pins seat on the drive spindle hub.




Older Model Centrifuges




Newer Model Centrifuges

NOTE Except for the JS-24.38 and JS-24.15 rotors, you can leave the rotor yoke in the centrifuge between runs unless spillage has occurred—in which case you should remove the buckets or carriers and yoke and clean the centrifuge and rotor components immediately, according to the instructions in the centrifuge and rotor instruction manuals. The JS-24.38 and JS-24.15 rotors must be removed from the centrifuge to install or remove buckets.


- 2 When the yoke is correctly seated, secure it to the drive spindle hub.
 - a. Rotors *with* tie-down knobs—hand tighten the tie-down knob. If the rotor is left in the centrifuge between runs, tighten the knob before each run.



- b. Rotors *without* tie-down knobs—secure the rotor with the tie-down screw, and tighten the screw with the torque bar or tie-down tool. If the rotor is left in the centrifuge between runs, ensure that the screw is *tight* before each run.




- c. JS-4.2A and JS-4.2SMA—secure the rotor to the drive hub with the tie-down bolt (368518). Tighten the bolt with the tie-down tool (368521), then remove the tool.




NOTE Older JS-4.2A or JS-4.2SMA rotors may be secured to the drive hub with tie down bolt (339031). Tighten the bolt with the spanner wrench (870666) or tie-down tool (368521), then remove the wrench or tool.

- d. JS-5.0—secure the rotor to the drive hub with the tie-down bolt (367824). Tighten the bolt with the tie-down tool (368521), then remove the tool.



Loading JS-24 Series Rotors

- 1 To install the rotor, carefully lift it up off the rotor stand with both hands—do not lift the rotor by the adapter—and place it on the drive hub.



Make sure that the rotor pins are perpendicular to the drive hub pins.

The pins must not rest on top of each other; turn the rotor to the right (clockwise) by hand to check for proper installation.

- 2 Turn the tie-down knob to the right (clockwise) to secure the rotor.

Symmetric and Balanced Loading


To ensure optimal performance and stability, *swinging-bucket rotors must be loaded symmetrically*. Two factors affect symmetric loading:

1. The buckets or carriers must be loaded symmetrically with respect to their pivotal axes (the pivotal axis runs parallel to the crossbar, see [Figure 5.2](#)).
2. The rotor should be loaded symmetrically with respect to its center of rotation.

This means that for best results you should load opposing buckets or carriers with the same type of labware containing the same amounts of fluid of equal density. Additionally, buckets or carriers placed opposite each other on the rotor yoke must balance to within a certain weight, typically 10 grams (see the applicable rotor manual for details). Do not exceed the rated maximum load for buckets or carriers.

NOTE The JS-4.2A and JS-4.2SMA swinging-bucket rotors incorporate ARIES (Automated Rotor Imbalance Equilibrating System) “Smart Balance” technology, which provides imbalance compensation for rotors with buckets that are up to 100 grams unbalanced due to different loading volumes or tube or bag breakage.


Beckman Coulter supplies buckets and carriers for most rotors in weight-matched sets to make balancing easier (the weight and date of manufacture are marked on the side of each bucket and bottom of each carrier). Modular disk adapters are also sold in weight-matched sets. However, there are variances in weight between sets, as well as variance in weight between previously purchased adapters. To prevent accidental imbalance, it is important to keep matched sets of adapters together and to weigh other adapters to be sure they are approximately the same. Marking matched sets will help you keep them together.



It is not necessary to completely fill all tubes, positions in adapters, or wells in microtiter plates; however, partially filled adapters or microtiter plates must be balanced with respect to the pivotal axis of the bucket or carrier as discussed below.

During a run, buckets and carriers swing 90 degrees from their at-rest position. The pivotal axis of a bucket or carrier can be imagined as a line extending across the bucket or carrier from one pivot pin to the other. If a bucket or carrier is loaded so that its weight is unequally distributed on either side of its pivotal axis, it will not hang vertically at rest and, more importantly, may not swing to a horizontal position during a run. As a result, extra stress will be placed on the bucket, carrier, tubes, and/or microtiter plates during the run, increasing the possibility of breakage or rotor imbalance.

Figure 5.2 Examples of Correctly and Incorrectly Loaded Buckets and Carriers*



* Contents of opposing buckets must be the same and each bucket must be balanced on its pivotal axis.


Loading Buckets

Buckets can be loaded before or after being installed on the rotor yoke. For best results, fill the appropriate labware first and then load the labware into the buckets. This is especially important when using blood bags—you can trip the imbalance detector in the centrifuge by pushing blood bags into cups within buckets that are installed in the rotor. You can also bend the centrifuge drive spindle.

Using Modular Disk Adapters

Assemble modular disk adapters as follows.

- 1 Select the appropriate adapter base and attach a bail to it.
- 2 Place the base and bail in an empty bucket or on the lab bench (not in the rotor).
- 3 Position one of the disks so that its grooves are aligned with the bail.
 - a. Push the disk down until the bail snaps into the grooves.



- 4 Add more disks until the height of the assembly is nearly as tall as the tubes you will be using. (If the height of the disks is very tall, you may have to push the bail into the grooves of the top disks by hand.)
 - a. Remove or add disks to the bail to accommodate shorter or longer tubes.
 - b. If the tubes fit too snugly in the adapter's rubber base, apply a light film of dusting power, such as talcum powder, to prevent the tubes from sticking.

Place each tube in an adapter so that its weight is balanced by a tube in a diametrically opposite position across the pivotal axis in the same adapter. Adapters placed in opposing buckets should also be filled the same way (see [Figure 5.2](#)). If you must run only one tube in an adapter, be sure this tube rests over the bucket's pivotal axis.

NOTE Be sure to run a tube of the same approximate weight in the same configuration in the opposite bucket.


Using Blood Bag Cups

Do not pour liquid directly into blood bag cups. Fit blood bags into cups before loading the cups into the rotor buckets. Load the blood bag cups as follows:

- 1 Load the cups so that the blood bags and tubing fit as far down as possible.
NOTE Make sure the bags stay as vertical as possible, with no folds at the top or corners. If folds are present, blood cells could remain in the folds and then mix with the plasma when the bag is removed.
- 2 Sandwich the tubing between the blood bag and any transfer packs (see [Figure 5.3](#)).
- 3 Make sure the loaded blood bag cups opposite each other on the rotor yoke are approximately the same weight (within limits listed in the applicable rotor manual).
In some rotors, balancing pads can be used if necessary to maintain weight balance.
- 4 Place loaded cups into rotor buckets.
 - a. If only two filled cups are run, place them in opposing buckets.
The remaining buckets should contain similar “blank” loads to prevent imbalance (either empty modular disk adapters or water-filled blood bags in cups).

CAUTION


If bucket covers or rotor lids are not used, make sure the superstructure of the blood bag protruding from the cup does not inhibit the bucket from reaching its horizontal position. If it does, remove the cup from the rotor and reposition the blood bag so that it seats further into the cup. Allowing the blood bags to contact the rotor yoke during centrifugation can cause the bucket to come off the pivot pins and can seriously damage both the rotor and the centrifuge.

Figure 5.3 Typical Blood Bag Loading Procedures (JS-24.3 Rotor Shown)

Loading Buckets Into the Rotor

NOTE JS-24 series rotors must have buckets attached before the rotor is put into the centrifuge.

- 1 If bucket covers or rotor lids are used to help contain spills and glass particles that could result from tube breakage, make sure cover O-rings are in good condition and lightly coated with silicone vacuum grease.
 - a. Before use, inspect Aeroseal cover sealing surfaces, especially the O-ring groove. It must be smooth and free of scratches.



- b. Also ensure that the top 2.54 cm (1 in.) of the bucket is clean and smooth; buckets with scratches or gouges in this surface will not seal properly.
 - c. Inspect the O-ring and plug for nicks, abrasions, and other damage.
 - d. Replace damaged components.

-
- 2 Load the filled buckets (and/or carriers) onto the rotor yoke pivot pins, following the instructions in the rotor manual.
Make sure that the buckets are properly seated by gently swinging them on the pivot pins.
-

NOTE All positions on the rotor yoke must contain either a bucket or a microtiter plate carrier during a run. Consult the applicable centrifuge instruction manual for operating instructions.

Using Microtiter Plate Carriers


Anodized aluminum microtiter plate carriers can be installed on the pivot pins in place of the buckets normally used with some swinging-bucket rotors. Each carrier allows centrifugation of up to three 96-well microtiter plates. (For complete information about the carriers, see publication GS6-TB-011, which accompanies the Micro Plus carriers, or publication J6-TB-009, which accompanies the J6 series carriers.)

Microplates will break if g -forces are too high. Rotor speed must be reduced when microplate carriers are used. If microplate carriers and buckets are centrifuged in the same run, run speed must be reduced to the speed allowable for the microplates. Refer to the applicable rotor manual for allowable run speeds.


If only two carriers are run, they must be installed opposite each other in the rotor, and the remaining positions on the yoke must be filled with either buckets or other carriers (they need not be loaded) to prevent rotor imbalance. (See [Symmetric and Balanced Loading](#), above.)

Micro Plus Carriers

-
- 1 To prevent microtiter plate breakage during centrifugation, place the flexible plastic pad, ridged side up, into the flat, indented area of the blue base (see [Figure 5.4](#)).
 - 2 Place the plate(s) on top of the pad, being careful not to spill the contents.
 - 3 Slide the base, pad, and plate assembly into the carrier until the base clicks into place.
-

Figure 5.4 The Micro Plus Microtiter Plate Carrier, Base, Pad, and Deep-Well Microtiter Plate


JS-5.9 and JS-5.3 Plate Carriers



High-impact thermoplastic carriers are used in the JS-5.9 and JS-5.3 rotor buckets to provide support to labware during centrifugation and facilitate loading and unloading buckets. Each rotor bucket can carry a 96-well kit for high-throughput processing (such as a DNA or RNA kit), or standard microplates used in the serial dilution of small liquid volumes—should contain five (JS-5.9) or six (JS-5.3) stacked 96-well polypropylene plates, two (stacked) deep-well plates, or one square-well plate per bucket. Proper loading of the plate carriers will ensure that there is adequate support for the wire handles.

NOTE When using stacked polypropylene plates, place a support pad beneath the bottom plate to prevent breakage during centrifugation. Use the support pad beneath all polystyrene plates.

- 1 If using polystyrene or stacked polypropylene plates, place a support pad in the carrier with the ridged-cross side down.



- 2 If using polystyrene or stacked polypropylene plates, place a support pad in the carrier with the ridged-cross side down.
- 3 After centrifugation, grasp the carrier by the wire handles and lift it straight up out of the bucket to unload it.

J6 Carriers

Carriers used with the JS-5.2 and JS-4.0 rotors are NOT interchangeable with those used with the JS-4.2, JS-4.2A, JS 4.2SM, JS-3.0, and JS-2.9 rotors. If you have more than one type of carrier, check the label on the side of the carrier to make sure that you are using the right one for your rotor.

[Table 5.2](#) lists carriers used with J6 rotors.

- 1 To prevent microtiter plate breakage during centrifugation, place the rubber pad that comes with each carrier on the bottom of the carrier.



- 2 Place the plate(s) on top of the pad, being careful not to spill the contents.

Table 5.2 Microplate Carriers Used with J6-Series Rotors

Rotor Type	Number of Carriers	Rotor Loads	Maximum Run Speed	Carrier Set Part Number
JS-5.2 or JS-4.0	4	12 single	2600 rpm	358680 (set of 2)
JS-4.2, JS-4.2A, JS 4.2SM, JS-3.0, or JS-2.9	6	18 single or 6 deep-well	2500 rpm	358682 (set of 2)

Operation

Refer to the centrifuge instruction manual for detailed operating information. For low-temperature runs, precool the rotor in the centrifuge or in a refrigerator before use—especially before short runs—to ensure that the rotor reaches the set temperature. (To ensure that the rotor reaches the required temperature during centrifugation, some temperature compensation may be required because of the mass of these rotors. Refer to [APPENDIX B](#) or to the rotor manual for tables listing temperature compensation units for various rotors.)

If you are using a microprocessor-controlled J2 or J6 series centrifuge, enter the rotor code (if the JS-5.2 rotor is used for example, enter code 5.2).

Sample Recovery

CAUTION

If disassembly reveals evidence of leakage, you should assume that some fluid escaped the container or rotor. Use appropriate decontamination procedures on the centrifuge, rotor, and accessories.

- 1 Remove the rotor lid (if applicable). Remove the buckets or carriers from the rotor.
- 2 Remove labware from the buckets or carriers.


NOTE Except for the JS-24.38 and JS-24.15 rotors, you can leave the rotor body or yoke in the centrifuge between runs unless spillage has occurred—in which case you should remove the buckets or carriers and yoke and clean the centrifuge and rotor components immediately, according to the instructions in the centrifuge and rotor instruction manuals. If the rotor is left in the centrifuge between runs, tighten the tie-down device before each run. The JS-24.38 and JS-24.15 rotors must be removed from the centrifuge to install or remove buckets.

Removing JS-24 Series Rotors

- 1 Remove the rotor from the centrifuge by lifting it straight up and off the drive hub.
- 2 Set the rotor on the rotor stand and carefully remove the buckets.

- 3 Remove the bucket caps and use the appropriate removal tool (listed in the rotor manual) to remove the spacers and tubes.

If floating spacers were used, remove them with the threaded end of the floating spacer removal tool (338765).



NOTE If the conical-shaped adapters that support konical tubes are difficult to remove after centrifugation, an extractor tool (354468) is available to facilitate removal.

See [CHAPTER 7](#) for instructions on the care of rotors, tubes or bottles, and accessories after a run.

Using Vertical-Tube and Rack-Type Rotors

Introduction



This section contains instructions for using vertical-tube or rack-type rotors in J series centrifuges. In addition to these instructions, observe procedures and precautions provided in the applicable rotor and centrifuge manuals.

Refer to [CHAPTER 2](#) for tube selection information, and [CHAPTER 3](#) for recommended filling and sealing requirements for each tube type and for sample recovery procedures. Refer to [CHAPTER 7](#) for information on the care of rotors and accessories.

Description

Refer to [Table 6.1](#) for general operating specifications for vertical-tube and rack-type rotors.


Table 6.1 General Specifications for Beckman Coulter J Series Vertical-Tube and Rack-Type Rotors

Rotor Profile and Name	Max Speed ^a / RCF/ k Factor	Critical Speed Range ^b (rpm)	Radial Distances (mm)			Number of Tubes × Nominal Capacity of Largest Tube	Nominal Rotor Capacity
			$r_{\max}$	r_{av}	$r_{\min}$		
 JV-20 (0° Angle)	20,000 rpm $41,619 \times g$ 206	600 to 800	93	80	67	8 × 39 mL	312 mL
 JR-3.2 (90° Angle at Speed)	3200 rpm $2280 \times g$ 25,606	600 to 800	199	80	67	320 × 1 mL	320 mL

a. Maximum speeds are based on a solution density of 1.7 g/mL for the JV-20 rotor and 1.2 g/mL for the JR-3.2 rotor.


b. Critical speed range is the range of speeds over which the rotor shifts so as to rotate about its center of mass. Passing through or running at the critical speed range is characterized by some vibration.

Vertical-Tube Rotors



Vertical-tube rotors (see [Figure 6.1](#)) hold tubes parallel to the axis of rotation; therefore, bands separate across the diameter of the tube rather than down the length of the tube (see [Figure 1.3](#)). Vertical-tube rotors are useful for separating and banding subcellular particles. These rotors have plugs that are screwed into the rotor cavities over sealed Quick-Seal tubes. The plugs (with spacers, when required) restrain the tubes in the cavities and provide support against the hydrostatic force generated by centrifugation. Refer to [CHAPTER 3](#) for information about filling and sealing Quick-Seal tubes for use in vertical-tube rotors.


Figure 6.1 Vertical-Tube Rotor



Rack-Type Rotors

The rack-type rotor (see [Figure 6.2](#)) holds a wide range of gamma-counter tubes in tube racks. Racks are loaded into removable trays, which are then loaded into carriers at a resting angle. During centrifugation, the carriers swing out to a completely horizontal position to provide uniform pelleting of samples. The nearly vertical position of the racks during centrifugation permits processing of up to 320 mL in one run.

Figure 6.2 Rack-Type Rotor



NOTE Although rotor components and accessories made by other manufacturers may fit in the Beckman Coulter rotor you are using, their safety in the rotor cannot be ascertained by Beckman Coulter. Use of other manufacturers' components or accessories in the Beckman Coulter rotor may void the rotor warranty, and should be prohibited by your laboratory safety officer. Only the components and accessories listed in the applicable rotor manual should be used.

Using a Vertical-Tube Rotor

Tubes and Bottles

Only Quick-Seal tubes, listed in the rotor manual, may be centrifuged in a vertical-tube rotor. Refer to [CHAPTER 3](#) for filling and sealing requirements of Quick-Seal tubes. Observe the maximum rotor speeds and fill volumes listed in the rotor manual.

Rotor Preparation and Loading

For runs at other than room temperature, refrigerate or warm the rotor beforehand for fast equilibration.

Prerun Safety Checks

Read all safety information in the rotor manual before using the rotor.

-
- 1 Make sure that the rotor and plugs are clean and show no signs of corrosion or cracking.
 - 2 Check the chemical compatibilities of all materials used. (Refer to [APPENDIX A](#).)
 - 3 Verify that the tubes, spacers, and floating spacers being used are in good condition and are listed in the rotor manual.
-

Rotor Preparation


-
- 1 Be sure that the plug threads are lightly but evenly lubricated with Spinkote lubricant (306812).
 - 2 Lubricate the rotor drive hole with silicone vacuum grease (335148).
-

3 Load the filled and sealed tubes symmetrically into the rotor.

Opposing tubes must be filled to the same level with liquid of the same density.

Refer to [Rotor Balance](#) in [CHAPTER 1](#).

4 Insert spacers and floating spacers, as listed in the rotor manual, to completely fill rotor cavities in use.




5 With the rotor in the rotor vise (332688), insert plugs over filled cavities *only*; do not insert plugs in empty cavities.

Tighten the plugs using the plug wrench provided (340632).

Operation

Refer to the centrifuge instruction manual for detailed operating information. For low-temperature runs, precool the rotor in the centrifuge or in a refrigerator before use—especially before short runs—to ensure that the rotor reaches the set temperature. (To ensure that the rotor reaches the required temperature during centrifugation, some temperature compensation may be required because of the mass of these rotors. Refer to [APPENDIX B](#) or to the rotor manual for tables listing temperature compensation units for various rotors.)

- If you are using an Avanti J series centrifuge, select the rotor number.
- If you are using a microprocessor-controlled J2 or J6 series centrifuge, enter the rotor code (enter code **20** for the JV-20 rotor).




Installing the Rotor

CAUTION


The centrifuge drive spindle can be bent or broken if the rotor is forced sideways or dropped onto it. Install the rotor by centering it over the spindle and carefully lowering it straight down.

- 1 Carefully lower the rotor straight down onto the drive spindle.
- 2 Rotate the rotor by hand until the drive pins seat on the drive spindle hub.
 - a. In *older model centrifuges*—be sure the pins in the rotor drive hole are located in the grooves of the drive spindle hub.



Older Model Centrifuges

- b. In *newer model centrifuges*—be sure the pins in the rotor drive hole are not sitting on top of the teeth on the drive spindle hub.



Newer Model Centrifuges

CAUTION

The pins located in the rotor hub must be seated correctly on the centrifuge drive spindle. Running a rotor that is not seated properly may result in rotor failure.

- 3 Secure the rotor to the drive spindle hub with the rotor tie-down assembly.

Removal and Sample Recovery




If disassembly reveals evidence of leakage, you should assume that some fluid escaped the container or rotor. Apply appropriate decontamination procedures to the centrifuge and accessories.


- 1 Remove the rotor tie-down assembly.
- 2 If the rotor straightens up and off the drive spindle. If the rotor sticks to the drive spindle, a rotor removal tool may be used.

NOTE Lubrication of the centrifuge drive spindle hub with Spinkote should prevent the rotor from sticking on all centrifuges except Avanti J series. Avanti J series centrifuges have acetal rings on the spindle hubs to prevent sticking and do not require lubrication.

- 3 Place the rotor in the rotor vise and use the plug wrench to remove the rotor plugs.



- 4 Remove spacers with the floating spacer removal tool (338765) and tubes with the tube removal tool (361668).




- 5 Refer to Section 3 for sample recovery methods.

Using a Rack-Type Rotor

Trays and Tubes

Two kinds of trays are available for use in the rack-type rotor to accommodate a variety of racks and tube sizes. The trays can be identified by color, as listed in the rotor manual. Some racks require the use of adapters, spacers, or frames to ensure a proper fit in the tray. Refer to the rotor manual to select compatible labware.



NOTE Tubes should be no longer than 105 mm for proper clearance.

Rotor Preparation and Loading

For runs at other than room temperature, refrigerate or warm the rotor beforehand for fast equilibration.

Prerun Safety Checks

Read all safety information in the rotor manual before using the rotor.

- 1 Make sure that the rotor yoke and carriers are clean and show no signs of corrosion or cracking.
- 2 Check the chemical compatibilities of all materials used. (Refer to [APPENDIX A](#).)
- 3 Verify that the tube racks, trays, adapters, and spacers being used are in good condition and are listed in the rotor manual.


Rotor Preparation

- 1 Lubricate the rotor drive hole with silicone vacuum grease (335148).
- 2 Load racks into either two or four trays, then load tubes into the racks.

Do not over-fill tubes; leave enough space to avoid spills during carrier loading.

NOTE If all tubes to the same level with liquid of the same density. Racks and tubes must be horizontally and vertically symmetrical during centrifugation.

- 3 Rest the end of the loaded tray on the carrier base.
 - a. Slide the tray down so that it passes under the hinge pins.
 - b. When it reaches the lower end of the carrier, seat the tray bottom completely into the carrier.



Operation

Refer to the centrifuge instruction manual for detailed operating information. For low-temperature runs, precool the rotor in the centrifuge or in a refrigerator before use—especially before short runs—to ensure that the rotor reaches the set temperature. (To ensure that the rotor reaches the required temperature during centrifugation, some temperature compensation may be required because of the mass of these rotors. Refer to [APPENDIX B](#) or to the rotor manual for tables listing temperature compensation units for various rotors.)

If you are using a microprocessor-controlled J6 series centrifuge, enter the rotor code (enter code **3.2** for the JR-3.2 rotor).


NOTE These rotors are not used in Avanti J series centrifuges.

Installing the Rotor

CAUTION


The centrifuge drive spindle can be bent or broken if the rotor is forced sideways or dropped onto it. Install the rotor by centering it over the spindle and carefully lowering it straight down.

- 1 Lift the rotor by the yoke and carefully lower it straight down onto the drive spindle.
- 2 Rotate the rotor by hand until the drive pins seat on the drive spindle hub.
 - In *older model centrifuges*—be sure the pins in the rotor drive hole are located in the grooves of the drive spindle hub.



Older Model Centrifuges

- In *newer model centrifuges*—be sure the pins in the rotor drive hole are not sitting on top of the teeth on the drive spindle hub.



Newer Model Centrifuges

CAUTION

The pins located in the rotor hub must be seated correctly on the centrifuge drive spindle. Running a rotor that is not seated properly may result in rotor failure.

- 3 Turn the tie-down knob to the right (clockwise) until the rotor is secure.

Removal and Sample Recovery



If disassembly reveals evidence of leakage, you should assume that some fluid escaped the container or rotor. Apply appropriate decontamination procedures to the centrifuge, rotor, and accessories.

1 Turn the tie-down knob to the left (counterclockwise) to release the rotor from the drive spindle.

2 Lift the rotor straight up and off the drive spindle.

If the rotor sticks to the drive spindle, a rotor removal tool may be used.

NOTE Lubrication of the centrifuge drive spindle hub with Spinkote should prevent the rotor from sticking.

3 To remove a tray from a carrier, lift the end of the tray just enough to clear the carrier. Slide the tray up so that it passes under the hinge pins.

Care and Maintenance

Introduction

This section provides information on the care of rotors and accessories. Included is a list of some common operating problems with suggestions for their solutions. Rotors and accessories should be kept in optimal condition, thus minimizing the chances of rotor or labware failure. In addition to these instructions, observe procedures and precautions provided in individual rotor manuals. Chemical Resistances (IN-175) provides the chemical compatibilities of rotor and accessory materials to various acids, bases, salts, and solvents.

Rotor Care

Rotor care involves not only careful operating procedures but also careful attention to:

- Regular cleaning, decontamination, and/or sterilization as required,
- Frequent inspection,
- Corrosion prevention, and
- Regular and proper lubrication.

Do not use sharp tools on a rotor, as the surface can get scratched. Corrosion begins in scratches and may open fissures in the rotor with continued use. The corrosion process accelerates with speed-induced stresses. The potential for damage from corrosion is greatest in aluminum rotors and components.

Cleaning

Wash rotors and rotor components immediately if salts or other corrosive materials are used or if spillage has occurred. DO NOT allow corrosive materials to dry on the rotor.

NOTE Do not wash rotor components or accessories in a dishwasher. Do not soak in detergent solution for long periods, such as overnight.

With normal usage, wash rotors frequently to prevent corrosion that can begin in scratches. Remove buckets from yokes before cleaning swinging-bucket rotors.



Do not immerse or spray a swinging bucket rotor yoke (or body) with water because liquid can become trapped in the hinge pin area and lead to corrosion.

- 1 Use plastic or wooden tools to remove O-rings or gaskets for cleaning—*do not use metal tools* that could scratch anodized surfaces.
 - a. Use a mild detergent such as Solution 555 (339555) and a soft brush to wash rotors and rotor components and accessories.
 - Dilute the detergent with water (10 parts water to 1 part detergent).
 - (Most laboratory detergents are too harsh for aluminum rotors and components.)
 - The Rotor Cleaning Kit (339558) contains two quarts of Solution 555 and brushes that will not scratch rotor surfaces.



- 2 Rinse thoroughly with water.
- 3 Air-dry the body or buckets upside down.

Do not use acetone to dry rotors.
- 4 Wipe clean the O-rings or gaskets regularly (lubricate after cleaning).

Replace them about twice a year or as required.

- 5 Frequently clean all surfaces that contact O-rings.
 - a. Regularly clean the threads of the rotor (lid, handle, buckets, cavities, and so on) with a nonmetal brush and a small amount of concentrated detergent, then rinse, and dry thoroughly.
 - b. Lubricate the threads as directed under [Lubrication](#), below.
- 6 Approximately once a week (or every 250 runs), clean the pins and bucket pin sockets of swinging bucket rotors to prevent buildup of residues.

After cleaning, lubricate pin sockets as described under [Lubrication](#), below.

 CAUTION

Do not use acetone, MEK (methylethylketone), chloroform, cyclohexane, or organic solvents on carbon-fiber canisters at any time. These substances will damage the epoxy resin surface material.

Decontamination

Rotors contaminated with radioactive or pathogenic materials must be decontaminated, following appropriate laboratory safety guidelines and/or other regulations.

NOTE. Strong bases and/or high-pH solutions can damage aluminum rotors and components.



- If a rotor (and/or accessories) becomes contaminated with radioactive material, it should be decontaminated using a solution that will not damage the anodized surfaces. Beckman Coulter has tested a number of solutions and found two that do not harm anodized aluminum: RadCon Surface Spray or IsoClean Solution (for soaking),^{*} and Radiacwash.[†]

NOTE. IsoClean can cause fading of colored anodized surfaces. Use it only when necessary, and do not soak rotor components longer than the minimum time specified in the IsoClean usage instructions. Then remove it promptly from surfaces.

While Beckman Coulter has tested these methods and found that they do not damage components, no guarantee of decontamination is expressed or implied. Consult your laboratory safety officer regarding the proper decontamination methods to use.




- If the rotor or other components are contaminated with toxic or pathogenic materials, follow appropriate decontamination procedures as outlined by appropriate laboratory safety guidelines and/or other regulations. Consult *Chemical Resistances* (IN-175) to select an agent that will not damage the rotor.

^{*} In U.S., contact Nuclear Associates (New York); in Eastern Europe and Commonwealth States, contact Victoreen GmbH (Munich); in South Pacific, contact Gammasonics Pty. Ltd. (Australia); in Japan, contact Toyo Medic Co. Ltd. (Tokyo).

[†] In U.S., contact Biodex Medical Systems (Shirley, New York); internationally, contact the U.S. office to find the dealer closest to you.

Sterilization and Disinfection

When sterilization or disinfection is a concern, consult your laboratory safety officer regarding proper methods to use. While Beckman Coulter has tested the following methods and found that they do not damage the rotor or components, no guarantee of sterility or disinfection is expressed or implied.



- Rotors and most rotor components (except those made of modified polyphenylene oxide) can be autoclaved at 121°C for up to an hour. Remove the lid and place the rotor (and/or buckets) in the autoclave upside-down. (O-rings and gaskets can be left in place on the rotor.)
- Ethanol (70%) may be used on all rotor components, including those made of plastic. Bleach (sodium hypochlorite) may be used, but may cause discoloration of anodized surfaces. Use the minimum immersion time for each solution, per laboratory standards.




Ethanol is a flammability hazard. Do not use it in or near operating centrifuges.


Inspection

Frequent and thorough inspection is crucial to maintaining a rotor in good operating condition.


- Periodically (at least monthly, depending on use) inspect the rotor, especially inside cavities and buckets, for rough spots, cracks, pitting, white powder deposits on aluminum rotors (frequently aluminum oxide), or heavy discoloration. If any of these signs are evident, do not run the rotor. Contact your Beckman Coulter representative for information about the Field Rotor Inspection Program and the Rotor Repair Program.
- Regularly check the condition of O-rings or gaskets and replace any that are worn or damaged.
- Regularly check that all sealing surfaces are smooth and undamaged to ensure proper sealing.



Before each use, inspect Aeroseal cover sealing surfaces, especially the O-ring groove. It must be smooth and free of scratches. Also ensure that the top 2.54 cm (1 in.) of the bucket is clean and smooth; buckets with scratches or gouges in this surface will not seal properly. Inspect the O-ring and plug for nicks, abrasions, and other damage. Replace damaged components with Beckman Coulter parts only; *do not use a substitute for the O-ring—it has been specifically selected for this application.*



Regularly check the condition of the Micro Plus carrier base and pad and do not use them if there are signs of damage. Retire the base from use after 1 year.



Before each use, inspect carbon-fiber canisters for cracks where carbon-fiber threads are visible. If any cracking or other damage is visible on the outside wall or near the aluminum-composite interface area, do not use the canister. Contact your Beckman Coulter representative. Retire the canister on the expiration date.

Lubrication


Regular and thorough lubrication can extend the useful life of rotor components.

O-Rings

Many rotors use O-rings or gaskets as seals to maintain atmospheric pressure in the rotor during a run. These O-rings and the rotor surfaces they bear against must be kept clean and evenly lubricated. After removing and cleaning rotor or bucket O-rings or gaskets, lightly but uniformly coat them with silicone vacuum grease and reposition them in the rotor.

NOTE Do not apply lubricant with a cotton-tipped swab. These swabs can leave lint on the O-ring or gasket that can interfere with the seal.

Pivot Pins and Buckets



JS-4.2, JS-4.2A, JS-4.2SMA, JS-5.9, JS-5.3, JS-5.0, and JS-4.3 Rotors—Approximately once a week, and after cleaning and/or autoclaving, lubricate the pin sockets with a lubricant such as Paint On Graphite Lubricant (977212). Allow the lubricant to dry for at least 5 minutes before installing the rotor in a centrifuge.

JS-5.2, JS-4.2, JS-4.2A, JS-4.0, and JS-3.0 Rotors—Lubricate the O-ring and plug of AeroSeal bucket covers with silicone vacuum grease. Also, lightly grease the inside top 1.27 cm (0.5 in.) of the bucket.

Field Rotor Inspection Program

The Field Rotor Inspection Program (FRIP) has two purposes:

- to prevent premature rotor failures by detecting conditions such as stress, corrosion, metal fatigue, damage, or wear in the anodized coatings; and
- to instruct laboratory personnel in the proper care of rotors.


Beckman Coulter has trained a group of experienced service engineers in the techniques of nondestructive evaluation. For more information about the program, contact your Beckman Coulter representative.*

Tube, Bottle, and Accessory Care

Proper care of tubes and bottles involves observing temperature, fill volume, and run speed limitations as well as careful cleaning and sterilization procedures.

Cleaning

Do not wash tubes and bottles in a commercial dishwasher — detergents and temperatures are too harsh.



- Wash tubes, bottles, adapters, and blood bag cups by hand, using a mild detergent, such as Solution 555 (339555) and a soft brush.
- Dilute the detergent with water (10 parts water to 1 part detergent).
- Disassemble multitube adapters for cleaning. After washing with Solution 555 and a soft brush, rinse them with water, then dry and reassemble.
- Polycarbonate bottles and tubes are vulnerable to attack by alkaline solutions and detergents, so use a detergent with pH less than 9, such as Solution 555. Do not use a brush with exposed metal; scratches in polycarbonate will cause early failure.
- Alcohol and acetone react unsatisfactorily with many tube and accessory materials. If a solvent must be used to rinse, dry, or decontaminate these materials, consult [APPENDIX A](#) to select an appropriate solvent.
- Do not dry tubes, bottles, or accessories in an oven. Labware should be air-dried.
- Quick-Seal, Ultra-Clear, and thinwall polypropylene tubes are intended for one-time use and should be discarded after use.


* Call 1-800-742-2345 (U.S.A. or Canada), outside the U.S., contact your local Beckman Coulter office.

Decontamination



Labware contaminated with radioactive or pathogenic solutions should be decontaminated or disposed of following appropriate safety guidelines and/or regulations. Consult *Chemical Resistances* (IN-175) to select an agent that will not damage the tube or bottle material.

Sterilization and Disinfection



Refer to [Table 7.1](#) for sterilization methods recommended for each container type.

Most tubes and accessories, *except those made of Ultra-Clear, polyethylene, Noryl, or cellulose propionate*, can be autoclaved at 121°C for about 20 minutes. Note that autoclaving reduces the lifetime of polycarbonate tubes. Also, polypropylene tubes may be permanently deformed if they are autoclaved many times or if they are handled or compressed before they cool. Tubes should be placed open-end down or supported in a rack if autoclaved.

CAUTION

Do not autoclave tubes or bottles with caps on. Pressure in a sealed container can cause an explosion. Pressures within the autoclave can cause partially sealed containers to collapse when the autoclave vents.

Table 7.1 Tube and Bottle Sterilization and Disinfection^a

Tube/Bottle Material	Autoclave ^b (121°C)	UV Irradiation	Ethylene Oxide	Formaldehyde	Ethanol (70%) ^c	Sodium Hypochlorite (10%)	Hydrogen Peroxide (10%)	Glutaraldehyde (2%)	Phenolic Derivatives
polypropylene	yes	no	yes	yes	yes	yes	yes	yes	no
Ultra-Clear	no	no	yes	yes ^d	yes	yes	yes	yes	no
polycarbonate	yes ^e	no	yes	yes ^d	no	yes ^f	yes	yes	no
polypropylene	yes	no	yes	yes	yes	yes ^g	yes ^h	yes	no
polyethylene	no	no	yes	yes	yes ⁱ	yes	yes	yes	yes
cellulose propionate	no	no	no	no	no	yes	yes	yes	no
stainless steel	yes	yes	yes	yes	yes ^j	no	yes	yes	no

- This information is provided as a guide to the use of sterilization and disinfection techniques for tube materials. Cold sterilization results shown are for short-duration (10-minute) soak periods; reactions may differ with extended contact. Refer to Appendix A of this manual for information about specific solutions.
- To avoid deformation, autoclave tubes or bottles open-end down in a tube rack at 15 psig for no more than 20 minutes (allow to cool before removing from tube rack). DO NOT autoclave capped or sealed tubes or bottles.
- Flammable; do not use in or near operating centrifuges.
- Do not use if there is methanol in the formula.
- Tube life will be reduced by autoclaving.
- Discoloration may occur.
- Can be used if diluted.
- Below 26°C only.
- Below 21°C only.
- Marginal.

JS-5.0 labware cups, cup covers, cup gaskets (369261 and 369257), and partitions can be autoclaved at 121°C for up to 20 minutes. Remove the plug and air-vent filter from each cup cover before autoclaving, and remove the gasket from the cup. To remove an air-vent filter, gently push it out from underneath the cover with a pencil or other non-metal tool that will not scratch the cover material. After autoclaving, insert a new air-vent filter into each cup cover. Thoroughly dry the gasket sealing surfaces before replacing the gasket.



Autoclaving will reduce the useful life of the labware cups, cup covers, cup gaskets, and partitions. After each autoclave cycle, examine these components for damage and DO NOT USE damaged components.

HarvestLine system liners can be gamma irradiated to a maximum dose of 40.0 kGy. Gamma irradiation causes the liners to become yellow, but does not affect their performance. *Do not steam or dry autoclave the liners or they will be damaged.* The liners are designed for single use only.

A cold sterilization method, such as immersion in 10% hydrogen peroxide for 30 minutes, may be used on Ultra-Clear tubes. Refer to [Table 7.1](#) to select cold sterilization materials that will not damage tubes and accessories.


While Beckman Coulter has tested these methods and found that they do not damage the components, no guarantee of sterility or disinfection is expressed or implied. When sterilization or disinfection is a concern, consult your laboratory safety officer regarding proper methods to use.

NOTE Multiwell plates can be purchased already sterilized.

Inspection


Inspect containers and accessories before use.

- Inspect tubes and bottles for cracks or any major deformities before using them.
- Do not use a tube that has become yellowed or brittle with age or excess exposure to ultraviolet light.
- Crazeing—the appearance of fine cracks on tubes and bottles—is the result of stress relaxation. If a crack approaches the outer wall of the tube or bottle, discard it.



- Discard any deformed or cracked adapters.

NOTE Replace the air-vent filter in each JS-5.0 cup cover after every 250 cycles, or after every autoclave cycle.



Tube and Bottle Storage

Tubes and bottles have an indefinite shelf life if properly stored. Store in a dark, cool, dry place away from ozone, chemical fumes, and ultraviolet light sources.

Returning a Rotor or Accessory to the Factory

Before returning a rotor or accessory for any reason, prior permission must be obtained from Beckman Coulter, Inc. The return authorization form may be obtained from your local Beckman Coulter sales office. The form, entitled Returned Material Authorization (RMA) for United States returns or *Returned Goods Authorization* (RGA) for international returns, should contain the following information:

- rotor type and serial number,
- history of use (approximate frequency of use),
- reason for the return,
- original purchase order number, billing number, and shipping number, if possible,
- name and email address of the person to be notified upon receipt of the rotor or accessory at the factory, and
- name and email address of the person to be notified about repair costs, etc.

To protect our personnel, it is the customer's responsibility to ensure that the parts are free from pathogens, chemical hazards, and/or radioactivity. Sterilization and decontamination **MUST** be done before returning the parts. Smaller items (such as tubes, bottles, and so on) should be enclosed in a sealed plastic bag.

All parts must be accompanied by a note, plainly visible on the outside of the box or bag, stating that they are safe to handle and that they are not contaminated with pathogens, chemical hazards, or radioactivity. Failure to attach this notification will result in return or disposal of the items without review of the reported problem.

Use the address label printed on the RMA/RGA form when mailing the rotor and/or accessories.

Customers located outside the United States should contact their local Beckman Coulter office.

Diagnostic Hints

Some of the more common operating problems experienced in centrifugation are listed in [Table 7.2](#) with suggestions for their solutions. Contact your Beckman Coulter Field Service representative if a problem cannot be corrected.

NOTE Use only the labware listed in the applicable rotor manual.

Table 7.2 Troubleshooting Chart

Symptom	Possible Cause and Suggested Action
Rotor	
Severe vibration	• Rotor imbalance. To balance the rotor load, fill all opposing tubes to the same level with liquid of the same density. Weight of opposing tubes must be distributed equally. Place tubes in a fixed-angle or vertical-tube rotor symmetrically, as illustrated in CHAPTER 1 (Figure 1.6). Detailed information about balancing swinging-bucket rotors is contained in CHAPTER 5. • Speed selected is within the rotor's critical speed range.^a Select a speed outside the critical speed range. (Refer to the applicable rotor manual for critical speed range.) • Rotor improperly tied-down. Make sure the rotor is properly secured to the drive spindle hub before centrifugation. If the rotor is left in the centrifuge between runs, tighten the tie-down device before each run. • Swinging-bucket rotor — Mishooked bucket, loose bucket cover, wrong type of bucket, mixed bucket types, opposing buckets not filled to the same level with liquids of the same density. Check loading procedures (refer to CHAPTER 5). • Swinging-bucket rotor—Pivot pin pockets not lubricated every 250 runs. Lubricate as described in Lubrication earlier in this chapter.
Rotor lid, canister cover, or bucket cover is difficult to remove after centrifugation	• Vacuum built up inside the container during centrifugation. Lift the vent plug on bucket or canister covers with vents to relieve the vacuum. • Threads contaminated with dirt, dried lubricant, or metal particles, or threads insufficiently lubricated cause rotor components to stick. Do not use excessive force to loosen components. Contact your Beckman Coulter representative. Routinely clean metal threads with concentrated Solution 555 that has been diluted to 10 parts water to 1 part detergent, then lubricate them with Spinkote.
Anodizing coming off where bucket or carrier contacts rotor pins on swinging-bucket rotor	Not an operational problem (some buckets are not anodized inside the pin pockets to facilitate swinging).
Adapter	
Adapters stick in buckets after centrifugation	Apply a thin film of powder, such as talcum powder, to the tube adapter rubber bases after cleaning or as required to prevent sticking.
Tube	
Tube leakage	
Tubes with cap assemblies	• Caps not properly secured. Caps must be properly seated on tubes and then fully tightened. • Cap components not dry before assembly. Thoroughly dry all components before assembling.

Table 7.2 Troubleshooting Chart (*Continued*)

Symptom	Possible Cause and Suggested Action
Tubes with snap-on caps	Tube too full; the meniscus must be kept lower to prevent leakage.
Uncapped tubes	Tube volume exceeds maximum uncapped volume. Refer to the rotor manual for tube volumes and speed reductions.
Quick-Seal tubes	Improperly sealed. After heat-sealing, squeeze the tube gently (if the tube contents may be disturbed) to test the seal for leaks. If the tube leaks, reseal it.
Tube cracking	- Tubes may crack or become brittle if they are used below their lower temperature limit. Before using tubes at other than stated temperature limits, evaluate them under centrifugation conditions. If sample is frozen in tubes, make sure that they are thawed to at least 2°C before centrifugation. - Tubes may become brittle with age and use. Dispose of brittle or cracked tubes.
Tube collapse	- Thinwall tube volume too low to provide tube wall support. Meniscus should be 2 to 3 mm below the tube top. Refer to the rotor manual for tube volumes. - Moisture between the tube and the cavity or bucket can cause the tube to float and collapse. Ensure that tubes and tube cavities or buckets are dry before inserting the tubes. - Reagent used that attacks the tube material. Refer to <i>Chemical Resistances</i> (IN-175) for chemical compatibilities of tube material and chemicals. - Tubes run above their rated speed. Refer to the applicable rotor manual for maximum speeds.
Bottle	
Bottle leakage (bottles with cap assemblies)	- Moisture or lubrication on cap or sealing surface. Ensure that the O-ring, plug, and bottle lip are dry and free of lubrication before use. - O-ring or gasket damaged or defective. Replace the O-ring or gasket. - Cap not tightened sufficiently. Tighten cap securely. - Sealing surface of the bottle is not smooth. Replace bottle. - Threaded caps without inserts or O-rings—Tube too full; these are not as liquid-tight as cap assemblies; therefore, the meniscus must be kept lower to prevent leakage.

Table 7.2 Troubleshooting Chart (*Continued*)

Symptom	Possible Cause and Suggested Action
Bottle leakage (uncapped bottles)	Bottle too full; the meniscus must be kept lower to prevent leakage. Refer to the rotor manual for fill volumes and speed reductions.
Bottle damage	• Fill volume too low to provide tube wall support. Refer to the rotor manual for fill volumes and speed reduction. • Moisture between the bottle and the cavity or bucket can cause the bottle to float and collapse. Ensure that bottles and cavities or buckets are dry before inserting them. • Reagent used that attacks the bottle material. Refer to <i>Chemical Resistances</i> (IN-175) for chemical compatibilities of bottle material and chemicals. • Bottles may crack or become brittle if they are used below their lower temperature limit. Before using bottles at other than stated temperature limits, evaluate them under centrifugation conditions. If sample is frozen in bottles, make sure that they are thawed to at least 2°C before centrifugation. • Bottles may become brittle with age and use. Dispose of brittle or cracked bottles. • Improper cleaning, decontamination, or sterilization procedures used. Refer to Table 7.1 for acceptable procedures and materials.

- a. Critical speed range is the range of speeds over which the rotor shifts so as to rotate about its center of mass. Passing through the critical speed range is characterized by some vibration.

Chemical Resistances for Beckman Coulter Centrifugation Products



List of Chemical Resistances

Appendix A is replicated in the separate pdf document,
Chemical Resistances (IN-175)

Temperature Compensation Tables

Introduction

This Appendix contains tables listing temperature compensation units for various rotors used in Beckman Coulter discontinued J2 and J6 series centrifuges.

Temperature Compensation

To ensure that the rotor reaches the required temperature during centrifugation, some temperature compensation may be required because of the mass of these rotors. The following tables list temperature compensation units for various rotors.

Refer to [CHAPTER 1](#) or the applicable rotor manual for procedures to set the temperature compensation for the model of J centrifuge being used.

NOTE When using an Avanti J series centrifuge, enter the run temperature according to the instructions in your centrifuge instruction manual. No additional input is required.

Table B.1 Temperature Compensation Settings for the J2-HC Centrifuge^a

Rotor	Speed (rpm)	Required Sample temperature (°C, green bar)						
		-20°C	-10°C	2°C	5°C	10°C	20°C	40°C
JA-25.50	18,000	N ^b	-8	-6	-6	-6	-6	-5
	15,000	-7	-6	-5	-5	-5	-4	-4
	10,000	-5	-3	-2	-3	-4	-3	-2
JA-25.15	18,000	-8	-7	-5	-6	-6	-5	+5
	15,000	-6	-5	-4	-4	-3	-4	+6
	10,000	-3	-2	-1	-2	-1	-1	+9
JA-21	18,000	N	-8	-8	-7	-6	-5	-3
	15,000	-6	-6	-5	-4	-4	-3	N
	10,000	-4	-2	-1	-1	-1	-1	N
JA-20.1	18,000	N	N	-7	-7	-7	-6	-4
	15,000	-7	-5	-5	-4	-4	-3	-2
	10,000	-4	-2	-1	-1	-1	-1	N

Table B.1 Temperature Compensation Settings for the J2-HC Centrifuge^a (Continued)

Rotor	Speed (rpm)	Required Sample temperature (°C, green bar)						
		-20°C	-10°C	2°C	5°C	10°C	20°C	40°C
JA-20	18,000	N	N	-6	-6	-5	-4	-3
	15,000	-6	-5	-4	-4	-3	-3	-2
	10,000	-4	-4	-3	-2	-2	-2	-2
JA-18.1 ^c	16,000	N	N	-8	-5	-4	-3	0
	15,000	N	N	-2	-2	-2	-2	0
	10,000	-4	-3	-3	-3	-3	-2	N
	8,000	-1	-2	-4	-3	-2	-1	N
JA-18	18,000	N	N	N	N	N	-8	-4
	17,000	N	N	N	-10	-9	-8	-3
	16,000	N	N	-10	-9	-8	-7	-2
	12,000	N	-6	-6	-5	-4	-3	-1
	8,000	-3	-2	-2	-1	-1	0	N
	5,000	-1	-1	-1	-1	0	+1	N
JA-17	17,000	N	N	-6	-6	-6	-5	-3
	15,000	N	-5	-4	-4	-4	-3	-2
	12,000	-4	-4	-3	-3	-3	-2	N
	8,000	-2	-2	-2	-2	-2	-1	N
JLA-16.250	14,000	N	N	7	-7	-7	-5	-4
	10,000	N	-4	-3	-3	-2	-1	N
	5,000	N	0	0	-3	-3	-3	N
JA-14	14,000	N	N	-5	-5	-5	-3	-2
	12,000	N	-3	-3	-3	-3	-2	-1
	10,000	-4	-2	-2	-2	-1	-1	N
	5,000	-1	0	0	0	0	0	N
JA-12	12,000	N	-3	-2	-2	-2	-2	-1
	10,000	-2	-2	-1	-1	-1	-1	-1
	5,000	-1	-1	0	0	0	0	0
JA-10	10,000	N	-3	-2	-1	0	+1	+2
	8,000	-3	-2	0	0	0	+1	+2
	5,000	-2	0	0	0	0	+1	+2
JLA-10.500	10,000	N	-6 ^d	-3	-3	-3	-1	0
	8,000	-3 [†]	-3	-3	-2	-2	-1	0
	5,000	-3 [†]	-2	-2	-1	0	0	+2
JS-13.1	13,000	N	-6	-9	-9	-9	-6	-3
	11,000	N	-5	-5	-5	-5	-4	-1
	8,000	-5	-3	-2	-2	-1	-1	N
	5,000	-4	-2	-1	-1	-1	0	N
JS-7.5	7,500	0	+2	+3	+3	+4	+4	+4
	5,000	0	+1	+2	+2	+2	+3	+3
	2,000	0	0	0	+1	+1	+2	N

Table B.1 Temperature Compensation Settings for the J2-HC Centrifuge^a (Continued)

Rotor	Speed (rpm)	Required Sample temperature (°C, green bar)						
		-20°C	-10°C	2°C	5°C	10°C	20°C	40°C
JS-4.3	4,300	0	0	0	0	0	0	+2
	3,000	0	0	0	0	0	0	+2
	1,500	0	0	0	0	0	0	N
JV-20	18,000	N	N	-9	-8	-7	-6	-6
	15,000	N	-7	-5	-5	-5	-3	-4
	10,000	-4	-3	-3	-3	-3	-1	0

- a. Interplate if intermediate values are required.
- b. N: indicates that the rotor cannot achieve the required temperature at this speed.
- c. **CAUTION:** for proper temperature control the JA-18.1 fixed-angle rotor must be derated in the J2-HC Centrifuge as follows: when the 25-degree-angle adapters are used, the maximum speed is 16,000 rpm; when the 45-degree-angle adapters are used, the maximum speed is 15,000 rpm.
- d. Above 30°C ambient temperature, this temperature may not be achieved at this speed.

Table B.2 Temperature Compensation Settings for the J2-21, J2-21B, J2-21C, and J2-HS Centrifuges^a

Rotor	Speed (rpm)	Required Sample temperature (°C, green bar)						
		-20°C	-10°C	2°C	5°C	10°C	20°C	40°C
JA-25.50	20,000	N ^b	N	-7	-5	-6	-5	-9
	18,000	N	-5	-5	-5	-5	-4	-7
	15,000	0	-2	-3	-3	-2	-2	-6
	10,000	0	0	-1	0	-1	0	-3
JA-25.15	21,000	N	N	-10	-10	-7	-6	-9
	18,000	N	-8	-8	-7	-4	-5	-5
	15,000	-6	-6	-5	-4	-4	-3	-2
	10,000	-3	-2	-1	-1	-1	-1	-2
JA-21	21,000	N	-8	-10	-10	-10	-8	-6
	18,000	N	-8	-8	-7	-6	-5	-3
	15,000	-6	-6	-5	-4	-4	-3	N
	10,000	-3	-2	-1	-1	-1	0	N
JA-20.1	20,000	N	N	-10	-9	-9	-9	-7
	18,000	N	N	-7	-7	-7	-6	-4
	15,000	-7	-5	-4	-4	-4	-3	N
	10,000	-2	-1	-1	0	0	0	N
JA-20	20,000	N	N	-9	-8	-7	-7	-6
	18,000	N	N	-6	-6	-5	-4	-3
	15,000	-5	-4	-4	-4	-3	-2	-1
	10,000	-2	-2	-1	0	0	0	0
JA-18.1	18,000	N	N	N	N	-8	-5	0
	17,000	N	N	N	N	-6	-5	0
	15,000	N	N	-9	-5	-4	-3	0
	10,000	-7	-7	-3	-3	-2	0	+2
	8,000	-6	-5	-3	-3	-1	0	+2

Table B.2 Temperature Compensation Settings for the J2-21, J2-21B, J2-21C, and J2-HS Centrifuges^a (Continued)

Rotor	Speed (rpm)	Required Sample temperature (°C, green bar)						
		–20°C	–10°C	2°C	5°C	10°C	20°C	40°C
JA-18	18,000	N	N	N	N	N	–10	–7
	17,000	N	N	N	–10	–9	–8	–6
	16,000	N	N	–10	–10	–9	–7	–5
	12,000	N	–6	–6	–6	–5	–4	–3
	8,000	–5	–4	–4	–3	–2	–1	0
	5,000	–4	–3	–2	–2	–1	0	0
JA-17	17,000	N	N	–7	–7	–7	–4	–2
	15,000	N	–5	–4	–4	–4	–3	–1
	12,000	–4	–4	–3	–3	–3	–1	N
	8,000	–1	–1	0	0	0	0	N
JLA-16.250	14,000	N	N	–7	–5	–4	–3	–3
	10,000	N	–4	–3	–3	–2	–1	N
	5,000	N	0	0	0	0	0	N
JA-14	14,000	N	N	–7	–7	–7	–5	–4
	12,000	N	–6	–5	–5	–4	–3	–2
	10,000	–4	–4	–3	–3	–2	–1	N
	5,000	–1	0	0	0	0	0	N
JA-12	12,000	N	–6	–6	–7	–7	–6	–9
	10,000	–3	–4	–5	–5	–5	–5	–9
	5,000	–1	–3	–3	–4	–4	–5	–9
JA-10	10,000	N	–3	–2	–1	0	+1	+2
	8,000	–3	–2	–1	–1	0	+1	+2
	5,000	–2	0	0	0	0	+1	+2
JLA-10.500	10,000	N	–5 ^c	–2	–1	+0	+2	+1
	8,000	–1 ^c	–1 ^c	+0	+2	+2	+3	+1
	5,000	–0 ^c	–1	+2	+1	+4	+3	+1
JS-13.1	13,000	N	–10	–10	–10	–10	–10	–9
	11,000	N	–8	–8	–8	–7	–6	–5
	8,000	–6	–5	–4	–4	–3	–3	N
	5,000	–4	–3	–1	–1	–1	–1	N
JS-7.5	7,500	–4	–3	+1	+1	+1	+2	+3
	5,000	–2	0	+3	+3	+3	+3	+4
	2,000	0	+2	+4	+4	+4	+4	+5
JV-20	20,000	N	N	–max	–max	–max	–max	–max
	18,000	N	N	–max	–max	–10	–10	–9
	15,000	N	N	–7	–7	–6	–5	–4
	10,000	N	N	–2	–2	–1	0	0

a. Interplate if intermediate values are required.

b. N: indicates that the rotor cannot achieve the required temperature at this speed.

c. Above 30°C ambient temperature, this temperature may not be achievable at this speed.

Table B.3 Temperature Compensation Settings for the J2-MI, J2-21M, J2-MC, and J2-21M/E Centrifuges

Rotor	Speed (rpm)	Required Sample Temperature (°C, green bar)						
		-20°C	-10°C	2°C	5°C	10°C	20°C	40°C
JA-25.50	20,000	N ^a	N	-7	-5	-6	-5	-9
	18,000	N	-5	-5	-5	-5	-4	-7
	15,000	0	-2	-3	-3	-2	-2	-6
	10,000	0	0	-1	0	-1	0	-3
JA-25.15	21,000	N	N	-8	-8	-7	-6	-9
	18,000	N	-8	-8	-8	-4	-5	-5
	15,000	-6	-6	-5	-4	-4	-3	-2
	10,000	-3	-2	-1	-1	-1	-1	-2
JA-21	21,000	N	N	-10	-10	-10	-8	-6
	18,000	N	-8	-8	-7	-6	-5	-3
	15,000	-6	-6	-5	-4	-4	-3	N
	10,000	-3	-2	-1	-1	-1	0	N
JA-20.1	20,000	N	N	-10	-9	-9	-9	-7
	18,000	N	N	-7	-7	-7	-6	-4
	15,000	-7	-5	-4	-4	-4	-3	N
	10,000	-2	-1	-1	0	0	0	N
JA-20	20,000	N	N	N	11	10	9	-6
	18,000	N	N	8	7.5	7	6	-3
	15,000	-5	-4	4	4	3	3	-1
	10,000	-2	-2	2	1	1	0	0
JA-18.1	18,000	N	N	N	N	-8	-5	0
	17,000	N	N	N	N	-6	-5	0
	15,000	N	N	-9	-5	-4	-3	0
	10,000	-7	-7	-3	-3	-2	0	+2
	8,000	-6	-5	-3	-3	-1	0	+2
JA-18	18,000	N	N	N	N	N	-10	-7
	17,000	N	N	N	-10	-9	-8	-6
	16,000	N	N	-10	-10	-9	-7	-5
	12,000	N	-6	-6	-6	-5	-4	-3
	8,000	-5	-4	-4	-3	-2	-1	0
	5,000	-4	-3	-2	-2	-1	0	0
JA-17	17,000	N	N	-7	-7	-7	-4	-2
	15,000	N	-5	-4	-4	-4	-3	-1
	12,000	-4	-4	-3	-3	-3	-1	N
	8,000	-1	-1	0	0	0	0	N
JLA-16.250	14,000	N	N	-7	-5	-4	-3	-3
	10,000	N	-4	-3	-3	-2	-1	N
	5,000	N	0	0	0	0	0	N
JA-14	14,000	N	N	-7	-7	-7	-5	-4
	12,000	N	-6	-5	-5	-4	-3	-2
	10,000	-4	-4	-3	-3	-2	-1	N
	5,000	-1	0	0	0	0	0	N

Table B.3 Temperature Compensation Settings for the J2-MI, J2-21M, J2-MC, and J2-21M/E Centrifuges (*Continued*)

Rotor	Speed (rpm)	Required Sample Temperature (°C, green bar)						
		-20°C	-10°C	2°C	5°C	10°C	20°C	40°C
JA-12	12,000	N	-6	-6	-7	-7	-6	-9
	10,000	-3	-4	-5	-5	-5	-5	-9
	5,000	-1	-3	-3	-4	-4	-5	-9
JA-10	10,000	N	-3	-2	-1	0	+1	+2
	8,000	-3	-2	-1	-1	0	+1	+2
	5,000	-2	0	0	0	0	+1	+2
JLA-10.500	10,000	N	-5 ^b	-2	-1	0	+2	+1
	8,000	-1 ^b	-1 ^b	0	+2	+2	+3	+1
	5,000	-0*	-1	+2	+1	+4	+3	+1
JS-13.1	13,000	N	-10	-10	-10	-10	-10	-9
	11,000	N	-8	-8	-8	-7	-6	-5
	8,000	-6	-5	-4	-4	-3	-3	N
	5,000	-4	-3	-1	-1	-1	-1	N
JS-7.5	7,500	-4	-3	+1	+1	+1	+2	+3
	5,000	-2	0	+3	+3	+3	+3	+4
	2,000	0	+2	+4	+4	+4	+4	+5
JV-20	20,000	N	N	N	N	-max	-max	-max
	18,000	N	N	N	-max	-10	-10	-9
	15,000	N	N	-7	-7	-6	-5	-4
	10,000	N	N	-2	-2	-1	0	0

a. N: indicates that the rotor cannot achieve the required temperature at this speed.

b. Above 30°C ambient temperature, this temperature may not be achievable at this speed.

Table B.4 Temperature Compensation Settings for the J6 Centrifuges

Rotor	Speed (rpm)	Required Sample Temperature (°C, green bar)						
		2°C	4°C	8°C	10°C	15°C	20°C	30°C
JS-5.2	5200	-3	-1	4	6	10	17	28
	4000	-1	1	5	7	12	19	29
	3000	0	2	7	9	14	20	30
	2000 and below	2	4	8	10	15	20	30
JS-4,2, JS-4.2SM, JS-4.2A, JS-4.2SMA	4200	-3	0	5	7	13	19	30
	3000	0	2	7	9	14	20	30
	2000 and below	2	4	8	10	15	20	30
All other rotors	all speeds	2	4	8	10	15	20	30

Gradient Materials

Introduction

This Appendix contains reference information on commonly used gradient materials. General instructions for filling and sealing tubes, including gradient preparation, are contained in [CHAPTER 3](#).

Gradient material selection depends on a number of factors, including the type of separation to be performed. Sucrose is used for rate zonal and isopycnic separations, and cesium chloride is often used for isopycnic separations. The basic requirement is that the gradient permit the type of separation. Additional considerations in selecting a gradient material include the following.

- Its density range should be sufficient to permit separation of the particles of interest by the chosen density gradient technique, without overstressing the rotor.
- It should not affect the biological activity of the sample.
- It should be neither hyperosmotic or hypoosmotic when the sample is composed of sensitive organelles.
- It should not interfere with the assay technique.
- It should be removable from the purified product.
- It should not absorb in the ultraviolet or visible range.
- It should be inexpensive and readily available; more expensive materials should be recoverable for reuse.
- It should be sterilizable.
- It should not be corrosive to the rotor.
- It should not be flammable or toxic to the extent that its aerosols could be hazardous.

The following charts are provided as a reference for information on commonly used gradient materials.

Table C.1 Commonly Used Gradient Materials with Their Solvents

Materials	Solvent	Maximum Density at 20°C
Sucrose (66%)	H ₂ O	1.32
Sucrose (65%)	D ₂ O	1.37
Silica sols	H ₂ O	1.30
Diodon	H ₂ O	1.37
Glycerol	H ₂ O	1.26
Cesium chloride	H ₂ O	1.91
	D ₂ O	1.98
Cesium formate	H ₂ O	2.010
Cesium acetate	H ₂ O	2.00
Rubidium chloride	H ₂ O	1.49
Rubidium formate	H ₂ O	1.85
Rubidium bromide	H ₂ O	1.63
Potassium acetate	H ₂ O	1.41
Potassium formate	H ₂ O	1.57
	D ₂ O	1.63
Sodium formate	H ₂ O	1.32
	D ₂ O	1.40
Lithium bromide	H ₂ O	1.83
Lithium chloride	D ₂ O	1.33
Albumin	H ₂ O	1.35
Sorbitol	H ₂ O	1.39
Ficoll	H ₂ O	1.17
Metrizamide	H ₂ O	1.46

Table C.2 Density, Refractive Index, and Concentration Data—Cesium Chloride at 25°C, Molecular Weight = 168.37^a

Density (g/cm ³) ^b	Refractive Index, η_D	% by Weight	mg/mL of Solution ^c	Molarity	Density (g/cm ³) ^a	Refractive Index, η_D	% by Weight	mg/mL of Solution ^b	Molarity
1.0047	1.3333	1	10.0	0.056	1.336	1.3657	34	454.2	2.698
1.0125	1.3340	2	20.2	0.119	1.3496	1.3670	35	472.4	2.806
1.0204	1.3348	3	30.6	0.182	1.363	1.3683	36	490.7	2.914
1.0284	1.3356	4	41.1	0.244	1.377	1.3696	37	509.5	3.026
1.0365	1.3364	5	51.8	0.308	1.391	1.3709	38	528.6	3.140
1.0447	1.3372	6	62.8	0.373	1.406	1.3722	39	548.3	3.257
1.0531	1.3380	7	73.7	0.438	1.4196	1.3735	40	567.8	3.372
1.0615	1.3388	8	84.9	0.504	1.435	1.3750	41	588.4	3.495
1.0700	1.3397	9	96.3	0.572	1.450	1.3764	42	609.0	3.617
1.0788	1.3405	10	107.9	0.641	1.465	1.3778	43	630.0	3.742
1.0877	1.3414	11	119.6	0.710	1.481	1.3792	44	651.6	3.870
1.0967	1.3423	12	131.6	0.782	1.4969	1.3807	45	673.6	4.001
1.1059	1.3432	13	143.8	0.854	1.513	1.3822	46	696.0	4.134
1.1151	1.3441	14	156.1	0.927	1.529	1.3837	47	718.6	4.268
1.1245	1.3450	15	168.7	1.002	1.546	1.3852	48	742.1	4.408
1.1340	1.3459	16	181.4	1.077	1.564	1.3868	49	766.4	4.552
1.1437	1.3468	17	194.4	1.155	1.5825	1.3885	50	791.3	4.700
1.1536	1.3478	18	207.6	1.233	1.601	1.3903	51	816.5	4.849
1.1637	1.3488	19	221.1	1.313	1.619	1.3920	52	841.9	5.000
1.1739	1.3498	20	234.8	1.395	1.638	1.3937	53	868.1	5.156
1.1843	1.3508	21	248.7	1.477	1.658	1.3955	54	893.3	5.317
1.1948	1.3518	22	262.9	1.561	1.6778	1.3973	55	922.8	5.481
1.2055	1.3529	23	277.3	1.647	1.699	1.3992	56	951.4	5.651
1.2164	1.3539	24	291.9	1.734	1.720	1.4012	57	980.4	5.823
1.2275	1.3550	25	306.9	1.823	1.741	1.4032	58	1009.8	5.998
1.2387	1.3561	26	322.1	1.913	1.763	1.4052	59	1040.2	6.178
1.2502	1.3572	27	337.6	2.005	1.7846	1.4072	60	1070.8	6.360
1.2619	1.3584	28	353.3	2.098	1.808	1.4093	61	1102.9	6.550
1.2738	1.3596	29	369.4	2.194	1.831	1.4115	62	1135.8	6.746
1.2858	1.3607	30	385.7	2.291	1.856	1.4137	63	1167.3	6.945
1.298	1.3619	31	402.4	2.390	1.880	1.4160	64	1203.2	7.146
1.311	1.3631	32	419.5	2.492	1.9052	1.4183	65	1238.4	7.355
1.324	1.3644	33	436.9	2.595					

a. Density data are from International Critical Tables.

b. Computed from the relationship $p^{25} = 10.2402 \eta_D^{25} - 12.6483$ for densities between 1.00 and 1.37, and $p^{25} = 10.8601 \eta_D^{25} - 13.4974$ for densities above 1.37 (Bruner and Vinograd, 1965).

c. Divide by 10.0 to obtain % w/v.

Table C.3 Density, Refractive Index, and Concentration Data—Sucrose at 20°C, Molecular Weight = 342.3^a

Density (g/cm ³)	Refractive Index, η_D	% by Weight	mg/mL of Solution ^b	Molarity	Density (g/cm ³) ^a	Refractive Index, η_D	% by Weight	mg/mL of Solution ^b	Molarity
0.9982	1.3330	0			1.1463	1.3883	34	389.7	1.138
1.0021	1.3344	1	10.0	0.029	1.1513	1.3902	35	403.0	1.177
1.0060	1.3359	2	20.1	0.059	1.1562	1.3920	36	416.2	1.216
1.0099	1.3374	3	30.3	0.089	1.1612	1.3939	37	429.6	1.255
1.0139	1.3388	4	40.6	0.119	1.1663	1.3958	38	443.2	1.295
1.0179	1.3403	5	50.9	0.149	1.1713	1.3978	39	456.8	1.334
1.0219	1.3418	6	61.3	0.179	1.1764	1.3997	40	470.6	1.375
1.0259	1.3433	7	71.8	0.210	1.1816	1.4016	41	484.5	1.415
1.0299	1.3448	8	82.4	0.211	1.1868	1.4036	42	498.5	1.456
1.0340	1.3464	9	93.1	0.272	1.1920	1.4056	43	512.6	1.498
1.0381	1.3479	10	103.8	0.303	1.1972	1.4076	44	526.8	1.539
1.0423	1.3494	11	114.7	0.335	1.2025	1.4096	45	541.1	1.581
1.0465	1.3510	12	125.6	0.367	1.2079	1.4117	46	555.6	1.623
1.0507	1.3526	13	136.6	0.399	1.2132	1.4137	47	570.2	1.666
1.0549	1.3541	14	147.7	0.431	1.2186	1.4158	48	584.9	1.709
1.0592	1.3557	15	158.9	0.464	1.2241	1.4179	49	599.8	1.752
1.0635	1.3573	16	170.2	0.497	1.2296	1.4200	50	614.8	1.796
1.0678	1.3590	17	181.5	0.530	1.2351	1.4221	51	629.9	1.840
1.0721	1.3606	18	193.0	0.564	1.2406	1.4242	52	645.1	1.885
1.0765	1.3622	19	204.5	0.597	1.2462	1.4264	53	660.5	1.930
1.0810	1.3639	20	216.2	0.632	1.2519	1.4285	54	676.0	1.975
1.0854	1.3655	21	227.9	0.666	1.2575	1.4307	55	691.6	2.020
1.0899	1.3672	22	239.8	0.701	1.2632	1.4329	56	707.4	2.067
1.0944	1.3689	23	251.7	0.735	1.2690	1.4351	57	723.3	2.113
1.0990	1.3706	24	263.8	0.771	1.2748	1.4373	58	739.4	2.160
1.1036	1.3723	25	275.9	0.806	1.2806	1.4396	59	755.6	2.207
1.1082	1.3740	26	288.1	0.842	1.2865	1.4418	60	771.9	2.255
1.1128	1.3758	27	300.5	0.878	1.2924	1.4441	62	788.3	2.303
1.1175	1.3775	28	312.9	0.914	1.2983	1.4464	62	804.9	2.351
1.1222	1.3793	29	325.4	0.951	1.3043	1.4486	63	821.7	2.401
1.1270	1.3811	30	338.1	0.988	1.3103	1.4509	64	838.6	2.450
1.1318	1.3829	31	350.9	1.025	1.3163	1.4532	65	855.6	2.500
1.1366	1.3847	32	363.7	1.063	1.3224	1.4558	66	872.8	2.550
1.1415	1.3865	33	376.7	1.100	1.3286	1.4581	67	890.2	2.864

a. Density data are from International Critical Tables.

b. Divide by 10.0 to obtain % w/v.

Table C.4 Density Conversion for Cesium and Rubidium Salts at 20°C

% w/w	CsCl	CsBr	CsI	Cs ₂ SO ₄	CsNO ₃	RbCl	RbBr	RbI	Rb ₂ SO ₄	RbNO ₃
1	1.00593	1.00612	1.00608	1.0061	1.00566	1.00561	1.00593	1.00591	1.0066	1.0053
2	1.01374	1.01412	1.01402	1.0144	1.01319	1.01307	1.01372	1.01370	1.0150	1.0125
4	1.02969	1.03048	1.03029	1.0316	1.02859	1.02825	1.02965	1.02963	1.0322	1.0272
6	1.04609	1.04734	1.04707	1.0494	1.04443	1.04379	1.04604	1.04604	1.0499	1.0422
8	1.06297	1.06472	1.06438	1.0676	1.06072	1.05917	1.06291	1.06296	1.0680	1.0575
10	1.08036	1.08265	1.08225	1.0870	1.07745	1.07604	1.08028	1.08041	1.0864	1.0731
12	1.09828	1.10116	1.10071	1.1071	1.09463	1.09281	1.09817	1.09842	1.1052	1.0892
14	1.11676	1.12029	1.11979	1.1275	1.11227	1.11004	1.11661	1.11701	1.1246	1.1057
16	1.13582	1.14007	1.13953	1.1484		1.12775	1.13563	1.13621	1.1446	1.1227
18	1.15549	1.16053	1.15996	1.1696		1.14596	1.15526	1.15605	1.1652	1.1401
20	1.17580	1.18107	1.18112	1.1913		1.16469	1.17554	1.17657	1.1864	1.1580
22	1.19679	1.20362	1.20305	1.2137		1.18396	1.19650	1.19781	1.2083	1.1763
24	1.21849	1.22634	1.22580	1.2375		1.20379	1.21817	1.21980	1.2309	1.1952
26	1.24093	1.24990	1.24942	1.2643		1.22421	1.24059	1.24257	1.2542	1.2146
28	1.26414	1.27435	1.27395			1.24524	1.26380	1.26616	1.2782	1.2346
30	1.28817	1.29973	1.29944			1.26691	1.28784	1.29061	1.3028	1.2552
35	1.35218	1.36764	1.36776			1.32407	1.35191	1.35598	1.3281	1.2764
40	1.42245	1.44275	1.44354			1.38599	1.42233	1.42806		
45	1.49993	1.52626	1.52803			1.45330	1.50010	1.50792		
50	1.58575	1.61970	1.62278			1.52675	1.58639	1.59691		
55	1.68137	1.72492					1.68254	1.69667		
60	1.78859							1.80924		
65	1.90966							1.93722		

Blood Component Separation

Introduction

This Appendix provides a basic overview of blood separation procedures using Beckman Coulter J series centrifuges.

Blood Bank Collection Overview

Blood is composed of plasma, red blood cells (RBC), white blood cells (WBC), and platelets. Approximately 40 to 45 percent of this volume is made up of red blood cells; most of the remainder is plasma, a watery substance that contains vital substances, including hormones and proteins.

Most whole blood collected undergoes fractionation, or separation into components, in order to use collected blood most efficiently. Termed “component therapy,” multiple use of different parts of the blood helps conserve this scarce resource and allows patients to receive only the components they need. As shelf life and storage requirements vary, conditions can be optimized by separating components.

In addition to collection of whole blood for separation into components, several techniques involve collection of whole blood, separation, collection of a fraction for infusion, and return of the remainder to the donor.

- In plasmapheresis a unit of blood is taken to obtain plasma, separated, and the red cells are immediately reinfused to the donor. Most plasmapheresis is performed for “source plasma,” which is not intended for intravenous transfusion, but separated by large-scale fractionation into clotting factors (especially factor VIII), albumin, and specific immunoglobulins.
- During plateletpheresis, whole blood is collected, platelets separated via centrifugations, and platelet-depleted red blood cells returned to the donor. The plasma may be returned to the donor or collected for fractionation into clotting factors and albumin.
- Leukapheresis is the separation of leukocytes, or white blood cells, from whole blood. The leukocyte-depleted and platelet-depleted red blood cells are continuously or intermittently returned to the donor.

Components and Typical Usage

Single-Donor Fresh Plasma

Single-donor fresh plasma is separated from whole blood within 4 to 8 hours after collection. If it is not used immediately, it may be frozen and stored (see [Shelf Life](#), below). Fresh and fresh-frozen plasma contain all plasma-clotting factors.

- Fresh plasma—used for the treatment of deficiencies of clotting factors V, XI, and XIII.
- Factor VII Concentrate—separated from platelet-rich plasma, useful for treating clotting deficiencies other than those mentioned above.
- Cryoprecipitated Antihemophilic Factor (factor VIII)—a protein concentrate separated from cell-free plasma (frozen, then thawed at 4°C), useful for treating hemophilia.
- Platelet Concentrates—separated from plasma, platelet concentrates are used to treat decreased platelet counts or failing platelet functions. Platelets must be separated by centrifugation within 6 hours of collection.

Single-Donor Plasma

Single-donor plasma can be separated from whole blood up to a few days after the expiration date, since no attempt is made to maintain the activity of the labile clotting factors. This plasma, which may be frozen and kept for up to 5 years, is used for expansion of blood volume (treatment of hypovolemic shock, caused by a dangerous reduction in blood volume).

Packed Red Blood Cells (PBC)

RBC, required when the oxygen-carrying capacity must be improved without overloading the cardiovascular system with extra fluid volume, are commonly administered to treat anemia.

Leukocyte-Depleted Red Blood Cells

These are prepared by removing most leukocytes and platelets from fresh whole blood. Obtained by differential separation, they are given to recipients with antileukocyte antibodies to prevent adverse reactions.

Shelf Life

Table D.1 lists approximate storage times for the separated components.

Table D.1 Blood Component Storage

Component	Approximate Storage Life
RBC with ADSOL ^a	42 days
RBC with SAG-M ^b	35 days
RBC with CPDA-1 ^c	25 days
RBC without preservative	21 days
RBC frozen without addition of cryoprotective agent	10 days
Plasma—frozen	1 year

a. ADSOL = dextrose-sodium chloride-mannitol-adenine.

b. SAG-M = saline-adenine-glucose-mannitol.

c. CPDA-1 = citrate-phosphate-dextrose-citrate-citric acid-adenine.

Freezing

A cryoprotective agent such as glycerol can be added to extend the life of frozen red blood cells. These cells can then be stored for up to 3 years at -80°C . Prior to use the cells are thawed and the glycerol is removed by washing.

If the plasma will not be separated within 15 hours of collection, it must be frozen within 6 hours of collection. Freezing must be carried out in a flash freezer with complete freezing accomplished within 1 or 2 hours of the time it is placed in the freezer.


Separation of Blood Components by Centrifugation

Centrifugation is the primary method for processing blood because it offers the required high throughput, reproducibility, and versatility. Most blood components can be separated in one or two runs. Generally, two types of centrifugation runs are performed (see Figure D.1). Soft spin runs, short centrifugation runs (3 to 5 minutes) at low g -forces (2000 to $3000 \times g$) at ambient temperature, are used to keep small cells or platelets in suspension while the larger cells sediment. This type of run is used to obtain platelet-rich plasma and red blood cell concentrate from whole blood.

Hard spin runs are longer (5 to 7 minutes), at higher g -forces (4000 to $5000 \times g$), at ambient temperature or at 4°C , and are used to separate fresh plasma from cellular components. Soft spin and hard spin techniques are often combined.

Blood Component Separation


Separation of Blood Components by Centrifugation



Donor blood is collected in plastic bags with one or more satellite bags (double, triple, or quad packs) containing anticoagulant and preservative. After each centrifugation run, the sedimented fraction is squeezed into its respective satellite bag. Common anticoagulants and preservatives include citrate-phosphate-dextrose (CPD), citrate-phosphate-dextrose-citrate-citric acid-adenine (CPDA-1), saline-adenine-glucose-mannitol (SAG-M), and dextrose-sodium chloride-mannitol-adenine (ADSOL).

Blood separations occur during centrifugation because of particle sedimentation. Using sedimentation theory, users can calculate sedimentation rates. For example, red blood cells settle at the approximate rate of 2 cm per hour in aqueous medium at 1 g, with higher force fields increasing the settling rate. Note that blood cells should not be subjected to high centrifugal force fields, as the cells can be damaged.

Figure D.1 Blood Component Separation



Beckman Coulter has centrifuges, rotors, and accessories designed to fit the special needs of blood component processing. Several rotors are available to accommodate single, double, triple, and quad blood bags. Blood bag cups rest in the rotor bucket and simplify processing, since they eliminate the need to remove buckets after each run. They also minimize clean-up downtime if a bag breaks—simply remove the cup and resume the run. Refer to the applicable rotor manual for blood bag cups used with each rotor. [Table D.2](#) lists blood bank methods that can be used for separating components in a variety of J6 series centrifuges.

Table D.2 Blood Bank Methods^a

Blood Components Products	Starting Material	Method	Temperature (°C)	Rotor	Capacity (No. of Cups)	Speed (rpm)	Time ^b (min)	Brake Setting ^c		
								J6-MI	J6-MC	J6-HC
Platelet-Rich Plasma and Red Blood Cells	Whole Blood	Soft Spin ^d	20 to 22	JS-4.2	6	2800	3.0	6	1	4.5–5.5
				JS-4.2SM	6	2850	3.0	6 or 7	1	4.5–5.5
				JS-5.2	4	3000	2.9	6	1	4.5–5.5
Platelet Concentrate	Platelet Rich Plasma	Hard Spin ³	20 to 22	JS-4.2	6	3850	6.0	6	1	4.5–5.5
				JS-4.2SM	6	3900	6.0	6 or 7	1	4.5–5.5
				JS-5.2	4	4100	6.0	6	1	4.5–5.5
Plasma (Cell-free) and Packed Cells	Whole Blood	Hard Spin	4	JS-4.2	6	3850	6.0	6	1	4.5–5.5
				JS-4.2SM	6	3900	6.0	6 or 7	1	4.5–5.5
				JS-5.2	4	4100	6.0	6	1	4.5–5.5
Cryoprecipitated Antihemophilic Factor	Plasma (Cell-free) (frozen, and then thawed at 4°C)	Hard Spin	4	JS-4.2	6	4200	7.2	Max.	Max.	Max.
				JS-4.2SM	6	4200	7.5	Max.	Max.	Max.
				JS-5.2	4	4500	7.1	Max.	Max.	Max.

a. The speeds, times, and brake settings shown here are intended to be guidelines only. Optimum conditions for separating blood components in each centrifuge must be determined by the user before carrying out actual separation runs.

b. Times include acceleration and time at maximum speed only. Deceleration time is not included.

c. Brake settings are estimated for a rotor fully loaded with 500-mL blood bags. When using other bags, brake settings should be increased to maintain comparable deceleration times.

d. Several methods for preparation of platelet-rich plasma and platelet concentrate are in current use. The speed and time ranges given have been estimated to be comparable to conditions specified in the following sources:

American Association of Blood Banks. *Technical Manual*, p. 359. 7th ed. Washington, 1977.

Humphreys, P. Private communication. Canadian Red Cross Society, Toronto, Ontario. (July 1977)

Kahn, R.A., Cossette, I., Friedman, L. I. *Transfusion* 16. 162-165 (1976)

Reiss, R. F., Katz, A. J. *Transfusion* 16, 370-374 (1976)

Slichter, S. J., Harker, L. A. *Transfusion* 16, 8-12 (1976)

Tips for Optimum Centrifugation Runs

Centrifugation generates high speeds, causing rotor heads and buckets to develop a gravity force of thousands of pounds. Observing the following tips will ensure safe and efficient operation.

 WARNING

Handle body fluids with care because they can transmit disease. No known test offers complete assurance that body fluids are free of micro-organisms. Some of the most virulent—Hepatitis (B and C) and HIV (I-V) viruses, atypical mycobacterium, and certain systemic fungi—further emphasize the need for aerosol protection.

- Never lift the centrifuge door while the instrument is running.
- Keep metal clips, needle holders, and sealed tube ends away from blood bags.
- Load opposing cups with equal weight, to ensure safety, optimum run efficiency, and long rotor life.
- Use weighted rubber disks for balancing.
- Load filled bags towards the outside wall of the bucket, away from the centrifuge drive spindle. Place the ADSOL bag between the blood bag and the plasma bag.

References

List of References

Documents listed below* can be obtained by calling Beckman Coulter Customer Service at 1-800-742-2345 (U.S.A. or Canada) or by contacting your local Beckman Coulter office. They are also available at www.beckmancoulter.com.

IN-181	<i>How to use Quick-Seal Tubes with the Beckman Coulter Cordless Tube Topper.</i>
IN-192	<i>Use and Care of Centrifuge Tubes and Bottles</i>
A-1792	<i>Rapid Pelleting of Bacteria in the Avanti J Centrifuge</i>
A-1796	<i>Rapid Protease Screening Using Ammonium Sulfate Precipitation and the Avanti J-25 High-Performance Centrifuge</i>
A-1803	<i>Isolation of Mitochondria from Saccharomyces Cerevisiae Using the Avanti J High-Performance Centrifuge A-1817</i>
	<i>Isolation of Plasma Membrane Sheets Using the Avanti J-30I High-Performance Centrifuge</i>
A-1818	<i>Baculovirus Purification Using the Avanti J-30I Centrifuge with the JS-24.38 and JS-24.15 Rotors</i>
A-1836	<i>Rapid Pelleting of Bacteria and Other Solids in the Avanti J-20 or J-20I Centrifuge</i>
A-1837	<i>Large- or Small-Scale Isolation of Chloroplasts Using the Avanti J Series of High-Performance Centrifuges</i>
A-1857	<i>Use of the JS-4.2 ARIES Self-Balancing Rotor in Blood Component Manufacturing</i>
A-1930	<i>Blood Cell Separation Using Commercial Gradient</i>
A2036	<i>Biocontainment Capability of the AeroSeal Covers for the JS-5.3 Rotor</i>
BR-8102	<i>High Performance, High Speed, High Capacity Rotors, Tubes & Accessories</i>
DS-528	<i>Use of the w2t Integrator</i>
DS-719	<i>Use of k Factor for Estimating Run Times</i>
DS-763	<i>Plasmid DNA Mini-Preparation Using the JA-18.1 Rotor</i>
DS-764	<i>Process 50 Liters of Sample in a Simple Run</i>
DS-769	<i>Multitube Carrier Adds Versatility</i>

* For detailed information on a rotor, see the applicable individual rotor manual.

References

List of References

DS-776	<i>Using k Factor to Compare Efficiency of Fixed-Angle Rotors</i>
DS-787	<i>New 4 × 50 mL Multitube Carrier for JS-7.5 Rotor</i>
DS-797	<i>New Large-Capacity Multitube Carrier Holds 12 × 75 mm Tubes in the JS-7.5 Rotor</i>
DS-828	<i>Using k Factors to Compare Rotor Efficiency</i>
DS-829	<i>Method for Plasmid DNA Mini-Preparation Using the JA-18.1 Rotor</i>
DS-885	<i>Using the JA-18 Rotor to Process Large Volumes Rapidly</i>
DS-916	<i>Discovery Video Series, "Principles and Practices of Centrifugation"</i>
DS-917	<i>JS-4.3 Rotor</i>
DS-7939	<i>MicroPlus Carrier</i>
DS-8028	<i>The JLA-10.500 Rotor with 3-Liter Capacity</i>
DS-8056	<i>Avanti® J-25 Rotor Data Sheet</i>
DS-8110	<i>Avanti® JS-24.15 and 24.38 Data Sheet</i>
DS-8113	<i>JLA 16.250 Rotor Data Sheet</i>
DS-8141	<i>Avanti® JA-30.50 Ti Data Sheet</i>
DS-8185	<i>JLA-8.1000 Rotor Data Sheet</i>
DS-8225	<i>JLA 9.1000 Rotor Data Sheet</i>
DS-8247	<i>JS-4.2A and JS-4.2SMA</i>
DS-8897	<i>JS-5.3 4-Place Swinging-Bucket Rotor for Avanti J-20XP</i>
DS-8898	<i>HarvestLine System Liners</i>
DS-8901	<i>JS-5.9 2-Place Swinging-Bucket Rotor for Avanti J-25 Series and J-30I</i>
DS-9240	<i>AllSpin JS-5.3 Rotor for Avanti J-E and J-26XP Centrifuges</i>
DS-9876	<i>Avanti J Series of BioSafe Centrifuge Systems</i>
DS-9980	<i>Avanti J-26 XP Series High-Performance Centrifuges</i>
DS-9982	<i>High Performance Rotors: Accelerate Your Sample Throughput and Increase Your Labs Versatility</i>
DS-9983	<i>Avanti J-30I High-Performance Centrifuge and Avanti J-HC High Capacity Bioprocessing Centrifuge Systems</i>
DS10009	<i>Avanti J-25 Series High-Performance Centrifuges</i>
FL-9987	<i>F10BCI 6x500y FIBERLite Carbon Fiber Rotor for Avanti J Series High-Performance Centrifuges</i>
SB-812	<i>J6 Series High-Capacity Centrifuges</i>
T-1735	<i>Modification of the Model J6-MC for Blood Component Preparation</i>
T-1741	<i>Optimizing Radioimmunoassays with the JR-3.2 Rack Rotor</i>
T-1783	<i>Higher Performance: The New Avanti J Series Switched Reluctance Drive</i>
T-1787	<i>Friction Reduction System Advantages in High-Performance Centrifugation</i>
T-1797	<i>Moving "High Speed" to Higher Performance Centrifugation: Applications History to Avanti J Series Performance</i>

Glossary

ADSOL — Dextrose-sodium chloride-mannitol-adenine; an additive used as a stabilizer for red blood cells that extends red cell life in CPD

Angular velocity, ω — Rate of rotation, measured in radians per second

$$\omega = \frac{2\pi \text{ rpm}}{60}$$

or

$$\omega = 0.10472 \text{ rpm}$$

Anodized coating — A thin, hard layer of aluminum oxide formed electrochemically on aluminum rotor and/or accessory surfaces as a protective coating for corrosion resistance

Autoclaving — Sterilization by heat (dry or steam)

Buoyant density — The density of a particle in a specified liquid medium

Buna N — Black nitrile rubber used for O-rings and gaskets in rotor assemblies; should be used at temperatures between –34 and 121°C (–30 and 250°F)

Centrifugal effect — Accumulated value of:

$$\int_{t_1}^{t_2} \omega^2 dt$$

where t is time and ω is angular velocity

Centrifugal force — In a centrifugal field, the force which causes a particle to move away from the center of rotation

Clearing factor k — Calculated for all Beckman Coulter high-speed rotors as a measure of the rotor's relative pelleting efficiency:

$$k = \frac{\ln(r_{\max}/r_{\min})}{\omega^2} \times \frac{10^{13}}{3600}$$

or

$$k = \frac{253303 \times \ln(r_{\max}/r_{\min})}{(\text{RPM} / 1000)^2}$$

Clearing time, t — $t = k/s$, where t is time in hours, k is the clearing factor of the rotor, and s is the sedimentation coefficient in Svedberg units (S)

CPD — Citrate-phosphate-dextrose; anti-coagulant and preservative

CPDA-1 — Citrate-phosphate-dextrose-citrate-citric acid-adenine; anti-coagulant and preservative

Cryoprecipitate — A precipitate, such as cryoglobulin or antihemophilic factor VIII, that results from cooling

CsCl — Cesium chloride; a high-density salt used in solution in isopycnic separations to separate particles based on their density

Cs₂SO₄ — Cesium sulfate; a salt, similar to CsCl, that will form its own gradient in solution

Delrin — Thermoplastic material (acetal homopolymer) used for most tube adapters (Delrin is a registered trademark of E.I. Du Pont de Nemours & Company.)

Density — Mass per unit volume

Density separation — A centrifugal separation process based on differences in particle densities

Differential separation — A centrifugal separation process based on differences in particle sizes

EPDM — Ethylene propylene rubber used for O-rings and pad adapters; should be used at temperatures between –57 and 120°C (–70 and 250°F)

Erythrocytes — See RBC (red blood cells)

Ethidium bromide — A fluorescent intercalating orange dye used commonly in the separation of DNA and in gel electrophoresis

Fixed-angle rotor — A rotor in which the tubes are held at an angle (usually 20 to 45 degrees) from the axis of rotation

Granulocytes — Generic name for three leukocyte (white blood cell) types characterized by having granules in their cytoplasm

Hard spin — Centrifugation run (5 to 7 minutes), at high *g*-forces (4000 to 5000 × *g*) at ambient temperature or at 4°C, used to separate fresh plasma from cellular components

HDPE — High density polyethylene used for adapters

Isopycnic — A method of particle separation or isolation based on particle buoyant density; particles are centrifuged until they reach a point in the gradient where the density of the particle is the same as the density of the gradient at that point

LDPE — Low density polyethylene used for tubes and bottles

Leukopheresis — Procedure in which leukocytes, or white blood cells, are separated from whole blood

Leukocytes — See WBC (white blood cells)

Lymphocyte — A type of leukocyte formed in the lymph nodes, other lymphoid tissue, and bone marrow; about a quarter of the white blood cells in the circulating blood are lymphocytes

Maximum volume — The maximum volume at which a tube should be filled for centrifugation (sometimes referred to as maximum fill volume or nominal fill volume)

Meniscus — The curved upper surface of a liquid column that is concave when the container walls are wetted by the liquid and convex when they are not

NaCl — Sodium chloride; a lower-density salt than CsCl, primarily used in lipoprotein type separations

Neoprene — Black synthetic elastomer used for O-rings in some tube caps and bottle cap assemblies; should be used at temperatures between -54 and 121°C (-65 and 250°F)

Noryl — Modified thermoplastic polyphenylene oxide (PPO) used for floating spacers (part of the *g*-Max system) and some polycarbonate bottle caps (Noryl is a registered trademark of GE Plastics.)

Pelleting — A centrifugal separation process in which particles in a sample sediment to the bottom of the tube (differential separation); differential pelleting separates particles of different sizes by successive centrifugation steps of progressively higher *g* force and/or longer run duration

PET — polyethylene terephthalate used in some adapters

Plasma — Major component of blood made up primarily of water, with substances such as albumin, globulins, coagulation factors, and electrolytes; distributes nutrients to the body, absorbs and carries away waste products

Plasmapheresis — Procedure in which whole blood is collected, platelets are separated via centrifugation, and platelet-poor red blood cells are returned to the donor; plasma is returned to the donor or collected for fractionation into clotting factors and albumin

Plateletpheresis — Procedure in which a unit of blood is taken to obtain plasma; following blood separation, red cells are immediately reinfused to the donor

Platelets — Blood component responsible for blood coagulation

Polypropylene — Random block copolymer of ethylene and propylene used for certain tubes.

Rack-type rotor — A rotor in which tubes are placed in gamma-counter racks; the racks are loaded into special plastic trays, which are then loaded into carriers that swing up to the horizontal position during centrifugation

Radel — Polyphenylsulfone (PPSU) used in plugs, cap closures, cannisters and other accessories

Rate zonal — A method of particle separation, based on differential rate of sedimentation, using a preformed gradient with the sample layered as a zone on top of the gradient

RBC — Red blood cells, or erythrocytes, carry oxygen to the tissues and carbon dioxide to the lungs for exhalation

RCF — Relative centrifugal field; the ratio of the centrifugal acceleration at a specified radius and speed ($r\omega^2$) to the standard acceleration of gravity (g) according to the following equation:

$$\text{RCF} = \frac{r\omega^2}{g}$$

where r is the radius in millimeters, ω is the angular velocity in radians per second ($2\pi \text{ RPM}/60$), and g is the standard acceleration of gravity (9807 mm/s^2). Thus the relationship between RCF and RPM is:

$$\text{RCF} = 1.12r \left(\frac{\text{RPM}}{1000} \right)^2$$

r_{max} — (Maximum radius) the position of the liquid in the tube at the maximum distance from the axis of rotation when the rotor is at speed

r_{min} — (Minimum radius) the position of the liquid in the tube at the minimum distance from the axis of rotation when the rotor is at speed

SAG-M — Saline-adenine-glucose-minitol; an additive used as a stabilizer for red cells that extends red cell life in CPD

Sedimentation — The settling out of particles from a suspension in the earth's field of gravity; in the centrifuge this process is accelerated and the particles move away from the axis of rotation

Sedimentation coefficient, s — Sedimentation velocity per unit of centrifugal force:

$$s = \frac{dr}{dt} \times \frac{1}{\omega^2 r}$$

SDS — Sodium dodecyl sulfate; an ionic detergent used in cell lysis and solubilizing of proteins

Silicone rubber — A large group of silicone elastomers used in various accessories; should be used at temperatures between -59 and 232°C (-75 and 450°F)

Soft spin — Short centrifugation run (3 to 5 minutes), at low g -forces (2000 to $3000 \times g$) at ambient temperature, used to keep small cells or platelets in suspension while the larger cellular components sediment; used to obtain platelet-rich plasma and red blood cell concentrate from whole blood

Solution 555 — Beckman Coulter concentrated rotor cleaning solution; recommended because it is a mild solution that has been tested and found effective and safe for Beckman Coulter rotors and accessories

Spinkote — Beckman Coulter lubricant for metal-to-metal contacts

Sucrose — A sugar (not a self-forming gradient) used in rate zonal separations; generally used in separating RNA, subcellular organelles, and cell membranes

Supernatant — The liquid above the sedimented material following centrifugation

Svedberg unit, S — A unit of sedimentation velocity:

$$1 S = 10^{-13} \text{ seconds}$$

Swinging-bucket rotor — A rotor in which the tubes or bottles are carried in buckets, microtiter plate carriers, or racks that swing up to the horizontal position during centrifugation (sometimes referred to as a horizontal or swing-out rotor)

Ultem — Polyetherimide (PEI)—used in adapters, covers, and spacers; should be used at temperatures between -29 and 204°C (-20 and 400°C) (Ultem is a registered trademark of GE Plastics.)

Vertical-tube rotor — A rotor in which the tubes or bottles are held parallel to the axis of rotation

Viton — Fluorocarbon elastomer used in high-temperature applications (Viton is a registered trademark of E.I. Du Pont de Nemours & Company.)

WBC — White blood cells, or leukocytes, protect the body against infection and many diseases

Wettable — Tube or bottle material that water or other aqueous solution will adhere to; the more wettable a tube or bottle material is, the more biological material, DNA, protein, cells, and so forth, will adhere to the walls

Beckman Coulter, Inc.

J Series Rotor Warranty

Subject to the conditions specified below and the warranty clause of the Beckman Coulter, Inc., terms and conditions of sale in effect at the time of sale, Beckman Coulter, Inc. agrees to correct either by repair, or, at its election, by replacement, any defects of material or workmanship which develop within seven (7) years after delivery of a J series rotor to the original buyer by Beckman Coulter, Inc. or by an authorized representative, provided that investigation and factory inspection by Beckman Coulter, Inc. discloses that such defect developed under normal and proper use. Should a Beckman Coulter centrifuge be damaged due to a failure of a rotor covered by this warranty, Beckman Coulter will supply free of charge all centrifuge parts required for repair.

Replacement

Any product claimed to be defective must, if requested by Beckman Coulter, Inc., be returned to the factory, transportation charges prepaid, and will be returned to Buyer with the transportation charges collect unless the product is found to be defective, in which case Beckman Coulter, Inc. will pay all transportation charges.

A defective rotor will be replaced by Beckman Coulter, Inc. at its then current list price less a credit based upon the age of the rotor (years since date of purchase). The Buyer shall not receive credit until the claimed defective rotor is returned to Beckman Coulter's Indianapolis, Indiana, facility or delivered to a Beckman Field Service representative.

The replacement price (cost to Buyer) for the respective rotor shall be calculated as follows:

$$\text{Replacement price} = \text{Current rotor list price} \times \frac{\text{years}}{7}$$

Conditions

1. Except as otherwise specifically provided herein, this warranty covers the rotor only and Beckman Coulter, Inc. shall not be liable for damage to accessories or ancillary supplies including but not limited to (i) tubes, (ii) tube caps, (iii) tube adapters, or (iv) tube contents.
2. This warranty is void if the rotor has been subjected to customer misuse such as operation or maintenance contrary to the instructions in the Beckman Coulter rotor or centrifuge manual.
3. This warranty is void if the rotor is operated with a rotor drive unit or in a centrifuge unmatched to the rotor characteristics, or is operated in a Beckman Coulter centrifuge that has been improperly disassembled, repaired, or modified.
4. Each bucket, whether purchased with a rotor assembly or purchased separately, is covered by this warranty for seven (7) years from the date of purchase, and will be replaced or repaired during such period according to the terms and conditions of this warranty. The date of manufacture marked on the bucket may be earlier than the date of purchase, and the expiration date marked on the bucket, which is seven (7) years after the date of purchase, may be correspondingly offset.
5. Buckets should not be used after the expiration date marked on the bucket. If at the time of purchase the marked expiration date is less than 7 years from the date of purchase, the expiration date becomes the date of purchase plus seven (7) years. Use of a bucket after such expiration date voids Beckman Coulter's warranty obligations with respect to any rotor and/or centrifuge in which such a bucket is used.

Disclaimer

IT IS EXPRESSLY AGREED THAT THE ABOVE WARRANTY SHALL BE IN LIEU OF ALL WARRANTIES OF FITNESS AND OF THE WARRANTY OF MERCHANTABILITY AND BECKMAN COULTER, INC. SHALL HAVE NO LIABILITY FOR SPECIAL OR CONSEQUENTIAL DAMAGES OF ANY KIND WHATSOEVER ARISING OUT OF THE MANUFACTURE, USE, SALE, HANDLING, REPAIR, MAINTENANCE, OR REPLACEMENT OF THE PRODUCT.

Related Documents

Rotors and Tubes CD (369668)

- Rotors and Tubes for Tabletop Preparative Ultracentrifuges
- Rotors and Tubes for J2, J6, Avanti J Series Centrifuges
- Rotors and Tubes for Preparative Ultracentrifuges
- Rotor Safety Bulletin
- Chemical Resistances for Beckman Coulter Centrifugation Products

Included with shipment of instrument.

Additional References

- Chemical Resistances for Beckman Coulter Centrifugation Products (IN-175)
- Beckman Coulter High Performance, High Speed, High Capacity Rotors, Tubes & Accessories catalog (BR-8102)
- HarvestLine System Liner Kit 369264 (J-TB-093)
- Use and Care of Centrifuge Tubes and Bottles (IN-192)
- The JE-6B Elutriation System and Rotor (JE6B-IM-9)
- The JE 5.0- Elutriation System (JE5-IM-13)
- Run Speeds for Stainless Steel Tubes (L5-TB-072)
- How to Use the Beckman Coulter Cordless Tube Topper with Quick-Seal Tubes (IN-181)
- Beckman Coulter Fraction Recovery System (L5-TB-081)
- JCF-Z Zonal and Continuous Flow Rotor (JCFZ-IM-12)
- Instructions for Using Micro Plus Multiwell Plate Carriers (GS6-TB-011)
- Instructions for Using Microplate Carriers (J6-TB-009)

Available in hard copy or electronic pdf by request.

www.beckmancoulter.com

